

Human Resistin ELISA KIT

Catalog Number

EA100063

Size

48 Tests



Human Resistin ELISA KIT

For the quantitative determination of human Resistin concentrations in cell culture supernates, serum, and plasma. This package insert must be read in its entirety before using this product. If you have questions or experience problems with this product, please contact our Technical Support staff. Our scientists commit themselves to providing rapid and effective help.

**FOR RESEARCH USE ONLY
NOT FOR USE IN DIAGNOSTIC PROCEDURES**

INTRODUCTION

Resistin, also known as Found In Inflammatory Zone 3 (FIZZ3) or Adipocyte Secreted Factor(ADSF), is a member of a protein family known as the Resistin-like molecules (RELMs). It is perhaps best known for its potential as a link between obesity and the development of insulin resistance (1). Other members of this family include RELM- α /FIZZ1 and RELM- γ , which are described in rodents but as yet have no identified human counterparts, and RELM- β /FIZZ2(2, 3). The Resistin amino acid (aa) sequence contains a putative N-terminal signal sequence and a motif containing 11 cysteine residues, 10 of which are characteristic of the RELM family(1-3). The protein is thought to be secreted as a dimer and potentially exists in higher order molecular structures resulting from interactions between Resistin dimers or other members of the RELM family (4-7). A splice variant in the rat, lacking the signal sequence and localized predominantly to the nucleus, has also been described (8). A large 3' intron is the primary reason that the mouse genomic sequence is 3-fold larger than the corresponding human sequence (9). Mouse and human Resistin share only 53 percent identity at the aa level and exhibit differences in expression patterns (1, 9, 10). In mouse, expression appears primarily in adipose tissues (1). Although some human studies suggest Resistin is expressed by adipose tissues as well, the most significant source appears to be blood mononuclear cells (11-13). In humans, Resistin is also reported to be expressed by pre-adipocytes (14), placenta (15), pancreatic islets (16), and primary leukemia cells (10). A receptor for Resistin has not yet been described.

Resistin acquired initial attention as a potential link between obesity and glucose regulation. Serum levels were shown to increase in diet-induced and genetic forms of obesity in mice (*ob/ob* and *db/db*) and decrease in response to insulin sensitizing drugs (TZDs) (1). In addition, function-blocking Resistin antibodies enhanced insulin actions while treatment with recombinant Resistin caused glucose intolerance and insulin resistance (1). Resistin knockout mice exhibit decreased fasting blood glucose levels as a result of reduced hepatic output (17). To establish a physiological role in humans, several studies have examined whether altered circulating Resistin levels are associated with type 2 diabetes, insulin resistance, and/or obesity. Although some demonstrate significant correlations (18-23), others report no correlation (23-28), suggesting that in humans fundamental questions remain regarding Resistin's role in these pathophysiological processes (29, 30). Resistin expression by human mononuclear cells could indicate a potential role in inflammation. In vitro,

Resistin expression by these cells is enhanced by treatment with several pro-inflammatory cytokines including IL-1 β , TNF- α , IFN- γ , or IL-6 (31). In addition, Resistin has been shown to activate endothelial cells in vitro, leading to the production of adhesion molecules, Endothelin-1, and chemokines (32, 33)..

PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for Resistin has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any Resistin present is bound by the immobilized antibody. Following incubation unbound samples are removed during a wash step, and then a detection antibody specific for Resistin is added to the wells and binds to the combination of capture antibody- Resistin in sample. Following a wash to remove any unbound combination, and enzyme conjugate is added to the wells. Following incubation and wash steps a substrate is added. A coloured product is formed in proportion to the amount of Resistin present in the sample. The reaction is terminated by addition of acid and absorbance is measured at 450nm. A standard curve is prepared from seven Resistin standard dilutions and Resistin sample concentration determined.




Figure 1: Schematic diagram of the assay

REAGENTS

1. Aluminium pouches with a Microwell Plate coated with antibody to human Resistin (8X12)
2. 2 vials human Resistin Standard lyophilized, 1000 pg/ml upon reconstitution
3. 2 vials concentrated Biotin-Conjugate anti-human Resistin antibody

4. 2 vials Streptavidin-HRP solution
5. 4 bottle Standard /sample Diluent
6. 1 bottle Biotin-Conjugate antibody Diluent
7. 1 bottle Streptavidin-HRP Diluent
8. 1 bottle Wash Buffer Concentrate 20x (PBS with 1% Tween-20)
9. 1 vial Substrate Solution
10. 1 vial Stop Solution
11. 4 pieces Adhesive Films
12. Package insert

NOTE: [96 Tests]

STORAGE

Table 1: Storage of the kit

Unopened Kit	Store at 2 – 8°C. Do not use past kit expiration date.	
Opened/ Reconstituted Reagents	Standard /sample Diluent	May be stored for up to 1 month at 2 – 8°C.**
	Concentrated Biotin-Conjugate	
	Streptavidin-HRP solution	
	Biotin-Conjugate antibody Diluent	
	Streptavidin-HRP Diluent	
	Wash Buffer Concentrate 20x	
	Substrate Solution	
	Stop Solution	
	Standard	Aliquot and store for up to 1 month at -20°C. Avoid repeated freeze-thaw cycles. Diluted standard shall not be reused.
	Microplate Wells	Return unused wells to the foil pouch containing the desiccant pack, reseal along entire edge of zip-seal. May be stored for up to 1 month at 2 – 8°C.**

**Provided this is within the expiration date of the kit

THE REQUIRED ITEMS (not provided, but can help to buy):

1. Microplate reader (450nm).
2. Micro-pipette and tips: 0.5-10, 2-20, 20-200, 200-1000 μL .
3. 37 °C incubator, double-distilled water or deionized water, coordinate paper, graduated cylinder.

PRECAUTIONS FOR USE

1. Store kit reagents between 2°C and 8°C. After use all reagents should be immediately returned to cold storage (2°C to 8°C).
2. Please perform simple centrifugation to collect the liquid before use.
3. To avoid cross contamination, please use disposable pipette tips.
4. The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and clothing protection when using this material. Avoid contact of skin or mucous membranes with kit reagents or specimens. In the case of contact with skin or eyes wash immediately with water.
5. Use clean, dedicated reagent trays for dispensing the washing liquid, conjugate and substrate reagent. Mix all reagents and samples well before use.
6. After washing microtiter plate should be fully pat dried. Do not use absorbent paper directly into the enzyme reaction wells.
7. Do not mix or substitute reagents with those from other lots or other sources. Do not use kit reagents beyond expiration date on label.
8. Each sample, standard, blank and optional control samples should be assayed in duplicate or triplicate.
9. Adequate mixing is very important for good result. Use a mini-vortexer at the lowest frequency or Shake by hand at 10min interval when there is no vortexer.
10. Avoid microtiter plates drying during the operation.
11. Dilute samples at the appropriate multiple, and make the sample values fall within the standard curve. If samples generate values higher than the highest standard, dilute the samples and repeat the assay.
12. Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time and temperature, and kit age can cause variation in binding.

13. This method can effectively eliminate the interference of the soluble receptors, binding proteins and other factors in biological samples.

SAMPLE COLLECTION AND STORAGE

1. **Cell Culture Supernates** - Remove particulates by centrifugation.
2. **Serum** - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at approximately 1000 x g. Remove serum, avoid hemolysis and high blood lipid samples.
3. **Plasma** - Recommended EDTA as an anticoagulant in plasma. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection.
4. Assay immediately or aliquot and store samples at -20°C. Avoid repeated freeze-thaw cycles.
5. Dilute samples at the appropriate multiple (recommended to do pre-test to determine the dilution factor).

REAGENT PREPARATION

1. Bring all reagents to room temperature before use.
2. **Wash Buffer** - Dilute 10 mL of Wash Buffer Concentrate into deionized or distilled water to prepare 200 mL of Wash Buffer. If crystals have formed in the concentrate Wash Buffer, warm to room temperature and mix gently until the crystals have completely dissolved.
3. **Standard** - Reconstitute the Standard with 1.0 mL of Standard /sample Diluent. This reconstitution produces a stock solution of 1000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions.

Pipette 500 µL of Standard/sample Diluent into the 500 pg/mL tube and the remaining tubes. Use the stock solution to produce a 2-fold dilution series (below). Mix each tube thoroughly and change pipette tips between each transfer. The 500 pg/mL standard serves as the high standard. The Standard/ sample Diluent serves as the zero standard (0 pg/mL).

If you do not run out of re-melting standard, store it at -20°C. Diluted standard shall not be reused.

4. Working solution of Biotin-Conjugate anti-human Resistin antibody: Make a 1:100 dilution of the concentrated Biotin-Conjugate solution with the Biotin-Conjugate antibody Diluent in a clean plastic tube.

The working solution should be used within one day after dilution.

5. Working solution of Streptavidin-HRP: Make a 1:100 dilution of the concentrated Streptavidin-HRP solution with the Streptavidin-HRP Diluent in a clean plastic tube.

The working solution should be used within one day after dilution.




Figure 2: Preparation of Resistin standard dilutions

GENERAL ELISA PROTOCOL

1. Prepare all reagents and working standards as directed in the previous sections.
2. Determine the number of microwell strips required to test the desired number of samples plus appropriate number of wells needed for running blanks and standards. Remove extra microwell strips from holder and store in foil bag with the desiccant provided at 2-8°C sealed tightly.
3. Add 100µL of Standard, control, or sample, per well. Cover with the adhesive strip provided. Incubate for 1.5 hours at 37°C.
4. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with Wash Buffer (350µL) using a squirt bottle, manifold dispenser or auto-washer. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add 100µL of the working solution of Biotin-Conjugate to each well. Cover with a new adhesive strip and incubate 1 hour at 37°C.
6. Repeat the aspiration/wash as in step 4.

7. Add 100µL of the working solution of Streptavidin-HRP to each well. Cover with a new adhesive strip and incubate for 30 minutes at 37°C. Avoid placing the plate in direct light.
8. Repeat the aspiration/wash as in step 4.
9. Add 100µL of Substrate Solution to each well. Incubate for 10-20 minutes at 37°C. Avoid placing the plate in direct light.
10. Add 100µL of Stop Solution to each well. Gently tap the plate to ensure thorough mixing.
11. Determine the optical density of each well immediately, using a microplate reader set to 450 nm.(optionally 630nm as the reference wave length;610-650nm is acceptable)

ASSAY PROCEDURE SUMMARY




Figure 3: Assay procedure summary

TECHNICAL HINTS

1. When mixing or reconstituting protein solutions, always avoid foaming.
2. To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.
3. To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
4. Substrate Solution should remain colorless until added to the plate. Stop Solution should be added to the plate in the same order as the Substrate Solution. Keep Substrate Solution protected from light. Substrate Solution should change from colorless to gradations of blue.
5. A standard curve should be generated for each set of samples assayed. According to the content of tested factors in the sample, appropriate diluted or concentrated samples, it is best to do pre-experiment.

CALCULATION OF RESULTS

1. Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density.
2. Create a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph.
3. The data may be linearized by plotting the log of the Resistin concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
4. This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

Table 2: Typical data using the Resistin ELISA (Measuring wavelength: 450nm, Reference wavelength: 650nm)

Standard (pg/ml)	OD.	OD.	Average	Corrected
0	0.050	0.047	0.049	—
7.8	0.097	0.091	0.094	0.045
15.625	0.131	0.125	0.128	0.079
31.25	0.213	0.207	0.210	0.161
62.5	0.376	0.369	0.373	0.324
125	0.643	0.637	0.640	0.591
250	1.152	1.143	1.148	1.099
500	2.092	2.078	2.085	2.036




Figure 4: Representative standard curve for Resistin ELISA. Resistin was diluted in serial two-fold steps in Sample Diluent.

Do not use this standard curve to derive test results. A standard curve must be run for each group of microwell strips assayed.

SENSITIVITY, SPECIFICITY AND REPEATABILITY

1. **REPEATABILITY:** The coefficient of variation of both intra-assay and inter-assay were less than 10%.
2. **SENSITIVITY:** The minimum detectable dose was 4 pg/mL.
3. **SPECIFICITY:** This assay recognizes both natural and recombinant human Resistin. The factors listed below were prepared at 100 ng/ml in Standard /sample Diluent and assayed for cross-reactivity and no significant cross-reactivity or interference was observed.

Table 3: Factors assayed for cross-reactivity

Recombinant human	Recombinant mouse	
Leptin	Leptin	
Leptin R	Leptin R	
LIF	LIF	
RELM- β	RELM- α	
	Resistin	

REFERENCES

1. Steppan, C.M. et al. (2001) Nature 409:307.
2. Steppan, C.M. et al. (2001) Proc. Natl. Acad. Sci. USA 98:502.
3. Gerstmayr, B. et al. (2003) Genomics 81:588.
4. Raghu, P. et al. (2004) Biochem. Biophys. Res. Commun. 313:642.
5. Banerjee, R.R. and M.A. Lazar (2001) J. Biol. Chem. 276:25970.
6. Chen, J. et al. (2002) J. Endocrinol. 175:499.
7. Aruna, B. et al. (2003) Biochemistry 42:10554.
8. Del Arco, A. et al. (2003) FEBS Lett. 555:243.
9. Ghosh, S. et al. (2003) Gene 305:27.
10. Yang, R.Z. et al. (2003) Biochem. Biophys. Res. Commun. 310:927.
11. McTernan, P.G. et al. (2002) J. Clin. Endocrinol. Metab. 87:2407.
12. Nagaev, I. and U. Smith (2001) Biochem. Biophys. Res. Commun. 285:561.
13. Savage, D.B. et al. (2001) Diabetes 50:2199.
14. Janke, J. et al. (2002) Obes. Res. 10:1.
15. Yura, S. et al. (2003) J. Clin. Endocrinol. Metab. 88:1394.

16. Minn, A.H. et al. (2003) *Biochem. Biophys. Res. Commun.* 310:641.
17. Banerjee, R.R. et al. (2004) *Science* 303:1195.
18. McTernan, P.G. et al. (2003) *J. Clin. Endocrinol. Metab.* 88:6098.
19. Degawa-Yamauchi, M. et al. (2003) *J. Clin. Endocrinol. Metab.* 88:5452.
20. Youn, B.S. et al. (2004) *J. Clin. Endocrinol. Metab.* 89:150.
21. Fujinami, A. et al. (2004) *Clin. Chim. Acta* 339:57.
22. Azuma, K. et al. (2003) *Obes. Res.* 11:997.
23. Silha, J.V. et al. (2003) *Eur. J. Endocrinol.* 149:331.
24. Fehmann, H.C. and J. Heyn (2002) *Horm. Metab. Res.* 34:671.
25. Pfutzner, A. et al. (2003) *Clin. Lab.* 49:571.
26. Kielstein, J.T. et al. (2003) *Am. J. Kidney Dis.* 42:62.
27. Lee, J.H. et al. (2003) *J. Clin. Endocrinol. Metab.* 88:4848.
28. Furuhashi, M. et al. (2003) *Clin. Endocrinol.* 59:507.
29. Smith, U. (2002) *Obes. Res.* 10:61.
30. Ukkola, O. (2002) *Eur. J. Endocrinol.* 147:571.
31. Kaser, S. et al. (2003) *Biochem. Biophys. Res. Commun.* 309:286.
32. Kawanami, D. et al. (2004) *Biochem. Biophys. Res. Commun.* 314:415.
33. Verma, S. et al. (2003) *Circulation* 108:736.

If you have any questions, please tell us!