Human PBEF/Visfatin ELISA KIT

Catalog Number EA100091 Size 48 Tests



Human PBEF/Visfatin ELISA KIT

For the quantitative determination of human PBEF/Visfatin concentrations in cell culture supernates, serum, and plasma. This package insert must be read in its entirety before using this product. If you have questions or experience problems with this product, please contact our Technical Support staff. Our scientists commit themselves to providing rapid and effective help.

FOR RESEARCH USE ONLY
NOT FOR USE IN DIAGNOSTIC PROCEDURES

phone: 086-010-56315211 fax: 086-010-56315212 <u>www.4abio.com</u> Email: info@4abio.com

INTRODUCTION

PBEF (preB cell colonyenhancing factor; also visfatin and nicotinamide phosphoribosyltransferase) is a 52 kDa member of the NAPRTase family of molecules. It functions both intracellularly and extracellularly, where it participates in NAD synthesis and insulin receptor activation, respectively. Human PBEF is 491 amino acids in length and contains no signal sequence. There is at least one alternative splice form that shows a 5 aa substitution for the Cterminal 128 amino acids (aa 364491). Over aa 27491, human PBEF shares 96%, 97%, and 96% aa identity with mouse, porcine, and canine PBEF, respectively.




Figure 1:Schematic diagram of the assay

REAGENTS

- 1. Aluminium pouches with a Microwell Plate coated with antibody to human PBEF/Visfatin (8X12)
- 2. 2 vials human PBEF/Visfatin Standard lyophilized, 50 ng/ml upon reconstitution
- 2 vials concentrated Biotin-Conjugate anti-human PBEF/Visfatin antibody
- 4. 2 vials Streptavidin-HRP solution
- 5. 4 bottle Standard /sample Diluent
- 6. 1 bottle Biotin-Conjugate antibody Diluent
- 7. 1 bottle Streptavidin-HRP Diluent
- 8. 1 bottle Wash Buffer Concentrate 20x (PBS with 1% Tween-20)
- 9. 1 vial Substrate Solution

- 10. 1 vial Stop Solution
- 11. 4 pieces Adhesive Films
- 12. Package insert

NOTE: [96 Tests]

STORAGE

Table 1: Storage of the kit

Unopened Kit	Store at 2 – 8°C. Do not use past kit expiration date.					
	Standard /sample Diluent	May be stored for up to 1 month at 2 – 8°C.**				
Opened/ Reconstituted Reagents	Concentrated Biotin-Conjugate					
	Streptavidin-HRP solution					
	Biotin-Conjugate antibody Diluent					
	Streptavidin-HRP Diluent					
	Wash Buffer Concentrate 20x					
	Substrate Solution					
	Stop Solution					
	Standard	Aliquot and store for up to 1				
		month at -20°C.				
		Avoid repeated freeze-thaw				
		cycles. Diluted standard				
		shall not be reused.				
	Microplate Wells	Return unused wells to the foil pouch containing the desiccant pack, reseal along entire edge of zip-seal. May				
				be stored for up to 1 month		
					at 2 – 8°C.**	

^{**}Provided this is within the expiration date of the kit

THE REQUIRED ITEMS (not provided, but can help to buy):

- 1. Microplate reader (450nm).
- 2. Micro-pipette and tips: 0.5-10, 2-20, 20-200, 200-1000µL.

phone : 086-010-56315211 fax : 086-010-56315212 <u>www.4abio.com</u> Email: info@4abio.com

3. 37 °C incubator, double-distilled water or deionized water, coordinate paper, graduated cylinder.

PRECAUTIONS FOR USE

- 1. Store kit regents between 2°C and 8°C. After use all reagents should be immediately returned to cold storage (2°C to 8°C).
- 2. Please perform simple centrifugation to collect the liquid before use.
- 3. To avoid cross contamination, please use disposable pipette tips.
- 4. The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and clothing protection when using this material. Avoid contact of skin or mucous membranes with kit reagents or specimens. In the case of contact with skin or eyes wash immediately with water.
- Use clean, dedicated reagent trays for dispensing the washing liquid, conjugate and substrate reagent. Mix all reagents and samples well before use.
- 6. After washing microtiter plate should be fully pat dried. Do not use absorbent paper directly into the enzyme reaction wells.
- 7. Do not mix or substitute reagents with those from other lots or other sources. Do not use kit reagents beyond expiration date on label.
- 8. Each sample, standard, blank and optional control samples should be assayed in duplicate or triplicate.
- 9. Adequate mixing is very important for good result. Use a mini-vortexer at the lowest frequency or Shake by hand at 10min interval when there is no vortexer.
- 10. Avoid microtiter plates drying during the operation.
- 11. Dilute samples at the appropriate multiple, and make the sample values fall within the standard curve. If samples generate values higher than the highest standard, dilute the samples and repeat the assay.
- 12. Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time and temperature, and kit age can cause variation in binding.
- 13. This method can effectively eliminate the interference of the soluble receptors, binding proteins and other factors in biological samples.

SAMPLE COLLECTION AND STORAGE

- 1. **Cell Culture Supernates** Remove particulates by centrifugation.
- 2. **Serum** Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at approximately 1000 x g. Remove serum, avoid hemolysis and high blood lipid samples.
- 3. **Plasma** Recommended EDTA as an anticoagulant in plasma. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection.
- 4. Assay immediately or aliquot and store samples at -20°C. Avoid repeated freeze-thaw cycles.
- 5. Dilute samples at the appropriate multiple (recommended to do pre-test to determine the dilution factor).

REAGENT PREPARATION

- 1. Bring all reagents to room temperature before use.
- 2. Wash Buffer Dilute 10mL of Wash Buffer Concentrate into deionized or distilled water to prepare 200mL of Wash Buffer. If crystals have formed in the concentrate Wash Buffer, warm to room temperature and mix gently until the crystals have completely dissolved.
- **3. Standard** Reconstitute the Standard with 0.5mL of Standard /sample Diluent. This reconstitution produces a stock solution of 100 ng/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions.
 - Pipette 250µL of Standard/sample Diluent into the 50 ng/mL tube and the remaining tubes. Use the stock solution to produce a 2-fold dilution series (below). Mix each tube thoroughly and change pipette tips between each transfer. The 100 ng/mL standard serves as the high standard. The Standard/ sample Diluent serves as the zero standard (0 pg/mL).
 - If you do not run out of re-melting standard, store it at -20°C. Diluted standard shall not be reused.
- 4. Working solution of Biotin-Conjugate anti-human PBEF/Visfatin antibody: Make a 1:100 dilution of the concentrated Biotin-Conjugate solution with the Biotin-Conjugate antibody Diluent in a clean plastic tube.
 - The working solution should be used within one day after dilution.
- Working solution of Streptavidin-HRP: Make a 1:100 dilution of the concentrated Streptavidin-HRP solution with the Streptavidin-HRP

Diluent in a clean plastic tube.

The working solution should be used within one day after dilution.




Figure 2: Preparation of PBEF/Visfatin standard dilutions

GENERAL ELISA PROTOCOL

- 1. Prepare all reagents and working standards as directed in the previous sections.
- Determine the number of microwell strips required to test the desired number of samples plus appropriate number of wells needed for running blanks and standards. Remove extra microwell strips from holder and store in foil bag with the desiccant provided at 2-8°C sealed tightly.
- Add 100μL of Standard, control, or sample, per well, then add 50μL
 of the working solution of Biotin-Conjugate to each well. Cover with
 the adhesive strip provided and incubate 2 hours at RT. Adequate
 mixing is very important for good result. Use a mini-vortexer at the
 lowest frequency.
- 4. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with Wash Buffer (350 μL) using a squirt bottle, manifold dispenser or auto-washer. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
- 5. Add 100µL of the working solution of Streptavidin-HRP to each well.

phone: 086-010-56315211 fax: 086-010-56315212 <u>www.4abio.com</u> Email: info@4abio.com

- Cover with a new adhesive strip and incubate for 30 minutes at RT. Avoid placing the plate in direct light.
- 6. Repeat the aspiration/wash as in step 4.
- 7. Add 100µL of Substrate Solution to each well. Incubate for 10-20 minutes at RT. Avoid placing the plate in direct light.
- 8. Add 100µL of Stop Solution to each well. Gently tap the plate to ensure thorough mixing.
- 9. Determine the optical density of each well immediately, using a microplate reader set to 450 nm.(optionally 650nm as the reference wave length;610-650nm is acceptable)

ASSAY PROCEDURE SUMMARY

Prepare all reagents and standards as directed



Add 100μL standard or samples to each well, and add 50μL working solution of Biotin-Conjugate antibody to each well, incubate 120 minutes.RT

Aspirate and wash 4 times

Add 100µL working solution of Streptavidin-HRP to each well, incubate 30 minutes,RT

Aspirate and wash 4 times

Add 100µL Substrate solution to each well, incubate 10-20 minutes,RT.Protect from light.

 \prod

Add 100µL Stop solution to each well. Read at 450nm within 30 minutes.

Figure 3: Assay procedure summary

phone: 086-010-56315211 fax: 086-010-56315212 <u>www.4abio.com</u> Email: info@4abio.com

TECHNICAL HINTS

- 1. When mixing or reconstituting protein solutions, always avoid foaming.
- 2. To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.
- 3. To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
- 4. Substrate Solution should remain colorless until added to the plate. Stop Solution should be added to the plate in the same order as the Substrate Solution. Keep Substrate Solution protected from light. Substrate Solution should change from colorless to gradations of blue.
- A standard curve should be generated for each set of samples assayed.
 According to the content of tested factors in the sample, appropriate diluted or concentrated samples, it is best to do pre-experiment.

CALCULATION OF RESULTS

- 1. Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density.
- 2. Create a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph.
- 3. The data may be linearized by plotting the log of the PBEF/Visfatin concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
- 4. This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

Table 2:Typical data using the PBEF/Visfatin ELISA (Measuring wavelength: 450nm,Reference wavelength: 650nm)

Standard (ng/ml)	OD.	OD.	Average	Corrected
0	0.089	0.085	0.087	
1.56	0.130	0.124	0.127	0.040
3.125	0.162	0.157	0.160	0.073
6.25	0.234	0.233	0.234	0.147
12.5	0.358	0.351	0.355	0.268
25	0.590	0.589	0.590	0.503
50	1.066	1.059	1.063	0.976
100	1.708	1.714	1.711	1.624




Figure 4:Representative standard curve for PBEF/Visfatin ELISA. PBEF/Visfatin was diluted in serial two-fold steps in Sample Diluent.

Do not use this standard curve to derive test results. A standard curve must be run for each group of microwell strips assayed.

SENSITIVITY, SPECIFICITY AND REPEATABILITY

- 1. **REPEATABILITY**: The coefficient of variation of both intra-assay and inter-assay were less than 10%.
- 2. **SENSITIVITY**: The minimum detectable dose was 0.70 ng/mL.

REFERENCES

- 1.Kitani,T,Okuno,S.,Fujisawa,H(2003) Growth phase-dependent changes in the subcellular localization of per-B-cell colony-enhancing factor,FEBSLett,544.,74-78
- 2.Martin, P.R., Shea, R.J., Mulks, M.H. (2001) Identication of a plasmid-encoded gene from Haemophilus ducreyi which confers NAD independence, J., Bacteriol, 183, 1168-1174
- 3.Ognjanovic,S.,Bao,S.,Yamamoto,S.Y.,Garibay-Tupas,J.,Samal,B.,Bryant-Green wood,G,D,(2001) Genomic organization of the gene coding for human-pre-B-cell colong enhancing factor and expression in human fetal membranes.J.Mol.Endocrinol.26,107-117
- 4.Ye,S,O.,Simon,B,A.,Maloney,J,P.,Zambelliweiner,A.,Gao,L.,Grant,A.,Easley,R,B.,Mcverry,B,J.,Tuder,R,M.,Standiford,T.,Brower,R,G.,Barnes,K,C.,Garcia,J,G.(2005)Pre-B-cell colony-enhancing facto rasapotential novel biomarkerin acute lung injury,Am,J.Respir,Crit.CareMed.171,360-370
- 5.Stephens JM., Vidal-Puig AJ, (2006) An update on visfatin/pre-B cell colony enhancing factor, an ubiquitously expressed, illusive cytokine that is regulated in obesity, Curr Opin Lipidol, 17(2);128-131

If you have any questions, please tell us!