

Phone: 1.888.267.4436 Fax: 301-340-9254 Email: techsupport@origene.com Web: www.origene.com

Mouse TNFR2 ELISA Kit

Catalog No. EA100370

Size 96T(8×12 divisible strips)

For quantitative detection of mouse TNFR2 in cell culture supernates and serum.


Typical Data Obtained from Mouse TNFR2

(TMB reaction incubate at 37°C for 23 min)

| Concentration(pg/ml) | 0.0 | 15.6 | 31.2 | 62.5 | 125 | 250 | 500 | 1000 |
|----------------------|-------|-------|-------|-------|-------|-------|-------|-------|
| O.D | 0.031 | 0.094 | 0.177 | 0.315 | 0.587 | 1.032 | 1.701 | 2.469 |

Typical Mouse TNFR2 ELISA Kit Standard Curve

This standard curve was generated at OriGene for demonstration purpose only. A standard curve must be run with each assay.



Range 15.6pg/ml-1000pg/ml

Sensitivity < 1pg/ml

Specificity Natural and recombinant mouse TNFR2

Cross-reactivity No detectable cross-reactivity with other relevant proteins

Storage

Store at 4°C for 6 months, at -20°C for 12 months. Avoid multiple freeze-thaw cycles (Shipped with wet ice.)

Precision

Intra-Assay Precision (Precision within an assay) Three samples of known concentration were tested on one plate to assess intra-assay precision.

Inter-Assay Precision (Precision between assays) Three samples of known concentration were tested in separate assays to assess inter-assay precision.

| | Intra-Assay Precision | | | Inter-Assay Precision | | |
|--------------------|-----------------------|-------|------|-----------------------|-------|------|
| Sample | 1 | 2 | 3 | 1 | 2 | 3 |
| n | 16 | 16 | 16 | 24 | 24 | 24 |
| Mean(pg/ml) | 92 | 263 | 492 | 104 | 253 | 527 |
| Standard deviation | 4.6 | 11.57 | 27.6 | 5.4 | 14.67 | 32.2 |
| CV(%) | 5 | 4.4 | 5.6 | 5.2 | 5.8 | 6.1 |

Principle

OriGene's mouse TNFR2 ELISA Kit was based on standard sandwich enzyme-linked immune-sorbent assay technology. A monoclonal antibody from rat specific for TNFR2 has been precoated onto 96-well plates. Standards(NSO, V23-G258) and test samples are added to the wells, a biotinylated detection polyclonal antibody from goat specific for TNFR2 is added subsequently and then followed by washing with PBS or TBS buffer. Avidin-Biotin-Peroxidase Complex was added and unbound conjugates were washed away with PBS or TBS buffer. HRP substrate TMB was used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the mouse TNFR2 amount of sample captured in plate.

Kit Components

| Description | Quantity | | |
|---|-----------------------|--|--|
| 96-well plate precoated with anti- mouse TNFR2 antibody | 1 | | |
| Lyophilized recombinant mouse TNFR2 standard | 10ng/tube×2 | | |
| Biotinylated anti- mouse TNFR2 antibody | 130µl(dilution 1:100) | | |
| Avidin-Biotin-Peroxidase Complex (ABC) | 130µl(dilution 1:100) | | |
| Sample diluent buffer | 30 ml | | |
| Antibody diluent buffer | 12ml | | |
| ABC diluent buffer | 12ml | | |
| TMB color developing agent | 10ml | | |
| TMB stop solution | 10ml | | |

Material Required But Not Provided

- 1. Microplate reader in standard size.
- 2. Automated plate washer.
- Adjustable pipettes and pipette tips. Multichannel pipettes are recommended in the condition of large amount of samples in the detection.
- 4. Clean tubes and Eppendorf tubes.
- 5. Washing buffer (neutral PBS or TBS).
 - ➤ Preparation of 0.01M **TBS**: Add 1.2g Tris, 8.5g Nacl; 450μl of purified acetic acid or 700μl of concentrated hydrochloric acid to 1000ml H₂O and adjust pH to 7.2-7.6. Finally, adjust the total volume to 1L.
 - ➤ Preparation of 0.01 M **PBS**: Add 8.5g sodium chloride, 1.4g Na₂HPO₄ and 0.2g NaH₂PO₄ to 1000ml distilled water and adjust pH to 7.2-7.6. Finally, adjust the total volume to 1L.

Notice for Application of Kit

- 1. To inspect the validity of experiment operation and the appropriateness of sample dilution proportion, pilot experiment using standards and a small number of samples is recommended.
- 2. The TMB Color Developing agent is colorless and transparent before using, contact us freely if it is not the case.
- 3. Before using the Kit, spin tubes and bring down all components to the bottom of tubes.
- 4. Duplicate well assay is recommended for both standard and sample testing.
- 5. Don't let 96-well plate dry, for dry plate will inactivate active components on plate.
- 6. Don't reuse tips and tubes to avoid cross contamination.
- 7. Avoid using the reagents from different batches together.
- 8. In order to avoid marginal effect of plate incubation due to temperature difference (reaction may be stronger in the marginal wells), it is suggested that the diluted ABC and TMB solution will be pre-warmed in 37°C for 30 min before using.

Preparation

1. Sample Preparation and Storage

Store samples to be assayed within 24 hours at 2-8°C. For long-term storage, aliquot and freeze samples at -20°C. Avoid repeated freeze-thaw cycles.

- > Cell culture supernates: Remove particulates by centrifugation, assay immediately or aliquot and store samples at -20°C.
- > Serum: Allow the serum to clot in a serum separator tube (about 4 hours) at room temperature. Centrifuge at approximately 1000 X g for 15 min. Analyze the serum immediately or aliquot and store samples at -20°C.

2. Sample Dilution Guideline

The user needs to estimate the concentration of the target protein in the sample and select a proper dilution factor so that the diluted target protein concentration falls near the middle of the linear regime in the standard curve. Dilute the sample using the provided diluent buffer. The following is a guideline for sample dilution. Several trials may be necessary in practice. The sample must be well mixed with the diluents buffer.

- Figh target protein concentration (10-100ng/ml). The working dilution is 1:100. i.e. Add 1μl sample into 99 μl sample diluent buffer.
- Medium target protein concentration (1-10ng/ml). The working dilution is 1:10. i.e. Add 10μl sample into 90 μl sample diluent buffer.
- > Low target protein concentration (15.6-1000pg/ml). The working dilution is 1:2. i.e. Add 50μl sample to 50 μl sample diluent buffer.
- ➤ Very Low target protein concentration ≤ 15.6pg/ml). No dilution necessary, or the working dilution is 1:2.

3. Reagent Preparation and Storage

- A. Reconstitution of the mouse TNFR2 standard: TNFR2 standard solution should be prepared no more than 2 hours prior to the experiment. Two tubes of TNFR2 standard (10ng per tube) are included in each kit. Use one tube for each experiment.
 - a. 10,000pg/ml of mouse TNFR2 standard solution: Add 1ml sample diluent buffer into one tube, keep the tube at room temperature for 10 min and mix thoroughly.
 - b. 1000pg/ml of mouse TNFR2 standard solution: Add 0.1ml of the above 10ng/ml TNFR2 standard solution into 0.9ml sample diluent buffer and mix thoroughly.
 - c. 500pg/ml→15.6pg/ml of mouse TNFR2 standard solutions: Label 6 Eppendorf tubes with 500pg/ml, 250pg/ml, 125pg/ml, 62.5pg/ml, 31.2pg/ml, 15.6pg/ml respectively. Aliquot 0.3ml of

the sample diluent buffer into each tube. Add 0.3ml of the above 1000pg/ml TNFR2 standard solution into 1st tube and mix. Transfer 0.3ml from 1st tube to 2nd tube and mix. Transfer 0.3ml from 2nd tube to 3rd tube and mix, and so on.

Note: The standard solutions are best used within 2 hours. The 10ng/ml standard solution should be stored at 4°C for up to 12 hours, or at -20°C for up to 48 hours. Avoid repeated freeze-thaw cycles.

- B. Preparation of biotinylated anti-mouse TNFR2 antibody working solution: The solution should be prepared no more than 2 hours prior to the experiment.
 - a. The total volume should be: 0.1ml/well x (the number of wells). (Allowing 0.1-0.2 ml more than total volume)
 - b. Biotinylated anti-mouse TNFR2 antibody should be diluted in 1:100 with the antibody diluent buffer and mixed thoroughly. (i.e. Add 1µl Biotinylated anti-mouse TNFR2 antibody to 99µl antibody diluent buffer.)
- C. Preparation of Avidin-Biotin-Peroxidase Complex (ABC) working solution: The solution should be prepared no more than 1 hour prior to the experiment.
 - a. The total volume should be: 0.1ml/well x (the number of wells). (Allowing 0.1-0.2 ml more than total volume)
 - b. Avidin- Biotin-Peroxidase Complex (ABC) should be diluted in 1:100 with the ABC dilution buffer and mixed thoroughly. (i.e. Add 1µl ABC to 99µl ABC diluent buffer.)

Assay Procedure

The ABC working solution and TMB color developing agent must be kept warm at 37°C for 30 min before use. When diluting samples and reagents, they must be mixed completely and evenly. Standard TNFR2 detection curve should be prepared for each experiment. The user will decide sample dilution fold by crude estimation of TNFR2 amount in samples.

- 1. Aliquot 0.1ml per well of the 1000pg/ml, 500pg/ml, 250pg/ml, 125pg/ml, 62.5pg/ml, 31.2pg/ml, 15.6pg/ml mouse TNFR2 standard solutions into the precoated 96-well plate. Add 0.1ml of the sample diluent buffer into the control well (Zero well). Add 0.1ml of each properly diluted sample of mouse cell culture supernates or serum to each empty well. See "Sample Dilution Guideline" above for details. It is recommended that each mouse TNFR2 standard solution and each sample be measured in duplicate.
- 2. Seal the plate with the cover and incubate at 37°C for 90 min.
- 3. Remove the cover, discard plate content, and blot the plate onto paper towels or other absorbent material. Do NOT let the wells completely dry at any time.
- 4. Add 0.1ml of biotinylated anti-mouse TNFR2 antibody working solution into each well and incubate the plate at 37°C for 60 min.
- 5. Wash plate 3 times with 0.01M TBS or 0.01M PBS, and each time let washing buffer stay in the wells for 1 min. Discard the washing buffer and blot the plate onto paper towels or other absorbent material. (Plate Washing Method: Discard the solution in the plate without touching the side walls. Blot the plate onto paper towels or other absorbent material. Soak each well with at least 0.3 ml PBS or TBS buffer for 1~2 minutes. Repeat this process two additional times for a total of THREE washes. Note: For automated washing, aspirate all wells and wash THREE times with PBS or TBS buffer, overfilling wells with PBS or TBS buffer. Blot the plate onto paper towels or other absorbent material.)
- 6. Add 0.1ml of prepared ABC working solution into each well and incubate the plate at 37°C for 30 min.
- 7. Wash plate 5 times with 0.01M TBS or 0.01M PBS, and each time let washing buffer stay in the wells for 1-2 min. Discard the washing buffer and blot the plate onto paper towels or other absorbent material. (See Step 5 for plate washing method).
- 8. Add 90µl of prepared TMB color developing agent into each well and incubate plate at 37°C in dark for 20-25 min (**Note:** For reference only, the optimal incubation time should be determined by end user. And

the shades of blue can be seen in the wells with the four most concentrated mouse TNFR2 standard solutions; the other wells show no obvious color).

- 9. Add 0.1ml of prepared TMB stop solution into each well. The color changes into yellow immediately.
- 10. Read the O.D. absorbance at 450nm in a microplate reader within 30 min after adding the stop solution.

For calculation, (the relative $O.D._{450}$) = (the $O.D._{450}$ of each well) – (the $O.D._{450}$ of Zero well). The standard curve can be plotted as the relative $O.D._{450}$ of each standard solution (Y) vs. the respective concentration of the standard solution (X). The mouse TNFR2 concentration of the samples can be interpolated from the standard curve. **Note:** if the samples measured were diluted, multiply the dilution factor to the concentrations from interpolation to obtain the concentration before dilution.

Summary

- 1. Add samples and standards and incubate the plate at 37°C for 90 min. Do not wash.
- 2. Add biotinylated antibodies and incubate the plate at 37°C for 60 min. Wash plate 3 times with 0.01M TBS.
- 3. Add ABC working solution and incubate the plate at 37°C for 30 min. Wash plate 5 times with 0.01M TBS.
- 4. Add TMB color developing agent and incubate the plate at 37°C in dark for 20-25 min.
- 5. Add TMB stop solution and read.

Background

Tumor necrosis factor receptor 2 (TNFR2) is one of receptors of TNF. TNF has proinflammatory and immunosuppressive properties that may segregate at the level of the 2 TNF receptors (TNFRs), TNFR1 and TNFR2. The genes for TNFR1, a 55-kDa protein, and TNFR2, a 70-kDa protein, have been mapped to human chromosomes 12 (12pter-cen) and 1 (1pter-p32), respectively. TNFR2 was induced on glomerular endothelial cells of nephritic kidneys, and TNFR2 expression on intrinsic cells, but not leukocytes, was essential for glomerulonephritis and glomerular complement deposition. TNFR1 promotes systemic immune responses and renal T cell death, while intrinsic cell TNFR2 plays a critical role in complement-dependent tissue injury. Therefore, therapeutic blockade specifically of TNFR2 may be a promising strategy in the treatment of immune-mediated glomerulonephritis.

Reference

- 1. Milatovich, A.; Song, K.; Heller, R. A.; Francke, U.: Tumor necrosis factor receptor genes, TNFR1 and TNFR2, on human chromosomes 12 and 1. Somat. Cell Molec. Genet. 17: 519-523, 1991.
- 2. Vielhauer, V.; Stavrakis, G.; Mayadas, T. N.: Renal cell-expressed TNF receptor 2, not receptor 1, is essential for the development of glomerulonephritis. J. Clin. Invest. 115: 1199-1209, 2005.