

Phone: 1.888.267.4436 Fax: 301-340-9254 Email: techsupport@origene.com Web: www.origene.com

Human CD21/CR2 ELISA Kit

Catalog No. EA100789 Size 96T(8×12 divisible strips)

For quantitative detection of human CD21 in cell culture supernates, serum and plasma(heparin, EDTA).

Typical Data Obtained from Human CD21


(TMB reaction incubate at 37°C for 20 min)

| Concentration(ng/ml) | 0.0 | 0.78 | 1.56 | 3.12 | 6.25 | 12.5 | 25 | 50 |
|----------------------|-------|-------|-------|-------|-------|-------|-------|-------|
| O.D | 0.179 | 0.230 | 0.259 | 0.323 | 0.474 | 0.817 | 1.264 | 1.825 |

Typical Human CD21 ELISA Kit Standard Curve

This standard curve was generated at OriGene for demonstration purpose only. A standard curve must be run with each assay.

Human CD21/CR2 ELISA Kit



Range 0.78ng/ml-50ng/ml

Sensitivity < 20pg/ml

Specificity Natural and recombinant human CD21

Cross-reactivity No detectable cross-reactivity with other relevant proteins

Storage

Store at 4°C for 6 months, at -20°C for 12 months. Avoid multiple freeze-thaw cycles (Shipped with wet ice.)

Precision

Intra-Assay Precision (Precision within an assay) Three samples of known concentration were tested on one plate to assess intra-assay precision.

Inter-Assay Precision (Precision between assays) Three samples of known concentration were tested in separate assays to assess inter-assay precision.

| | Intra-Assay Precision | | | Inter-Assay Precision | | |
|--------------------|-----------------------|------|------|-----------------------|------|------|
| Sample | 1 | 2 | 3 | 1 | 2 | 3 |
| n | 16 | 16 | 16 | 24 | 24 | 24 |
| Mean(ng/ml) | 9.7 | 18.4 | 34.5 | 12.2 | 21.6 | 38.8 |
| Standard deviation | 0.44 | 0.77 | 1.24 | 0.77 | 1.23 | 1.86 |
| CV(%) | 4.5 | 4.2 | 3.6 | 6.3 | 5.7 | 4.8 |

Principle

OriGene's human CD21 ELISA Kit was based on standard sandwich enzyme-linked immune-sorbent assay technology. A monoclonal antibody from mouse specific for CD21 has been precoated onto 96-well plates. Standards(CHO, Ile21-71) and test samples are added to the wells, a biotinylated detection polyclonal antibody from goat specific for CD21 is added subsequently and then followed by washing with PBS or TBS buffer. Avidin-Biotin-Peroxidase Complex was added and unbound conjugates were washed away with PBS or TBS buffer. HRP substrate TMB was used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the human CD21 amount of sample captured in plate.

Kit Components

| Description | Quantity | | |
|--|-----------------------|--|--|
| 96-well plate precoated with anti- human CD21 antibody | 1 | | |
| Lyophilized recombinant human CD21 standard | 50ng/tube×2 | | |
| Biotinylated anti- human CD21 antibody | 130µl(dilution 1:100) | | |
| Avidin-Biotin-Peroxidase Complex (ABC) | 130µl(dilution 1:100) | | |
| Sample diluent buffer | 30 ml | | |
| Antibody diluent buffer | 12ml | | |
| ABC diluent buffer | 12ml | | |
| TMB color developing agent | 10ml | | |
| TMB stop solution | 10ml | | |

Material Required But Not Provided

- 1. Microplate reader in standard size.
- 2. Automated plate washer.
- 3. Adjustable pipettes and pipette tips. Multichannel pipettes are recommended in the condition of large amount of samples in the detection.
- 4. Clean tubes and Eppendorf tubes.
- Washing buffer (neutral PBS or TBS).
 - ▶ Preparation of 0.01M **TBS**: Add 1.2g Tris, 8.5g Nacl; 450µl of purified acetic acid or 700µl of concentrated hydrochloric acid to 1000ml H₂O and adjust pH to 7.2-7.6. Finally, adjust the total volume to 1L.
 - ➤ Preparation of 0.01 M **PBS**: Add 8.5g sodium chloride, 1.4g Na₂HPO₄ and 0.2g NaH₂PO₄ to 1000ml distilled water and adjust pH to 7.2-7.6. Finally, adjust the total volume to 1L.

Notice for Application of Kit

- 1. To inspect the validity of experiment operation and the appropriateness of sample dilution proportion, pilot experiment using standards and a small number of samples is recommended.
- 2. The TMB Color Developing agent is colorless and transparent before using, contact us freely if it is not the case.
- 3. Before using the Kit, spin tubes and bring down all components to the bottom of tubes.
- 4. Duplicate well assay is recommended for both standard and sample testing.
- 5. Don't let 96-well plate dry, for dry plate will inactivate active components on plate.
- 6. Don't reuse tips and tubes to avoid cross contamination.
- 7. Avoid using the reagents from different batches together.
- 8. In order to avoid marginal effect of plate incubation due to temperature difference (reaction may be stronger in the marginal wells), it is suggested that the diluted ABC and TMB solution will be pre-warmed in 37°C for 30 m in before using.

Preparation

1. Sample Preparation and Storage

Store samples to be assayed within 24 hours at 2-8°C. For long-term storage, aliquot and freeze samples at -20°C. Avoid repeated freeze-thaw cycles.

- > Cell culture supernates: Remove particulates by centrifugation, assay immediately or aliquot and store samples at -20°C.
- > Serum: Allow the serum to clot in a serum separator tube (about 4 hours) at room temperature. Centrifuge at approximately 1000 X g for 15 min. Analyze the serum immediately or aliquot and store samples at -20°C.
- ➤ Plasma: Collect plasma using heparin or EDTA as an anticoagulant. Centrifuge for 15 min at 1500 x g within 30 min of collection. Assay immediately or aliquot and store samples at -20°C.

2. Sample Dilution Guideline

The user needs to estimate the concentration of the target protein in the sample and select a proper dilution factor so that the diluted target protein concentration falls near the middle of the linear regime in the standard curve. Dilute the sample using the provided diluent buffer. The following is a guideline for sample dilution. Several trials may be necessary in practice. The sample must be well mixed with the diluents buffer.

- Figh target protein concentration (500-5000ng/ml). The working dilution is 1:100. i.e. Add 1 μl sample into 99 μl sample diluent buffer.
- Medium target protein concentration (50-500ng/ml). The working dilution is 1:10. i.e. Add 10 μl sample into 90 μl sample diluent buffer.
- **Low target protein concentration (0.78-50ng/ml).** The working dilution is 1:2. i.e. Add 50 μl sample to 50 μl sample diluent buffer.
- > Very Low target protein concentration (≤0.78ng/ml). No dilution necessary, or the working dilution is 1:2

3. Reagent Preparation and Storage

- A. Reconstitution of the human CD21 standard: CD21 standard solution should be prepared no more than 2 hours prior to the experiment. Two tubes of CD21 standard (50ng per tube) are included in each kit. Use one tube for each experiment.
 - a. 50ng/ml of human CD21 standard solution: Add 1ml sample diluent buffer into one tube, keep the tube at room temperature for 10 min and mix thoroughly.
 - b. 25ng/ml→0.78ng/ml of human CD21 standard solutions: Label 6 Eppendorf tubes with 25ng/ml, 12.5ng/ml, 6.25ng/ml, 3.12ng/ml, 1.56ng/ml, 0.78ng/ml respectively. Aliquot 0.3ml of the sample diluent buffer into each tube. Add 0.3 ml of the above 50ng/ml CD21 standard

solution into 1st tube and mix. Transfer 0.3 ml from 1st tube to 2nd tube and mix. Transfer 0.3ml from 2nd tube to 3rd tube and mix, and so on.

Note: The standard solutions are best used within 2 hours. The 50ng/ml standard solution should be stored at 4°C for up to 12 hours, or at -20°C for up to 48 hours. Avoid repeated freeze-thaw cycles.

- B. Preparation of biotinylated anti-human CD21 antibody working solution: The solution should be prepared no more than 2 hours prior to the experiment.
 - a. The total volume should be: 0.1ml/well x (the number of wells). (Allowing 0.1-0.2 ml more than total volume)
 - b. Biotinylated anti-human CD21 antibody should be diluted in 1:100 with the antibody diluent buffer and mixed thoroughly. (i.e. Add 1µl Biotinylated anti-human CD21 antibody to 99µl antibody diluents buffer.
- C. Preparation of Avidin-Biotin-Peroxidase Complex (ABC) working solution: The solution should be prepared no more than 1 hour prior to the experiment.
 - a. The total volume should be: 0.1ml/well x (the number of wells). (Allowing 0.1-0.2 ml more than total volume)
 - b. Avidin- Biotin-Peroxidase Complex (ABC) should be diluted in 1:100 with the ABC dilution buffer and mixed thoroughly. (i.e. Add 1µl ABC to 99µl ABC diluent buffer.)

Assay Procedure

The ABC working solution and TMB color developing agent must be kept warm at 37°C for 30 min before use. When diluting samples and reagents, they must be mixed completely and evenly. Standard CD21 detection curve should be prepared for each experiment. The user will decide sample dilution fold by crude estimation of CD21 amount in samples.

- 1. Aliquot 0.1ml per well of the 50ng/ml, 25ng/ml, 12.5ng/ml, 6.25ng/ml, 3.12ng/ml, 1.56ng/ml, 0.78ng/ml human CD21 standard solutions into the precoated 96-well plate. Add 0.1ml of the sample diluent buffer into the control well (Zero well). Add 0.1ml of each properly diluted sample of human cell culture supernates, serum or plasma(heparin, EDTA) to each empty well. See "Sample Dilution Guideline" above for details. It is recommended that each human CD21 standard solution and each sample be measured in duplicate.
- 2. Seal the plate with the cover and incubate at 37°C for 90 min.
- 3. Remove the cover, discard plate content, and blot the plate onto paper towels or other absorbent material. Do NOT let the wells completely dry at any time.
- 4. Add 0.1ml of biotinylated anti-human CD21 antibody working solution into each well and incubate the plate at 37°C for 60 min.
- 5. Wash plate 3 times with 0.01M TBS or 0.01M PBS, and each time let washing buffer stay in the wells for 1 min. Discard the washing buffer and blot the plate onto paper towels or other absorbent material. (Plate Washing Method: Discard the solution in the plate without touching the side walls. Blot the plate onto paper towels or other absorbent material. Soak each well with at least 0.3 ml PBS or TBS buffer for 1~2 minutes. Repeat this process two additional times for a total of THREE washes. Note: For automated washing, aspirate all wells and wash THREE times with PBS or TBS buffer, overfilling wells with PBS or TBS buffer. Blot the plate onto paper towels or other absorbent material.)
- 6. Add 0.1ml of prepared ABC working solution into each well and incubate the plate at 37°C for 30 min.
- 7. Wash plate 5 times with 0.01M TBS or 0.01M PBS, and each time let washing buffer stay in the wells for 1-2 min. Discard the washing buffer and blot the plate onto paper towels or other absorbent material. (See Step 5 for plate washing method).
- 8. Add 90µl of prepared TMB color developing agent into each well and incubate plate at 37°C in dark for 20-25 min (**Note:** For reference only, the optimal incubation time should be determined by end user. And

the shades of blue can be seen in the wells with the four most concentrated human CD21 standard solutions; the other wells show no obvious color).

- 9. Add 0.1ml of prepared TMB stop solution into each well. The color changes into yellow immediately.
- 10. Read the O.D. absorbance at 450nm in a microplate reader within 30 min after adding the stop solution.

For calculation, (the relative $O.D._{450}$) = (the $O.D._{450}$ of each well) – (the $O.D._{450}$ of Zero well). The standard curve can be plotted as the relative $O.D._{450}$ of each standard solution (Y) vs. the respective concentration of the standard solution (X). The human CD21 concentration of the samples can be interpolated from the standard curve.

Note: if the samples measured were diluted, multiply the dilution factor to the concentrations from interpolation to obtain the concentration before dilution.

Summary

- 1. Add samples and standards and incubate the plate at 37°C for 90 min. Do not wash.
- 2. Add biotinylated antibodies and incubate the plate at 37°C for 60 min. Wash plate 3 times with 0.01M TBS.
- 3. Add ABC working solution and incubate the plate at 37°C for 30 min. Wash plate 5 times with 0.01M TBS.
- 4. Add TMB color developing agent and incubate the plate at 37°C in dark for 20-25 min.
- 5. Add TMB stop solution and read.

Background

Complement receptor type 2(CR2), also known as C3DR or CD21, is a protein that in humans is encoded by the CR2 gene. It is mapped to 1q32.2. CR2 is involved in the complement system. It binds to iC3b(inactive derivative of C3b), C3dg, or C3d. CR2 is the membrane protein on B lymphocytes to which the Epstein-Barr virus (EBV) binds during infection of these cells. CR2, together with CD19, CD81, and CD225, forms the B-cell coreceptor complex, which lowers the activation threshold of the B-cell antigen receptor. Genetic variations in this gene are associated with susceptibility to systemic lupus erythematosus type 9 (SLEB9).

Reference

- Frank K, Atkinson JP (2001). "Complement system." In Austen KF, Frank K, Atkinson JP, Cantor H. eds.Samter's Immunologic Diseases, 6th ed. Vol. 1, p. 281-298, Philadelphia: Lippincott Williams & Wilkins,
- Thiel, J., Kimmig, L., Salzer, U., Grudzien, M., Lebrecht, D., Hagena, T., Draeger, R., Volxen, N., Bergbreiter, A., Jennings, S., Gutenberger, S., Aichem, A., and 10 others.Genetic CD21 deficiency is associated with hypogammaglobulinemia. J. Allergy Clin. Immun. 129: 801-810, 2012.
- Yefenof, E., Klein, G., Jondal, M., Oldstone, M. B. A. Surface markers on human B- and T-lymphocytes.
 IX. Two color immunofluorescence studies on the association between EBV receptors and complement receptors on the surface of lymphoid cell lines. Int. J. Cancer 17: 693-700, 1976.