

## Product Information

### ***Salmonella typhi* IgG ELISA kit**

Catalog Number: EA100959

Storage Temperature: 2 – 8°C

## Instruction for Use

### **Intended Use**

The *Salmonella* IgG ELISA Kit is intended for the detection of IgG antibody to *Salmonella* in human serum or plasma.

### **Background**

*Salmonella typhi* is the causative agent of typhoid fever a contagious infection of the intestines that affects the whole body. In developing countries, typhoid often occurs in epidemics. Most people in the United States get typhoid as a result of visiting another country where the food or water supply has been contaminated. Symptoms usually start 1 to 3 weeks after exposure to the bacteria. Symptoms include: high fever, headache, sore throat, vomiting, diarrhea, skin rash and weakness. The symptoms may take 2 weeks or more to go away. Typhoid is spread when a person drinks or eats food and water contaminated by human waste (stool or urine) containing *Salmonella typhi* bacteria. A person who no longer has symptoms may still transmit the bacteria as a carrier. Testing for immunoglobulin G (IgG), IgA, and IgM antilipopolysaccharide (LPS) of *Salmonella typhi* antibodies by enzyme-linked immunosorbent assay (ELISA) showed that the levels of all three classes of immunoglobulin anti-LPS of *S. typhi* were higher in typhoid patients than in healthy or febrile nontyphoidal groups. The ELISA assay was much more sensitive and specific than any combination of the Widal test, and hence it could be a useful tool for the serologic diagnosis of typhoidal fever with a single blood sample.

### **Principle of the Test**

Diluted patient serum is added to wells coated with purified antigen. IgG specific antibody, if present, binds to the antigen. All unbound materials are washed away and the enzyme conjugate is added to bind to the antibody-antigen complex, if present. Excess enzyme conjugate is washed off and substrate is added. The plate is incubated to allow the hydrolysis of the substrate by the enzyme. The intensity of the color generated is proportional to the amount of IgG specific antibody in the sample.

### **Components**

<b>MATERIALS PROVIDED</b>	<b>96 Tests</b>
1. Microwells coated with <i>Salmonella typhi</i> antigen	12x8x1
2. Sample Diluent: 1 bottle (ready to use)	22 ml
3. Calibrator: 1 Vial (ready to use)	1ml
4. Positive Control: 1 vial (ready to use)	1ml

5. Negative Control: 1 vial (ready to use)	1ml
6. Enzyme conjugate: 1 bottle (ready to use)	12ml
7. TMB Substrate: 1 bottle (ready to use)	12ml
8. Stop Solution: 1 bottle (ready to use)	12ml
9. Wash concentrate 20X: 1 bottle	25ml

**Materials and Equipment Required but Not Provided**

1. Distilled or deionized water
2. Precision pipettes
3. Disposable pipette tips
4. ELISA reader capable of reading absorbance at 450nm
5. Absorbance paper or paper towel

**Disclaimer**

This product is for research use only and not intended for diagnostic procedures.

Control sera and sample diluent contain preserved with sodium azide. Sodium azide may react with lead and copper plumbing to form explosive metal azide. On disposal, flush with a large amount of water.

Potential hazardous materials:

The calibrator and controls contain human source components which have been tested and found non-reactive for hepatitis B surface antigen as well as HIV antibody with FDA license reagents. However, there is no test method that can offer complete assurance that HIV, Hepatitis B virus or other infectious agents are absent. These reagents should be handled at the Biosafety Level 2, as recommended in the Centers for Disease Control/National Institutes of Health manual, "Biosafety in Microbiological and Biomedical Laboratories." 1984.

**Specimen Collection and Preparation**

1. Collect blood specimens and separate the serum.
2. Specimens may be refrigerated at 2–8°C for up to seven days or frozen for up to six months. Avoid repetitive freezing and thawing.

**Reagent Preparation**

1. Prepare 1X Wash buffer by adding Wash Concentrate (25 ml, 20X) to 475 ml of distilled or deionized water
- . Store at room temperature (18-26°C).

## Assay Procedure

- Before proceeding with the assay, bring all reagents, serum references and controls to room temperature (18-26°C). Gently mix all reagents before use
  - The components in this kit are intended for use as an integral unit. The components of different lots should not be mixed
  - It is recommended that standards, control and serum samples be run in duplicate
1. Place the desired number of coated strips into the holder.
  2. Negative control, positive control, and calibrator are ready to use. Prepare 1:101 dilution of test samples, by adding 5 µl of the sample to 0.5 mL of sample diluent. Mix well.
  3. Dispense 100 µl of diluted sera, calibrator and controls into the appropriate wells. For the reagent blank, dispense 100 µl sample diluent in 1A well position. Tap the holder to remove air bubbles from the liquid and mix well. Incubate for 20 minutes at room temperature.
  4. Remove liquid from all wells. Wash wells three times with 300 µl of 1X wash buffer. Blot on absorbance paper or paper towel.
  5. Dispense 100 µl of enzyme conjugate to each well and incubate for 20 minutes at room temperature.
  6. Remove enzyme conjugate from all wells. Wash wells three times with 300 µl of 1X wash buffer. Blot on absorbance paper or paper towel.
  7. Dispense 100 µl of TMB substrate and incubate for 10 minutes at room temperature. Add 100 µl of stop solution.
  8. Read O.D. at 450 nm using ELISA reader within 15 min. A dual wavelength is recommended with reference filter of 600-650 nm

## Calculation of Results

1. Check Calibrator Factor (CF) value on the calibrator bottle. This value might vary from lot to lot. Make sure you check the value on every kit.
2. Calculate the cut-off value: Calibrator OD x Calibrator Factor (CF).
3. Calculate the Ab (Antibody) Index of each determination by dividing the O.D. value of each sample by cut-off value.

## Limitations of the Test

1. Lipemic or hemolyzed samples may cause erroneous results.

## References

1. Quiroga T; Goycoolea M; Tagle R; Gonzalez F; Rodriguez L; Villarroel L. Diagnosis of typhoid fever by two serologic methods. Enzyme-linked immunosorbent assay of antilipopolysaccharide of Salmonella typhi antibodies and Widal test. Diagn Microbiol Infect Dis 1992; 15(8):651-6.
2. Jesudason MV; Sridharan G; Arulselvan R; Babu PG; John TJ. Diagnosis of typhoid fever by the detection of anti- LPS & anti-flagellin antibodies by ELISA. Indian J Med Res 1998;107:204-7.
3. Mekara Y; Maneekarn N; Vithayasai V; Makonkawkeyoon S. Determination of antibody from typhoid patients against lipopolysaccharide and protein antigens of Salmonella typhi. Asian Pac J Allergy Immunol 1990; 8(2):95- 101.
4. Sippel JE; Hanafy HM; Diab AS; Prato C; Arroyo R. Serodiagnosis of typhoid fever in pediatric patients by anti-LPSELISA. Trans R Soc Trop Med Hyg 1987; 81(6):1022-6.
5. Vitale G; Librizzi R; Mocciaro C; Friscia I; Blandino E; Usticino V; Mansueto S; Di Fiore M; Reina G; Gambino G. An ELISA method in the diagnosis of typhoid fever. J Clin Lab Immunol 1990; 31(4):195-9.