

Product Information

Cytomegalovirus (CMV) IgM ELISA Kit

Catalog Number: EA101063

Storage Temperature: 2 – 8°C

Instruction for Use

THIS KIT IS INTENDED FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

Intended Use

The **ORIGENE Cytomegalie Virus (CMV) IgM Enzyme Immunoassay Kit** provides materials for measurement of IgM-class antibodies to Cytomegalie Virus (CMV) in human serum and plasma.

1 PRINCIPLE OF THE TEST

The **ORIGENE Cytomegalovirus (CMV) IgM ELISA Kit** is a solid phase enzyme-linked immunosorbent assay (ELISA). Specimen samples are diluted with *Sample Diluent* and additionally incubated with *IgG-RF-Sorbent* to eliminate competitive inhibition from specific IgG. This pretreatment avoids false low results. Microtiter wells as a solid phase are coated with inactivated grade 2 Cytomegalovirus (CMV) antigen (strain AD-169). **Diluted sample** specimens and **ready-for-use controls** are pipetted into these wells. During incubation Cytomegalie Virus (CMV)-specific antibodies of high specimens and controls are bound to the immobilized antigens. After a washing step to remove unbound sample and control material horseradish peroxidase conjugated anti-human IgM antibodies are dispensed into the wells. During a second incubation this anti-IgM conjugate binds specifically to IgM antibodies resulting in the formation of enzyme-linked immune complexes. After a second washing step to remove unbound conjugate the immune complexes formed (in case of high results) are detected by incubation with TMB substrate and development of a blue color. The blue color turns into yellow by stopping the enzymatic indicator reaction with sulfuric acid. The intensity of this color is directly proportional to the amount of Cytomegalie Virus (CMV)-specific IgM antibody in the sample specimen. Absorbance at 450 nm is read using an ELISA microtiter plate reader.

2 WARNINGS AND PRECAUTIONS

- For professional use only.
- Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood.
- All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
- Avoid contact with Stop Solution containing 0.2 mol/L H₂SO₄. It may cause skin irritation and burns.
- TMB substrate has an irritant effect on skin and mucosa. In case of possible contact, wash eyes with an abundant volume of water and skin with soap and abundant water. Wash contaminated objects before reusing them. If inhaled, take the person to open air.
- The microplate contains snap-off strips. Unused wells must be stored at 2 °C to 8 °C in the sealed foil pouch and used in the frame provided
- Pipetting of samples and reagents must be done as quickly as possible and in the same sequence for each step.

- Use reservoirs only for single reagents. This especially applies to the substrate reservoirs. Using a reservoir for dispensing a substrate solution that had previously been used for the conjugate solution may turn solution colored. Do not pour reagents back into vials as reagent contamination may occur.
- Mix the contents of the microplate wells thoroughly to ensure good test results. Do not reuse microwells.
- Do not let wells dry during assay; add reagents immediately after completing the rinsing steps.
- Allow the reagents to reach room temperature (21 °C – 26 °C) before starting the test. Temperature will affect the absorbance readings of the assay. However, values for the samples will not be affected.
- Never pipette by mouth and avoid contact of reagents and specimens with skin and mucous membranes.
- Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.
- Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
- Handling should be in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
- Do not use reagents beyond expiry date as shown on the kit labels.
- All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiter plate readers.
- Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
- Chemicals and prepared or used reagents have to be treated as hazardous waste according the national biohazard safety guideline or regulation.

3 REAGENTS

Reagents provided

1. **Microtiterwells**, 12 x 8 (break apart) strips, 96 wells;
Wells coated with inactivated grade 2 Cytomegalovirus (CMV) antigen (strain AD-169).
(incl. 1 strip holder and 1 cover foil)
2. **Sample Diluent ***, 1 vial, 100 mL, ready to use,
colored yellow; pH 7.2 ± 0.2 .
3. **IgG-RF-Sorbent***, 1 vial, 6.5 mL, ready to use,
colored yellow;
Contains anti-human IgG-class antibody.
4. **Standard (Standard 1-3)***, 3 vials, ready to use;
Standard 1: 2.0 mL; Standard 2-3: 1 mL each
Concentrations: **50, 200, 400 OU/mL**,
colored yellow, white caps.
5. **High Control ***, 1 vial, 1.0 mL, ready to use;
colored yellow, red cap.
6. **Low Control ***, 1 vial, 2.0 mL, ready to use;
colored yellow, yellow cap.
7. **Enzyme Conjugate ***, 1 vial, 20 mL, ready to use,
colored red,
antibody to human IgM conjugated to horseradish peroxidase.
8. **Substrate Solution**, 1 vial, 14 mL, ready to use,
Tetramethylbenzidine (TMB).
9. **Stop Solution**, 1 vial, 14 mL, ready to use,
contains 0.2 mol/L H₂SO₄,
Avoid contact with the stop solution. It may cause skin irritations and burns.
10. **Wash Solution ***, 1 vial, 30 mL (20X concentrated for 600 mL), pH 6.5 ± 0.1
see „Preparation of Reagents“.

* contain non-mercury preservative

Equipment and material required but not provided

- A microtiter plate calibrated reader (450/620nm ± 10 nm)
- Calibrated variable precision micropipettes
- Incubator 37 °C
- Manual or automatic equipment for rinsing wells
- Vortex tube mixer
- Deionised or (freshly) distilled water
- Timer
- Absorbent paper

Storage Conditions and stability of the Kit

When stored at 2 °C to 8 °C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date.

Opened reagents must be stored at 2 °C to 8 °C. Microtiter wells must be stored at 2 °C to 8 °C. Once the foil bag has been opened, care should be taken to close it tightly again.

Opened kits retain activity for two months if stored as described above.

Reagent Preparation

Allow all reagents and required number of strips to reach room temperature prior to use.

Wash Solution

Dilute *Wash Solution 1+19* (e.g. 10 mL + 190 mL) with fresh and germ free redistilled water. This diluted wash solution has a pH value of 7.2 ± 0.2 .

Consumption: ~ 5 mL per determination.

Crystals in the solution disappear by warming up to 37 °C in a water bath. Be sure that the crystals are completely dissolved before use.

The diluted Wash Solution is stable for 4 weeks at 2 °C to 8 °C.

Disposal of the Kit

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Material Safety Data Sheets.

Damaged Test Kits

In case of any severe damage to the test kit or components, ORIGENE has to be informed in writing, at the latest, one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

4 SPECIMEN COLLECTION AND PREPARATION

Serum or plasma can be used in this assay.

Do not use haemolytic, icteric or lipaemic specimens.

Specimen Collection

Serum:

Collect blood by venipuncture (e.g. Sarstedt Monovette for serum), allow to clot, and separate serum by centrifugation at room temperature. Do not centrifuge before complete clotting has occurred. Samples with anticoagulant may require increased clotting time.

Plasma:

Whole blood should be collected into centrifuge tubes containing anti-coagulant (e.g. Sarstedt Monovette with the appropriate plasma preparation) and centrifuged immediately after collection.

Specimen Storage and Preparation

Specimens should be capped and may be stored for up to 3 days at 2 °C to 8 °C prior to assaying.

Specimens held for a longer time should be frozen only once at -20 °C prior to assay. Thawed samples should be inverted several times prior to testing.

Specimen Dilution

Prior to assaying each sample specimen is first to be diluted with *Sample Diluent*. For the absorption of rheumatoid factor these prediluted samples then have to be incubated with *IgG-RF-Sorbent*

1. Dilute each sample specimen **1+50** with *Sample Diluent*;
e.g. 10 µL of specimen + 0.5 mL of *Sample Diluent*. **Mix well.**
2. Dilute this prediluted sample **1+1** with *IgG-RF-Sorbent*
e.g. 60 µL prediluted sample + 60 µL *IgG-RF-Sorbent*. **Mix well**
3. **Let stand for at least 15 minutes at room temperature, mix well or overnight at 2 °C – 8 °C and mix well again.**
4. Take 100 µL of these pretreated samples for the ELISA.

Please note: Controls are ready for use and must not be diluted!

5 ASSAY PROCEDURE**General Remarks**

- **It is very important to bring all reagents, samples and controls to room temperature before starting the test run!**
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination
- Absorbance is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.
- Close reagent vials tightly immediately after use to avoid evaporation and microbial contamination.
- To avoid cross-contamination and falsely elevated results pipette samples and dispense conjugate without splashing accurately to the bottom of wells.
- During 37°C incubation cover microtiter strips with foil to avoid evaporation.

Test Procedure

Prior to commencing the assay, dilute *Wash Solution*, **prepare samples as described in point 5.3**, mix well before pipette and establish carefully the **distribution and identification plan** supplied in the kit for all specimens and controls.

1. Select the required number of microtiter strips or wells and insert them into the holder.

Please allocate at least:

- | | | |
|---------|--------------|-------------------------------|
| 1 well | (e.g. A1) | for the substrate blank, |
| 1 well | (e.g. B1) | for the <i>Low Control</i> |
| 3 wells | (e.g. C1-E1) | for the <i>Standard 1-3</i> |
| 1 well | (e.g. F1) | for the <i>High Control</i> . |

It is left to the user to determine standards, control and samples in duplicate.

2. Dispense

- | | |
|--------------------------------------|---|
| 100 µL of <i>Low Control</i> | into well B1 |
| 100 µL of <i>Standard 1</i> | into well C1 |
| 100 µL of <i>Standard 2</i> | into well D1 |
| 100 µL of <i>Standard 3</i> | into well E1 |
| 100 µL of <i>High Control</i> | into well F1 and |
| 100 µL of each diluted sample | <u>with new disposable tips</u> into appropriate wells. |

Leave well A1 for substrate blank!

3. Cover wells with foil supplied in the kit. Incubate for **60 minutes at 37 °C**.
4. Briskly shake out the contents of the wells.
Rinse the wells 5 times with diluted *Wash Solution* (**300 µL per well**). Strike the wells sharply on absorbent paper to remove residual droplets.

Important note:

The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!

5. Dispense **100 µL** *Enzyme Conjugate* into each well, **except A1**.
6. Incubate for **30 minutes at room temperature (20 °C to 25 °C)**.
Do not expose to direct sun light!
7. Briskly shake out the contents of the wells.
Rinse the wells 5 times with diluted *Wash Solution* (300 µL per well). Strike the wells sharply on absorbent paper to remove residual droplets.
8. Add **100 µL** of *Substrate Solution* into all wells.
9. Incubate for **exactly 10 minutes at room temperature (20 °C to 25 °C) in the dark**.
10. Stop the enzymatic reaction by adding **100 µL** of *Stop Solution* to each well.
Any blue color developed during the incubation turns into yellow.
Note: Very high samples can cause dark precipitates of the chromogen!
11. Read the optical density at **450/620 nm** with a microtiter plate reader **within 30 minutes** after adding the *Stop Solution*.

Measurement

Adjust the ELISA microplate or microstrip reader **to zero** using the **substrate blank in well A1**.

If - due to technical reasons - the ELISA reader cannot be adjusted to zero using the substrate blank in well A1, subtract the absorbance value of well A1 from all other absorbance values measured in order to obtain reliable results!

Measure the absorbance of all wells **at 450 nm** and record the absorbance values for each control and sample in the distribution and identification plan.

Dual wavelength reading using 620 nm as reference wavelength is recommended.

Where applicable **calculate the mean absorbance values** of all duplicates.

6 CALCULATION OF RESULTS

Validation of the Test Run

The test run may be considered valid provided the following criteria are met:

Substrate blank in A1:	Absorbance value lower than 0.100
Low Control in B1:	Absorbance value lower than 0.200
Standard 1 (Calibrator) in C1 :	Absorbance value between 0.350 – 0.900
Standard 2 in D1 :	Absorbance value between 0.800 – 1.500
Standard 3 in E1 :	Absorbance value between 1.100 – 2.000
High Control in F1 :	Absorbance value between 0.650 – 3.000

Calculation of quantitative Results

In order to obtain **quantitative results in OU/mL** (OU = ORIGENE Units) plot the (mean) absorbance values of the *Low Control* and the 3 Standards 1, 2 and 3 on (linear/linear) graph paper in a system of coordinates against their corresponding concentrations (**0, 50, 200 and 400 OU/mL**) and draw a standard calibration curve (absorbance values on the vertical y-axis, concentrations on the horizontal x-axis).

Read results from this standard curve employing the (mean) absorbance values of each specimen and control.

All suitable computer programs available can be used for automated result reading and calculation. The following mathematical functions can be used: 4 PL (4 Parameter Logistics) curve fit, Linear regression or Point to Point calculation of the standard curve. We use ORIGENE regression program for windows (4 parameter Rodbart regression). If other regression software is used, the obtained values have to be validated by the user.

NOTE: Values of additionally (1:10, in total 1:1000) diluted specimen samples must be multiplied by the appropriate dilution factor in order to obtain correct results! (Dilution: 1:10 = Dilution factor: 10). (See chapter "5.3 Specimen Dilution").

Interpretation of quantitative Results

Normal value ranges for this ELISA should be established by each laboratory based on its own populations in the geographical areas serviced.

Calculation of qualitative Results

Absorbance value of **Standard 1 (Calibrator) = CALB**

Example: 0.4 = CALB

Interpretation of qualitative Results

LOW Mean OD sample < OD CALB –10%

GREY ZONE OD CALB -10% ≤ Mean OD sample ≤ OD CALB +10%

Repeat test 2 - 4 weeks later - with new samples

Results in the second test again in the grey zone → LOW

HIGH Mean OD sample > OD Calb +10 %

7 QUALITY CONTROL

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results.

It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results.

Please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.

After checking the above mentioned items without finding any error contact your distributor or ORIGENE directly.

Method Comparison

The ORIGENE ELISA was compared with the Virion/Serion CMV IgM ELISA. 85 serum samples are assayed.















n= 85		Diamed Eurogen ELISA	
		high	low
ORIGENE ELISA Lot 1	high	16	0
	low	0	69



7.1 Agreement: 100%

8 REFERENCES

1. Lentz, E.B. et al., Detection of Antibody to CMV Induced Early Antigens and Comparison with Four Serologic Assays and Presence of Viruria in Blood Donors, *J. of Clin. Microbiol.*, 26, 133-35 (1988)
2. Von Loon, H.M. et al., Direct Enzyme-Linked Immunosorbent Assay that uses Peroxidase-Labelled Antigens for Determination of Immunoglobulin M Antibody to CMV, *Journal of Clinical Microbiology* 13, 416-422 (1981)
3. Matas, A.J., Simmons, R.L., Fryd, D. and Jajarian, J.S., Persistent, Recurrent and Late CMV Infection, *Transplantation Proceedings*, Vol. XIII, Nr. 1, 114-124 (1981)

9 SHORT INSTRUCTIONS FOR USE

	<p>All reagents and specimens must be allowed to come to room temperature (18-25°C) before use.</p>
	<p>Leave well A1 for substrate Blank. Dispense 100 µl of Standards and Control into appropriate wells.</p>
	<p>Dispense 100 µl of sample into selected wells. (Please note special sample treatment, point 5.3!)</p>
	<p>Cover wells with foil. Incubate for 60 minutes at 37 °C.</p>
	<p>Briskly shake out the contents of the wells.</p>
	<p>Rinse the wells 5 times with diluted Wash Solution (300 µl per well).</p>
	<p>Strike the wells sharply on absorbent paper to remove residual droplets.</p>
	<p>Dispense 100 µl of Enzyme-Conjugate into each well.</p>
	<p>Incubate for 30 minutes at room temperature.</p>
	<p>Briskly shake out the contents of the wells.</p>
	<p>Rinse the wells 5 times with diluted Wash Solution (300 µl per well).</p>
	<p>Strike the wells sharply on absorbent paper to remove residual droplets.</p>
	<p>Add 100 µl of Substrate Solution to each well.</p>
	<p>Incubate for 10 minutes at room temperature.</p>

 A photograph of a microplate reader, a laboratory instrument used for measuring the absorbance of samples in a microplate.	Stop the reaction by adding 100 μ l of Stop Solution to each well.
 A photograph of a microplate, a multi-well plate used for high-throughput screening and assays.	Determine the absorbance of each well at 450 nm.

Version 2, last updated March 6, 2016