

## Product Information

### CA 15-3 ELISA Kit

Catalog Number: EA101114

Storage Temperature: 2 – 8°C

## Instruction for Use

**THIS KIT IS INTENDED FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.**

### 1 INTRODUCTION

#### 1.1 Intended Use

The **ORIGENE CA 15-3 ELISA** is an enzyme immunoassay for measurement of CA15-3 in serum and plasma.

### 2 PRINCIPLE OF THE TEST

The ORIGENE CA 15-3 ELISA Kit is a solid phase enzyme-linked immunosorbent assay (ELISA) based on the sandwich principle. The microtiter wells are coated with a mouse monoclonal antibody directed towards a unique antigenic site of the CA15-3 molecule. An aliquot of sample containing endogenous CA15-3 is incubated in the coated well with enzyme conjugate, which is an anti-CA15-3 antibody conjugated with horseradish peroxidase. After incubation the unbound conjugate is washed off. The amount of bound peroxidase is proportional to the concentration of CA15-3 in the sample. Having added the substrate solution, the intensity of colour developed is proportional to the concentration of CA15-3 in the sample.

### 3 WARNINGS AND PRECAUTIONS

1. All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
2. Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood.
3. The microplate contains snap-off strips. Unused wells must be stored at 2 °C to 8 °C in the sealed foil pouch and used in the frame provided.
4. Pipetting of samples and reagents must be done as quickly as possible and in the same sequence for each step.
5. Use reservoirs only for single reagents. This especially applies to the substrate reservoirs. Using a reservoir for dispensing a substrate solution that had previously been used for the conjugate solution may turn solution colored. Do not pour reagents back into vials as reagent contamination may occur.
6. Mix the contents of the microplate wells thoroughly to ensure good test results. Do not reuse microwells.
7. Do not let wells dry during assay; add reagents immediately after completing the rinsing steps.
8. Allow the reagents to reach room temperature (21-26°C) before starting the test. Temperature will affect the absorbance readings of the assay. However, values for the samples will not be affected.
9. Never pipet by mouth and avoid contact of reagents and specimens with skin and mucous membranes.
10. Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.

11. Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
12. Handling should be done in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
13. Do not use reagents beyond expiry date as shown on the kit labels.
14. All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiterplate readers.
15. Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
16. Avoid contact with *Stop Solution* containing 0.5 M H<sub>2</sub>SO<sub>4</sub>. It may cause skin irritation and burns.
17. Some reagents contain Proclin 300, BND and/or MIT as preservatives. In case of contact with eyes or skin, flush immediately with water.
18. TMB substrate has an irritant effect on skin and mucosa. In case of possible contact, wash eyes with an abundant volume of water and skin with soap and abundant water. Wash contaminated objects before reusing them. If inhaled, take the person to open air.
19. Chemicals and prepared or used reagents have to be treated as hazardous waste according to the national biohazard safety guideline or regulation.

#### 4 REAGENTS

##### 4.1 Reagents provided

- 1) **Microtiterwells**, 12 x 8 (break apart) strips, 96 wells;  
Wells coated with anti-CA15-3 antibody (monoclonal).
- 2) **Zero Standard**, 1 vial, 3 mL, ready to use;  
Contains non-mercury preservative
- 3) **Standard (Standard 1-4)**, 4 vials, 0.5 mL each, ready to use  
Concentrations: 25 - 50 - 100 - 200 U/mL;  
Contain non-mercury preservative.
- 4) **Control Low & High**, 2 vials, 0.5 mL each, lyophilized,  
see „Reagent Preparation“  
For control values and ranges please refer to vial label or QC-Datasheet.  
Contain non-mercury preservative.
- 5) **Assay Buffer**, 1 vial, 30 mL, ready to use:  
Contains non-mercury preservative.
- 6) **Enzyme Conjugate**, 1 vial, 14 mL, ready to use,  
Anti-CA15-3 antibody conjugated to horseradish peroxidase;  
Contains non-mercury preservative.
- 7) **Substrate Solution**, 1 vial, 14 mL, ready to use,  
Tetramethylbenzidine (TMB).
- 8) **Stop Solution**, 1 vial, 14 mL, ready to use,  
Contains 0.5 M H<sub>2</sub>SO<sub>4</sub>,  
Avoid contact with the stop solution. It may cause skin irritations and burns.
- 9) **Wash Solution**, 1 vial, 30 mL (40X concentrated),  
see „Preparation of Reagents“.

**Note:** Additional *Zero Standard* for sample dilution is available upon request.

#### 4.2 Materials required but not provided

- A microtiter plate calibrated reader ( $450 \pm 10$  nm).
- Calibrated variable precision micropipettes.
- Absorbent paper.
- Distilled or deionized water
- Timer
- Graph paper or software for data reduction

#### 4.3 Storage Conditions

When stored at 2 °C to 8 °C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date.

Opened reagents must be stored at 2 °C to 8 °C. Microtiter wells must be stored at 2 °C to 8 °C. Once the foil bag has been opened, care should be taken to close it tightly again.

Opened kits retain activity for two months if stored as described above.

#### 4.4 Reagent Preparation

Bring all reagents and required number of strips to room temperature prior to use.

#### **Control**

Reconstitute the lyophilized content with 0.5 mL distilled water and let stand for 10 minutes in minimum. Mix the controls several times before use.

**Note:** *The reconstituted controls are stable for 2 days at 2 °C to 8 °C.*

*For longer storage the reconstituted controls should be apportioned and stored at –20 °C.*

#### **Wash Solution**

Add deionized water to the 40X concentrated Wash Solution.

Dilute 30 mL of concentrated *Wash Solution* with 1170 mL deionized water to a final volume of 1200 mL.

*The diluted Wash Solution is stable for 2 weeks at room temperature.*

#### 4.5 Disposal of the Kit

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Material Safety Data Sheet.

#### 4.6 Damaged Test Kits

In case of any severe damage to the test kit or components, ORIGENE has to be informed in writing, at the latest, one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

## 5 SPECIMEN COLLECTION AND PREPARATION

Serum or plasma (EDTA, heparin or citrate plasma) can be used in this assay.

Do not use haemolytic, icteric or lipaemic specimens.

Please note: Samples containing sodium azide should not be used in the assay.

### 5.1 Specimen Collection

#### Serum:

Collect blood by venipuncture (e.g. Sarstedt Monovette # 02.1388.001), allow to clot, and separate serum by centrifugation at room temperature. Do not centrifuge before complete clotting has occurred. Samples containing anticoagulant may require increased clotting time.

#### Plasma:

Whole blood should be collected into centrifuge tubes containing anti coagulant and centrifuged immediately after collection.

(e.g. for EDTA plasma Sarstedt Monovette – red cap - # 02.166.001;  
for Heparin plasma Sarstedt Monovette – orange cap - # 02.165.001;  
for Citrate plasma Sarstedt Monovette – green cap - # 02.167.001)

### 5.2 Specimen Storage and Preparation

Specimens should be capped and may be stored for up to 48 hours at 2 °C to 8 °C prior to assaying.

Specimens held for a longer time (at least six months) should be frozen only once at -20 °C prior to assay. Thawed samples should be inverted several times prior to testing.

### 5.3 Specimen Dilution

If in an initial assay, a specimen is found to contain more than the highest standard, the specimens can be diluted with *Zero Standard* and reassayed as described in “Assay Procedure”.

For the calculation of the concentrations this dilution factor has to be taken into account.

#### Example:

- a) dilution 1:10: 10 µL sample + 90 µL *Zero Standard* (mix thoroughly)
- b) dilution 1:100: 10 µL dilution a) 1:10 + 90 µL *Zero Standard* (mix thoroughly).

## 6 ASSAY PROCEDURE

### 6.1 General Remarks

- All reagents and specimens must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination.
- Absorbance is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.

## 6.2 Test Procedure

Each run must include a standard curve.

1. Secure the desired number of Microtiter wells in the frame holder.
2. Dispense **10 µL** of each **Standard, Control** and **samples** with new disposable tips into appropriate wells.
3. Dispense **250 µL Assay Buffer** into each well.  
Thoroughly mix for 10 seconds. It is important to have a complete mixing in this step.
4. Incubate for **60 minutes** at room temperature.
5. Briskly shake out the contents of the wells.  
Rinse the wells **4 times** with diluted *Wash Solution* (400 µL per well). Strike the wells sharply on absorbent paper to remove residual droplets.

### Important note:

The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!

6. Dispense **100 µL Enzyme Conjugate** into each well.
7. Incubate for **60 minutes** at room temperature.
8. Briskly shake out the contents of the wells.  
Rinse the wells **4 times** with diluted *Wash Solution* (400 µL per well). Strike the wells sharply on absorbent paper to remove residual droplets.
9. Add **100 µL of Substrate Solution** to each well.
10. Incubate for **15 minutes** at room temperature.
11. Stop the enzymatic reaction by adding **100 µL of Stop Solution** to each well.
12. Determine the absorbance (OD) of each well at **450 ± 10 nm** with a microtiter plate reader.  
It is recommended that the wells be read **within 10 minutes** after adding the *Stop Solution*.

## 6.3 Calculation of Results

1. Calculate the average absorbance values for each set of standards, controls and samples.
2. Manual method: Using linear graph paper, construct a standard curve by plotting the mean absorbance obtained from each standard against its concentration with absorbance value on the vertical (Y) axis and concentration on the horizontal (X) axis.
3. Using the mean absorbance value for each sample determine the corresponding concentration from the standard curve.
4. Automated method: The results in the IFU have been calculated automatically using a 4 PL (4 Parameter Logistics) curve fit. 4 Parameter Logistics is the preferred method. Other data reduction functions may give slightly different results.
5. The concentration of the samples can be read directly from this standard curve. Samples with concentrations higher than that of the highest standard have to be further diluted or reported as > 200 U/mL. For the calculation of the concentrations this dilution factor has to be taken into account.

### 6.3.1 Example of Typical Standard Curve

The following data is for demonstration only and **cannot** be used in place of data generations at the time of assay.

Standard	Optical Units (450 nm)
Standard 0 (0 U/mL)	0.02
Standard 1 (25 U/mL)	0.45
Standard 2 (50 U/mL)	0.82
Standard 3 (100 U/mL)	1.43
Standard 4 (200 U/mL)	2.03

## 7 QUALITY CONTROL

- Good laboratory practice requires that controls be run with each calibration curve. A statistically significant number of controls should be assayed to establish mean values and acceptable ranges to assure proper performance.
- It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results. It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results.
- Employ appropriate statistical methods for analysing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials results should be considered invalid.
- In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.
- After checking the above mentioned items without finding any error contact your distributor or ORIGENE directly.

## 8 LEGAL ASPECTS

### 8.1 Liability

Any modification of the test kit and/or exchange or mixture of any components of different lots from one test kit to another could negatively affect the intended results and validity of the overall test. Such modification and/or exchanges invalidate any claim for replacement.

Claims submitted due to customer misinterpretation of laboratory results are also invalid. Regardless, in the event of any claim, the manufacturer's liability is not to exceed the value of the test kit. Any damage caused to the test kit during transportation is not subject to the liability of the manufacturer.

**9 REFERENCES / LITERATURE**

1. Duffy M.J., Evoy D., and McDermott E.W.: CA 15-3: uses and limitation as a biomarker for breast cancer. *Clin. Chim. Acta* 2010; 411(23-24):1869-74.
2. Hilkens, J., et al.: Monoclonal antibodies against human milkfat globule membranes useful in carcinoma research. *Prot. Biol. Fluids* 1984;31:1013-1016.
3. Sekine, H., et al.: Purification and characterization of a high molecular weight glycoprotein detectable in human milk and breast carcinomas. *J. Immunol.* 1985;135(5):3610-3615.
4. Fujino, N., et al.: Clinical evaluation of an immunoradiometric assay for CA 15-3 antigen associated with human mammary carcinomas: Comparison with carcinoembryonic antigen. *Jpn. J. Clin. Oncol.* 1986;46:335-345.
5. Colomer, R., et al.: CA 15-3 Early results of a new breast cancer marker. *Anticancer Research* 1986;6:683-684.
6. Harris L. et al.: American Society of Clinical Oncology 2007 update of recommendations for the use of tumor markers in breast cancer. *J. Clin. Oncol* 2007; 25(33):5287-312.
7. Chittoor S, Swain S: Adjuvant Therapy in Early Breast Cancer. *Am Fam Physician* 1991;44:453-462.
8. Duffy M.J.: Serum tumor markers in breast cancer: are they of clinical value? *Clin. Chem.* 2006; 52(3):345-51.
9. Ugrinska A., et al.: Circulating tumor markers and nuclear medicine imaging modalities: breast, prostate and ovarian cancer. *Q.J. Nucl. Med.* 2002; 46(2):88-104.
10. Hayes, D.F., et al.: Comparison of circulating CA 15-3 and carcinoembryonic antigen levels in patients with breast cancer. *J. Clin. Oncol.* 1986;4(10):1542-1550.
11. Pons-Anticet, D.F.M., et al.: Value of CA 15-3 in the follow-up of breast cancer patients. *Brit. J. Cancer* 1987;55(5):567-569.
12. Ruibal A, Genollá J, Rosell M, Gris JM, Colomer R: Serum CA 15.3 levels in patients with non-tumoral diseases, and establishment of a threshold for tumoral activity. Results in 1219 patients. *Int J Biol Markers* 1(3):159-60, 1986.
13. Colomer R, Ruibal A, Navarro M, Encabo G, Sole LA, Salvador L: Circulating CA 15.3 levels in breast cancer. Our present experience. *Int J Biol Markers* 1(2):89-92, 1986.

Version 2, last updated March 6, 2016