

Product Information

DHT-optimized ELISA Kit

Catalog Number: EA101127

Storage Temperature: 2 – 8°C

Instruction for Use

THIS KIT IS INTENDED FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

1 INTRODUCTION

1.1 Intended Use

The **ORIGENE DHT-optimized ELISA** is an enzyme immunoassay for the quantitative measurement of total 5 α -Dihydrotestosterone (DHT) in human serum or plasma (EDTA, Li-heparin- or citrate plasma).

2 PRINCIPLE OF THE TEST

The ORIGENE DHT-optimized ELISA Kit is a solid phase enzyme-linked immunosorbent assay (ELISA), based on the principle of competitive binding. The microtiter wells are coated with a polyclonal antibody directed towards an antigenic site of the DHT molecule. During the first incubation, the DHT in the added sample competes with the added enzyme conjugate, which is DHT-horseradish peroxidase conjugate, for binding to the coated antibody. After incubation the unbound conjugate is washed off. The amount of bound peroxidase conjugate is inversely proportional to the concentration of DHT in the sample. After addition of the substrate solution, the intensity of color developed is inversely proportional to the concentration of DHT in the sample.

3 WARNINGS AND PRECAUTIONS

1. This kit is for research use only. For professional use only.
2. All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
3. Before starting the assay, read the instructions completely and carefully. Use the valid version of instructions for use provided with the kit. Be sure that everything is understood.
4. The microplate contains snap-off strips. Unused wells must be stored at 2 °C - 8 °C in the sealed foil pouch and used in the frame provided.
5. Pipetting of samples and reagents must be done as quickly as possible and in the same sequence for each step.
6. Use reservoirs only for single reagents. This especially applies to the substrate reservoirs. Using a reservoir for dispensing a substrate solution that had previously been used for the conjugate solution may turn solution colored. Do not pour reagents back into vials as reagent contamination may occur.
7. Mix the contents of the microplate wells thoroughly to ensure good test results. Do not reuse microwells.
8. Do not let wells dry during assay; add reagents immediately after completing the rinsing steps.
9. Allow the reagents to reach room temperature (20 °C - 26 °C) before starting the test. Temperature will affect the absorbance readings of the assay. However, values for the samples will not be affected.
10. Never pipet by mouth and avoid contact of reagents and specimens with skin and mucous membranes.

11. Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.
12. Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
13. Handling should be done in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
14. Do not use reagents beyond expiry date as shown on the kit labels.
15. All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiterplate readers.
16. Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
17. Avoid contact with *Stop Solution* containing 0.5 M H₂SO₄. It may cause skin irritation and burns.
18. Some reagents contain Proclin 300, BND and/or MIT as preservatives. In case of contact with eyes or skin, flush immediately with water.
19. TMB substrate has an irritant effect on skin and mucosa. In case of possible contact, wash eyes with an abundant volume of water and skin with soap and abundant water. Wash contaminated objects before reusing them. If inhaled, take the person to open air.
20. Chemicals and prepared or used reagents have to be treated as hazardous waste according to the national biohazard safety guideline or regulation.

4 REAGENTS

4.1 Reagents provided

1. **Microtiterwells**, 12 x 8 (break apart) strips, 96 wells;
Wells coated with anti-DHT antibody (polyclonal).
2. **Sealing Film**, 1 sheet;
clear plastic to cover wells during 37 °C incubation
3. **Zero Standard**, 1 vial, 3 mL, ready to use.
0 pg/mL;
Contains non-mercury preservative.
4. **Standard (Standard 1-5)**, 5 vials, 1 mL, ready to use,
Concentrations: 25; 100; 250; 625; 1500 pg/mL
Contain non-mercury preservative.
5. **Control Low & High**, 2 vials, 1 mL each, ready to use,
For control values and ranges please refer to vial label or QC-Datasheet.
Contain non-mercury preservative.
6. **Conjugate Diluent**, 1 vial, 4 mL, ready to use,
Contains non-mercury preservative.
7. **Enzyme Conjugate 10X concentrate**, 1 vial, 0.5 mL,
DHT conjugated to horseradish peroxidase;
See "Reagent Preparation".
Contains non-mercury preservative.
8. **Substrate Solution**, 1 vial, 14 mL, ready to use, keep away from direct sunlight.
Tetramethylbenzidine (TMB).
9. **Stop Solution**, 1 vial, 14 mL, ready to use,
contains 0.5 M H₂SO₄,
Avoid contact with the stop solution. It may cause skin irritations and burns.

10. **Wash Solution**, 1 vial, 30 mL (40X concentrated),
See “Reagent Preparation”.

4.2 Materials required but not provided

- A microtiter plate calibrated reader (450 nm ± 10 nm)
- Calibrated variable precision micropipettes.
- Incubator at 37 °C
- Absorbent paper.
- Distilled or deionized water
- Timer
- Semi logarithmic graph paper or software for data reduction

4.3 Storage Conditions

When stored at 2 °C - 8 °C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date.

Opened reagents must be stored at 2 °C - 8 °C. Microtiter wells must be stored at 2 °C - 8 °C. Once the foil bag has been opened, care should be taken to close it tightly again.

Opened kits retain activity for 2 months if stored as described above.

4.4 Reagent Preparation

Bring all reagents and required number of strips to room temperature prior to use.

Wash Solution

Add deionized water to the 40X concentrated *Wash Solution*.

Dilute 30 mL of concentrated *Wash Solution* with 1170 mL deionized water to a final volume of 1200 mL.

The diluted Wash Solution is stable for 2 weeks at room temperature.

Enzyme Conjugate

Dilute *Enzyme Conjugate* concentrate 1:10 in *Conjugate Diluent*.

Stability of the prepared Enzyme-Conjugate: 1 week at 2 °C - 8 °C in a sealed container.

Example:

If the whole plate is used, dilute 0.36 mL *Enzyme Conjugate 10X conc.* with 3.24 mL *Conjugate Diluent* to a total volume of 3.6 mL.

If the whole plate is not used at once, prepare the required quantity of *Enzyme Conjugate* by mixing *Enzyme Conjugate 10X conc.* with *Conjugate Diluent* as shown in the table:

No. of strips	Enzyme Conjugate 10X conc. (μL)	Conjugate Diluent (mL)
1	30	0.27
2	60	0.54
4	120	1.08
6	180	1.62
8	240	2.16
10	300	2.70
12	360	3.24

4.5 Disposal of the Kit

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Material Safety Data Sheet.

4.6 Damaged Test Kits

In case of any severe damage to the test kit or components, ORIGENE has to be informed in writing, at the latest, one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

5 SAMPLE COLLECTION AND PREPARATION

Human serum or plasma (EDAT, lithium heparin- or citrate plasma) can be used in this assay.

Do not use hemolytic, icteric or lipemic specimens.

Please note: Samples containing sodium azide should not be used in the assay.

5.1 Sample Collection

Serum:

Collect blood by venipuncture (e.g. Sarstedt Monovette for serum), allow to clot, and separate serum by centrifugation at room temperature. Do not centrifuge before complete clotting has occurred. Samples containing anticoagulant may require increased clotting time.

Plasma:

Whole blood should be collected into centrifuge tubes containing anti-coagulant (e.g. Sarstedt Monovette with the appropriate plasma preparation) and centrifuged immediately after collection.

Whole blood should not be frozen before centrifugation.

5.2 Sample Storage

Samples must be stored tightly capped and may be stored for up to 5 days at 2 °C - 8 °C prior to assaying.

Samples held for a longer time should be frozen only once at -20 °C (for up to 12 months) prior to assaying. If stored frozen, freeze only once. Thawed samples should be inverted several times prior to testing.

5.3 Sample Dilution

If in an initial assay, a sample is found to contain more than the highest standard, the specimens can be diluted with *Zero Standard* and re-assayed as described in Assay Procedure.

For the calculation of the concentrations this dilution factor has to be taken into account.

Example:

dilution 1:2: 75 μ L sample + 75 μ L *Zero Standard* (mix thoroughly)

We recommend to dilute the samples not more than 1:4.

6 ASSAY PROCEDURE

6.1 General Remarks

- All reagents and specimens must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination.
- Absorbance is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.

6.2 Test Procedure

Each run must include a standard curve. The controls serve as internal controls for the reliability of the test procedure. They must be assayed with each test run.

1. Secure the desired number of Microtiter wells in the frame holder.
2. Dispense **75 μ L** of each **Standard, Control** and **sample** with new disposable tips into appropriate wells.
3. Dispense **25 μ L Enzyme Conjugate** into each well.
Thoroughly mix for 10 seconds. It is important to have a complete mixing in this step.
4. Cover the wells with sealing film or equivalent.
Incubate for **60 minutes at 37 °C**.
5. Wash the wells as follows:
If the wash step is performed manually:
 Briskly shake out the contents of the wells.
 Rinse the wells **3 times** with diluted *Wash Solution* (300 μ L per well).
If an automated plate washer is used:
 Rinse the wells **3 times** with diluted *Wash Solution* (400 μ L per well).

At the of the washing step, always strike the wells sharply on absorbent paper to remove residual droplets.

Important note:

The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!

6. Add **100 μ L** of **Substrate Solution** to each well.
7. Incubate for **15 (\pm 5) minutes** at room temperature.
8. Stop the enzymatic reaction by adding **100 μ L** of **Stop Solution** to each well.
9. Determine the absorbance (OD) of each well at **450 \pm 10 nm** with a microtiter plate reader.
It is recommended that the wells be read **within 10 minutes** after adding the *Stop Solution*.

6.3 Calculation of Results

1. Calculate the average absorbance values for each set of standards, controls and samples.
2. Using semi-logarithmic graph paper, construct a standard curve by plotting the mean absorbance obtained from each standard against its concentration with absorbance value on the vertical (Y) axis and concentration on the horizontal (X) axis.
3. Using the mean absorbance value for each sample determine the corresponding concentration from the standard curve.
4. Automated method: The results in the Instructions for Use have been calculated automatically using a 4-Parameter curve fit. (4 Parameter Rodbard or 4 Parameter Marquardt are the preferred methods.) Other data reduction functions may give slightly different results.
5. The concentration of the samples can be read directly from this standard curve. Samples with concentrations higher than that of the highest standard have to be further diluted or reported as > 1500 pg/mL. For the calculation of the concentrations this dilution factor has to be taken into account.

6.3.1 Example of Typical Standard Curve

The following data is for demonstration only and **cannot** be used in place of data generations at the time of assay.

Standard	Optical Units (450 nm)
Zero Standard (0 pg/mL)	2.409
Standard 1 (25 pg/mL)	1.939
Standard 2 (100 pg/mL)	1.364
Standard 3 (250 pg/mL)	0.869
Standard 4 (625 pg/mL)	0.449
Standard 5 (1500 pg/mL)	0.189

7 QUALITY CONTROL

Good laboratory practice requires that controls be run with each calibration curve. A statistically significant number of controls should be assayed to establish mean values and acceptable ranges to assure proper performance.

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results.

The controls and the corresponding results of the QC-Laboratory are stated in the QC certificate added to the kit. The values and ranges stated on the QC sheet always refer to the current kit lot and should be used for direct comparison of the results.

It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results.

Employ appropriate statistical methods for analyzing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials results should be considered invalid.

In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.

After checking the above mentioned items without finding any error contact your distributor or ORIGENE directly.

8 ASSAY CHARACTERISTICS

8.1 Assay Dynamic Range

The range of the assay is between 6.0 - 1500 pg/mL.

8.2 Cross Reactivity

The following substances were tested for cross reactivity of the assay:

Substance	Concentration tested (ng/ml)	Mean Cross Reactivity (%)
17-OH-Progesterone	20	0.0
Aldosterone	1	0.0
Androstenedione	10	0.67
Corticosterone	1	0.05
Danazol	1	0.0
DHEA	30	0.09
DHEA-S	10000	0.01
Estradiol	2	0.0
Estriol	40	0.44
Estrone	2.4	0.0
Ethisterone	1	0.0
Progesterone	40	0.0
Testosterone	8.5	3.22

9 LIMITATIONS OF USE

Reliable and reproducible results will be obtained when the assay procedure is performed with a complete understanding of the package insert instruction and with adherence to good laboratory practice.

Any improper handling of samples or modification of this test might influence the results.

9.1 Interfering Substances

Hemoglobin (up to 4 mg/mL), Bilirubin (up to 0.5 mg/mL) and Triglyceride (up to 30 mg/mL) have no influence on the assay results.

9.2 Drug Interferences

Until today no substances (drugs) are known to us, which have an influence to the measurement of DHT in a sample.

9.3 High-Dose-Hook Effect

No hook effect was observed in this test.

10 LEGAL ASPECTS

10.1 Reliability of Results

The test must be performed exactly as per the manufacturer's instructions for use. Moreover the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable national standards and/or laws. This is especially relevant for the use of control reagents. It is important to always include, within the test procedure, a sufficient number of controls for validating the accuracy and precision of the test.

The test results are valid only if all controls are within the specified ranges and if all other test parameters are also within the given assay specifications. In case of any doubt or concern please contact ORIGENE.

10.2 Liability

Any modification of the test kit and/or exchange or mixture of any components of different lots from one test kit to another could negatively affect the intended results and validity of the overall test. Such modification and/or exchanges invalidate any claim for replacement.

Claims submitted due to customer misinterpretation of laboratory results are also invalid. In the event of any claim, the manufacturer's liability is not to exceed the value of the test kit. Any damage caused to the test kit during transportation is not subject to the liability of the manufacturer.

11 REFERENCES / LITERATURE

1. Pang S, Riddick L: Hirsutism. In Lifshitz F: Pediatric Endocrinology, A Clinical Guide, Second Edition. Marcel Dekker, New York, 1990, 259-291.
2. Mostaghel EA. Steroid hormone synthetic pathways in prostate cancer. *Transl Androl Urol.* 2013 Sep;2(3):212-227.
3. Askew EB et al. Modulation of androgen receptor activation function 2 by testosterone and dihydrotestosterone. *J Biol Chem.* 2007 Aug 31;282(35):25801-16.
4. Shanbhag VP, Södergård R. The temperature dependence of the binding of 5 alpha-dihydrotestosterone, testosterone and estradiol to the sex hormone globulin (SHBG) of human plasma. *J Steroid Biochem.* 1986 Feb;24(2):549-55.
5. Kalogera E et al. Androgen glucuronides analysis by liquid chromatography tandem-mass spectrometry: could it raise new perspectives in the diagnostic field of hormone-dependent malignancies? *J Chromatogr B Analyt Technol Biomed Life Sci.* 2013 Dec 1;940:24-34.
6. Rothman MS et al. Reexamination of testosterone, dihydrotestosterone, estradiol and estrone levels across the menstrual cycle and in postmenopausal women measured by liquid chromatography-tandem mass spectrometry. *Steroids.* 2011 Jan;76(1-2):177-82.
7. Imperato-McGinley J, Zhu YS. Androgens and male physiology the syndrome of 5alpha-reductase-2 deficiency. *Mol Cell Endocrinol.* 2002 Dec 30;198(1-2):51-9.
8. Flores E et al. Steroid 5alpha-reductase inhibitors. *Mini Rev Med Chem.* 2003 May;3(3):225-37.
9. Wilson Jd. Syndromes of androgen resistance. 1992 *Biol Reprod* 46:168-173.
10. Azziz R, Carmina E, Sawaya ME. Idiopathic hirsutism. *Endocr Rev.* 2000 Aug;21(4):347-62.
11. Münzker J et al. Testosterone to dihydrotestosterone ratio as a new biomarker for an adverse metabolic phenotype in the polycystic ovary syndrome. *J Clin Endocrinol Metab.* 2015 Feb;100(2):653-60

Version 3, last updated September 30, 2022