



# Murine IL-2 ELISA Kit

Instructions for use

Catalogue numbers:                   1x48 tests: EA101684  
  1x96 tests: EA101685  
  2x96 tests: EA101686

**For research use only**

Fast Track Your Research.....

## Table of Contents

1. Intended Use .....	2
2. Principle Of The Method .....	2
3. Reagents Provided And Reconstitution .....	2
4. Materials Required But Not Provided.....	3
5. Storage Instructions .....	3
6. Specimen Collection, Processing & Storage .....	3
7. Safety & Precautions For Use .....	4
8. Assay Preparation .....	5
8.1. Assay Design.....	5
8.2. Preparation Of Wash Buffer.....	5
8.3. Preparation Of Standard Diluent Buffer .....	5
8.4. Preparation Of Standard.....	6
8.5. Preparation Of Biotinylated Anti-mIL-2 .....	6
8.6. Preparation Of Streptavidin-HRP .....	6
9. Method.....	7
10. Data Analysis.....	8
11. Assay Limitations.....	8
12. Performance Characteristics .....	9
12.1. Sensitivity .....	9
12.2. Precision.....	9
12.3. Dilution Parallelism .....	10
12.4. Recovery .....	11
12.5. Stability .....	11
12.6. Specificity .....	11
13. References .....	12
14. Assay Summary.....	13

# MurineIL-2 ELISA KIT

## 1. Intended use

The OriGeneMurine IL-2 ELISA is to be used for the in-vitro quantitative determination of murine interleukin-2 (mIL-2) in murine serum, buffered solutions or cell culture medium. The assay will recognize both natural and recombinant murine IL-2.

**This kit has been configured for research use only.**

## 2. Principle of the method

The murine IL-2 Kit is a solid phase sandwich Enzyme Linked-Immuno-Sorbent Assay (ELISA). A monoclonal antibody specific for mIL-2 has been coated onto the wells of the microtiter strips provided. Samples including standards of known concentrations of mIL-2 concentrations and unknowns are pipetted into these wells.

During the first incubation, the antigen and a biotinylated monoclonal antibody specific for mIL-2 are simultaneously incubated.

After washing, the enzyme (streptavidin-peroxydase) is added. After incubation and washing to remove all the unbound enzyme, a substrate solution which is acting to the bound enzyme is added to induce a coloured reaction product. The intensity of this coloured product is directly proportional to the concentration of mIL-2 present in the samples.

## 3. Reagents provided and reconstitution

Reagents (Store@2-8°C)	Quantity 1x48 well kit Cat no. EA101684	Quantity 1x96 well kit Cat no. EA101685	Quantity 2x96 well kit Cat no. EA101686	Reconstitution
96 well microtiter strip plate	1/2	1	2	Ready to use (Pre-coated)
Plastic plate covers	2	2	4	n/a
Standard mIL-2 500 pg/ml	1	2	4	Reconstitute as directed on the vial (see Assay preparation, section 8)
Standard Diluent (Buffer)	1 (25ml)	1 (25ml)	1 (25ml)	10x Concentrate, dilute in distilled water (see reagent preparation, section 8)
Biotinylated anti-mIL-2	1 (0.4ml)	1 (0.4ml)	2 (0.4ml)	Dilute in Biotinylated Antibody Diluent (see Assay preparation, section 8)
Biotinylated Antibody Diluent	1 (7.5ml)	1 (7.5ml)	1 (13ml)	Ready to use
Streptavidin-HRP	1 (5µl)	2 (5µl)	4 (5µl)	Add 0.5ml of HRP Diluent prior to use (see Assay preparation, section 8)
HRP Diluent	1 (23ml)	1 (23ml)	1 (23ml)	Ready to use
Wash Buffer	1 (10ml)	1 (10ml)	2 (10ml)	200x Concentrate dilute in distilled water (see Assay preparation, section 8)
TMB Substrate	1 (11ml)	1 (11ml)	1 (24ml)	Ready to use
H <sub>2</sub> SO <sub>4</sub> stop reagent	1 (11ml)	1 (11ml)	2 (11ml)	Ready to use

## 4. Materials required but not provided

- Microtiter plate reader fitted with appropriate filters (450nm required with optional 630nm reference filter)
- Microplate washer or wash bottle
- 10, 50, 100, 200 and 1,000µl adjustable single channel micropipettes with disposable tips
- 50-300µl multi-channel micropipette with disposable tips
- Multichannel micropipette reagent reservoirs
- Distilled water
- Vortex mixer
- Miscellaneous laboratory plastic and/or glass, if possible sterile

## 5. Storage Instructions

Store kit reagents between 2 and 8°C. Immediately after use remaining reagents should be returned to cold storage (2-8°C). Expiry of the kit and reagents is stated on box front labels. The expiry of the kit components can only be guaranteed if the components are stored properly, and if, in case of repeated use of one component, the reagent is not contaminated by the first handling.

**Wash Buffer:** Once prepared store at 2-8° C for up to 1 week

**Standard Diluent Buffer:** Once prepared store at 2-8° C for up to 1 week

**Standards :** Once prepared use immediately and do not store

**Biotinylated Secondary Antibody:** Once prepared use immediately and do not store

**Streptavidin-HRP:** Once prepared use immediately and do not store

## 6. Specimen collection, processing & storage

Cell culture supernatants, serum, plasma or other biological samples will be suitable for use in the assay. Remove serum from the clot or red cells, respectively, as soon as possible after clotting and separation.

**Cell culture supernatants:** Remove particulates and aggregates by spinning at approximately 1000 x g for 10 min.

**Storage:** If not analysed shortly after collection, samples should be aliquoted (250-500µl) to avoid repeated freeze-thaw cycles and stored frozen at -70°C. Avoid multiple freeze-thaw cycles of frozen specimens.

**Recommendation:** Do not thaw by heating at 37°C or 56°C. Thaw at room temperature and make sure that sample is completely thawed and homogeneous before use. When possible avoid use of badly haemolysed or lipemic sera. If large amounts of particles are present these should be removed prior to use by centrifugation or filtration.

## 7. Safety & precautions for use

- Handling of reagents, serum or plasma specimens should be in accordance with local safety procedures, e.g. CDC/NIH Health manual : " Biosafety in Microbiological and Biomedical Laboratories" 1984
- Laboratory gloves should be worn at all times
- Avoid any skin contact with H<sub>2</sub>SO<sub>4</sub> and TMB. In case of contact, wash thoroughly with water
- Do not eat, drink, smoke or apply cosmetics where kit reagents are used
- Do not pipette by mouth
- When not in use, kit components should be stored refrigerated or frozen as indicated on vials or bottles labels
- All reagents should be warmed to room temperature before use. Lyophilized standards should be discarded after use
- Once the desired number of strips has been removed, immediately reseal the bag to protect the remaining strips from deterioration
- Cover or cap all reagents when not in use
- Do not mix or interchange reagents between different lots
- Do not use reagents beyond the expiration date of the kit
- Use a clean disposable plastic pipette tip for each reagent, standard, or specimen addition in order to avoid cross contamination, for the dispensing of H<sub>2</sub>SO<sub>4</sub> and substrate solution, avoid pipettes with metal parts
- Use a clean plastic container to prepare the washing solution
- Thoroughly mix the reagents and samples before use by agitation or swirling
- All residual washing liquid must be drained from the wells by efficient aspiration or by decantation followed by tapping the plate forcefully on absorbent paper. Never insert absorbent paper directly into the wells
- The TMB solution is light sensitive. Avoid prolonged exposure to light. Also, avoid contact of the TMB solution with metal to prevent colour development. Warning TMB is toxic avoid direct contact with hands. Dispose off properly
- If a dark blue colour develops within a few minutes after preparation, this indicates that the TMB solution has been contaminated and must be discarded. Read absorbance's within 1 hour after completion of the assay
- When pipetting reagents, maintain a consistent order of addition from well-to-well. This will ensure equal incubation times for all wells
- Follow incubation times described in the assay procedure
- Dispense the TMB solution within 15 min of the washing of the microtitre plate

## 8. Assay Preparation

Bring all reagents to room temperature before use

### 8.1. Assay Design

Determine the number of microwell strips required to test the desired number of samples plus appropriate number of wells needed for running zeros and standards. Each sample, standard and zero should be tested **in duplicate**. Remove sufficient Microwell Strips for testing from the pouch immediately prior to use. Return any wells not required for this assay with desiccant to the pouch. Seal tightly and return to 2-8°C storage.

**Example plate layout**(example shown for a 6 point standard curve)

	Standards		Sample Wells									
	1	2	3	4	5	6	7	8	9	10	11	12
A	500	500										
B	250	250										
C	125	125										
D	62.5	62.5										
E	31.25	31.25										
F	15.6	15.6										
G	Blank	Blank										
H												

*All remaining empty wells can be used to test samples in duplicate*

### 8.2. Preparation of Wash Buffer

Dilute the (200x) Wash Buffer concentrate 200 fold with distilled water to give a 1x working solution. Pour entire contents (10 ml) of the Washing Buffer Concentrate into a clean 2,000 ml graduated cylinder. Bring final volume to 2,000 ml with glass-distilled or deionized water. Mix gently to avoid foaming. Transfer to a clean wash bottle and store at 2°-8°C for up to 1 week.

### 8.3. Preparation of Standard Diluent Buffer

Add the contents of the vial (10x concentrate) to 225ml of distilled water before use.

This solution can be stored at 2-8°C for up to 1 week.

#### 8.4. Preparation of Standard

Standard vials must be reconstituted with the volume of Standard Diluent shown on the vial immediately prior to use. This reconstitution gives a stock solution of 500 pg/ml of mIL-2. **Mix the reconstituted standard gently by inversion only.** Serial dilutions of the standard are made directly in the assay plate to provide the concentration range from 500 to 15.6pg/ml. A fresh standard curve should be produced for each new assay.

- Immediately after reconstitution add 200µl of the reconstituted standard to wells A1 and A2, which provides the highest concentration standard at 500pg/ml
- Add 100µl of Standard Diluent to the remaining standard wells B1 and B2 to F1 and F2
- Transfer 100µl from wells A1 and A2 to B1 and B2. Mix the well contents by repeated aspirations and ejections taking care not to scratch the inner surface of the wells
- Continue this 1:1 dilution using 100µl from wells B1 and B2 through to wells F1 and F2 providing a serial diluted standard curve ranging from 500pg/ml to 15.6pg/ml
- Discard 100µl from the final wells of the standard curve (F1 and F2)

Alternatively these dilutions can be performed in separate clean tubes and immediately transferred directly into the relevant wells.

#### 8.5. Preparation of Biotinylated anti-mIL-2

It is recommended this reagent is prepared immediately before use. Dilute the biotinylated anti-mIL-2 with the Biotinylated Antibody Diluent in an appropriate clean glass vial using volumes appropriate to the number of required wells. Please see example volumes below:

Number of wells required	Biotinylated Antibody (µl)	Biotinylated Antibody Diluent (µl)
16	40	1060
24	60	1590
32	80	2120
48	120	3180
96	240	6360

#### 8.6. Preparation of Streptavidin-HRP

It is recommended to centrifuge vial for a few seconds in a microcentrifuge to collect all the volume at the bottom.

Dilute the 5µl vial with 0.5ml of HRP diluent **immediately before use.** Do-not keep this diluted vial for future experiments. Further dilute the HRP solution to volumes appropriate for the number of required wells in a clean glass vial. Please see example volumes below:

Number of wells required	Streptavidin-HRP (µl)	Streptavidin-HRP Diluent (ml)
16	30	2
24	45	3
32	60	4
48	75	5
96	150	10

## 9. Method

We strongly recommend that every vial is mixed thoroughly without foaming prior to use except the standard vial which must be mixed gently by inversion only.

Prepare all reagents as shown in section 8.

**Note:** Final preparation of Biotinylated anti-mIL-2 (section 8.5) and Streptavidin-HRP (section 8.6) should occur immediately before use.

Assay Step		Details
1.	Addition	<b>Prepare Standard curve</b> as shown in section 8.4
2.	Addition	Add 100µl of each <b>standard, sample and zero</b> in duplicate to appropriate number of wells
3.	Addition	Add 50µl of diluted <b>biotinylated anti-mIL-2</b> to all wells
4.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for <b>3 hours</b>
5.	Wash	Remove the cover and wash the plate as follows: a) Aspirate the liquid from each well b) Dispense 0.3 ml of <b>1x washing solution</b> into each well c) Aspirate the contents of each well d) Repeat step b and c another two times
6.	Addition	Add 100µl of <b>Streptavidin-HRP</b> solution into all wells
7.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for <b>30 min</b>
8.	Wash	Repeat wash step 5.
9.	Addition	Add 100µl of ready-to-use <b>TMB Substrate Solution</b> into all wells
10.	Incubation	Incubate in the dark for <b>10-20 minutes*</b> at room temperature. Avoid direct exposure to light by wrapping the plate in aluminium foil
11.	Addition	Add 100µl of <b>H<sub>2</sub>SO<sub>4</sub>:Stop Reagent</b> into all wells
<p><b>Read the absorbance</b> value of each well (immediately after step 11.) on a spectrophotometer using 450 nm as the primary wavelength and optionally 630 nm as the reference wave length (610 nm to 650 nm is acceptable).</p>		

*\*Incubation time of the substrate solution is usually determined by the ELISA reader performance. Many ELISA readers only record absorbance up to 2.0 O.D. Therefore the colour development within individual microwells must be observed by the analyst, and the substrate reaction stopped before positive wells are no longer within recordable range*



## 10. Data Analysis


Calculate the average absorbance values for each set of duplicate standards and samples. Ideally duplicates should be within 20% of the mean.

Generate a linear standard curve by plotting the average absorbance of each standard on the vertical axis versus the corresponding murine IL-2 standard concentration on the horizontal axis.

The amount of mIL-2 in each sample is determined by extrapolating OD values against mIL-2 standard concentrations using the standard curve.

**Example mIL-2 Standard curve**

Standard	IL-2 Concentration (pg/ml)	OD (450nm) mean	CV (%)
1	500	2.510	0.2
2	250	1.572	3.4
3	125	0.925	2.8
4	62.5	0.494	0.16
5	31.25	0.256	7.2
6	15.6	0.151	2.1
Zero	0	0.025	



**Note;** curve shown above should not be used to determine results. Every laboratory must produce a standard curve for each set of microwell strips assayed.

## 11. Assay limitations

Do not extrapolate the standard curve beyond the maximum standard curve point. The dose-response is non-linear in this region and good accuracy is difficult to obtain. Concentrated samples above the maximum standard concentration must be diluted with Standard diluent or with your own sample buffer to produce an OD value within the range of the standard curve. Following analysis of such samples always multiply results by the appropriate dilution factor to produce actual final concentration.

The influence of various drugs on end results has not been investigated. Bacterial or fungal contamination and laboratory cross-contamination may also cause irregular results.

Improper or insufficient washing at any stage of the procedure will result in either false positive or false negative results. Completely empty wells before dispensing fresh Washing Buffer, fill with Washing Buffer as indicated for each wash cycle and do not allow wells to sit uncovered or dry for extended periods.

Disposable pipette tips, flasks or glassware are preferred, reusable glassware must be washed and thoroughly rinsed of all detergents before use.

As with most biological assays conditions may vary from assay to assay therefore **a fresh standard curve must be prepared and run for every assay.**

## 12. Performance Characteristics

### 12.1. Sensitivity

The minimum detectable dose of murine IL-2 was determined to be **3.4 pg/ml**. This was determined by adding 2 standard deviations to the mean OD obtained when the zero standard was assayed 36 times (in 6 different experiments).

### 12.2. Precision

#### Intra-Assay

Reproducibility within the assay was evaluated in three independent experiments. Each assay was carried out with 6 replicates (3 duplicates) in 2 murine pooled serum, 2 in culture media and 2 in standard diluent with samples containing different concentrations of mIL-2.

**The overall intra-assay coefficient of variation has been calculated to be 3.4%.**

	Sample	Observed value 1 (pg/ml)	Observed value 2 (pg/ml)	Observed value 3 (pg/ml)	Mean	SD	CV%
Session 1	Sample 1	390.5	373.5	365.3	376.4	12.9	3.4%
	Sample 2	136.4	139.4	127.5	134.4	6.2	4.6%
	Sample 3	460.4	465.2	431.6	452.4	18.2	4.0%
	Sample 4	176.6	163.9	175.2	172.9	8.1	4.7%
	Sample 5	436.2	408.4	391.6	412.1	22.5	5.5%
	Sample 6	162.8	150.1	150.2	154.4	7.3	4.7%
Session 2	Sample 1	353.8	360.3	356.7	356.9	3.2	0.9%
	Sample 2	138.6	138.9	127.0	134.9	6.8	5.1%
	Sample 3	365.5	367.1	379.0	370.5	7.4	2.0%
	Sample 4	148.4	149.3	146.6	148.1	1.4	0.9%
	Sample 5	336.5	330.3	380.4	342.4	15.9	4.6%
	Sample 6	134.0	135.3	131.2	133.5	2.1	1.6%
Session 3	Sample 1	386.5	354.7	371.9	371.9	15.9	4.3%
	Sample 2	140.4	144.4	144.4	134.8	4.8	3.4%
	Sample 3	406.5	426.6	426.6	429.1	12.9	2.9%
	Sample 4	167.0	167.0	167.0	161.0	3.5	2.1%
	Sample 5	370.8	375.7	375.7	398.8	14.9	3.9%
	Sample 6	152.6	149.6	149.6	158.4	4.5	2.9%
						<b>Mean</b>	<b>3.4%</b>

### Inter-Assay

Assay to assay reproducibility within one laboratory was evaluated in three independent experiments. Each assay was carried out with 6 replicates (3 duplicates) in 2 murine pooled serum, 2 in culture media and 2 in standard diluent with samples containing different concentrations of mL-2.

**The calculated overall coefficient of variation was 6.8%.**

	Observed values	Sample 1	Sample 2	Sample 3	Sample 4	Sample 6	Sample 7
Session 1	Observed value 1	390.5	136.4	460.4	179.6	436.2	162.8
	Observed value 1	373.5	139.4	465.2	163.9	408.4	150.1
	Observed value 1	365.3	127.5	431.6	175.2	391.6	150.2
Session 2	Observed value 1	353.8	138.6	365.5	148.4	336.5	134.0
	Observed value 1	360.3	138.9	367.1	149.3	330.3	135.4
	Observed value 1	356.7	127.0	379.0	146.6	360.4	131.2
Session 3	Observed value 1	386.5	140.4	406.5	167.0	370.8	152.6
	Observed value 1	354.7	144.4	426.6	167.0	375.7	149.6
	Observed value 1	371.9	138.8	429.1	161.0	398.8	158.4
Mean		368.13	136.6	414.6	162.0	378.7	147.1
SD		13.5	5.8	37.6	11.8	34.7	11.15
%CV		<b>3.7</b>	<b>4.3</b>	<b>9.1</b>	<b>7.3</b>	<b>9.0</b>	<b>7.8</b>
		<b>Mean</b>		<b>6.8%</b>			

### 12.3. Dilution Parallelism

Spiked murine serum and spiked culture media with different levels of murine IL-2 were analysed at serial two fold dilutions (1:2-1:4) with two replicates each. Diluted specimens demonstrated good linearity compared to the expected concentrations: recoveries ranged from 85% to 123% with an **overall mean recovery of 93%**.

Experiment	Sample	Dilution	Expected value (pg/ml)	Observed value (pg/ml)	Recovery %
Experiment 1	Sample 1	Neat		306.5	
		1/2	153.2	137.6	90%
		1/4	68.8	63.5	92%
	Sample 2	Neat		299.7	
		1/2	148.9	141.5	95%
		1/4	70.7	60.3	85%
	Sample 3	Neat		404.3	
		1/2	202.2	189.8	94%
		1/4	94.9	84.4	89%
Experiment 2	Sample 1	Neat		261.0	
		1/2	130.5	160.0	123%
		1/4	80	68.8	86%
	Sample 2	Neat		274.6	
		1/2	137.3	122.5	89%
		1/4	61.3	58.8	96%
	Sample 3	Neat		427.4	
		1/2	213.7	193.1	90%
		1/4	96.6	86.9	90%
				<b>Mean</b>	<b>93%</b>

## 12.4. Recovery

The spike recovery was evaluated by spiking several concentrations (High/Medium/Low) of murine IL-2 in murine serum and culture media in three experiments.

Recoveries ranged from 76% to 141% with an **overall mean recovery of 103%**.

Sample	Expected concentration (pg/ml)	Observed concentration (pg/ml)	n	Recovery %
Sample 1	500	378.2	3	76%
Sample 2	250	235.7	3	94%
Sample 3	500	576.5	3	115%
Sample 4	150	141.5	3	94%
Sample 5	75	72.3	3	96%
Sample 6	75	106.1	3	141%
			<b>Mean</b>	<b>103%</b>

## 12.5. Stability

### Storage Temperature

Aliquots of spiked serum or culture media samples were stored at  $-20^{\circ}\text{C}$ ,  $2-8^{\circ}\text{C}$ , room temperature (RT) and at  $37^{\circ}\text{C}$  and the mIL-2 level was determined after 24h.

For culture media samples, we observed no significant loss.

For spiked serum, we observed a significant loss of 45% when stored at  $37^{\circ}\text{C}$ .

### Freeze-Thaw stability

Aliquots of spiked serum or culture media were stored frozen at  $-20^{\circ}\text{C}$  and thawed up to 5 times and mIL-2 level was determined.

No significant loss is observed.

## 12.6. Specificity

The assay recognizes natural and recombinant murine IL-2. To define specificity of this ELISA, several proteins were tested for cross reactivity. There was no cross reactivity observed for any protein tested (murine IL-6, murine IL-10, murine GM-CSF, murine IFN $\gamma$ , murine TNF $\alpha$ , murine IL-17A, murine CD138, murine IL-27, murine IL-23 and human IL-2) assayed at concentration from 5 to 100ng/ml

### 13. References

Garcia-Perez, M. A. et al. Alterations in the phenotype and function of immune cells in ovariectomy-induced osteopenic mice. *Hum. Reprod.*, 2006; 21(4):880-7

Li Xiangming et al. A novel HBV DNA vaccine based on T cell epitopes and its potential therapeutic effect in HBVtransgenic mice. *Int. Immunol.*, 2005; 17(10):1293-302

Zhu, M. et al. Enhancement of DNA vaccine potency against herpes simplex virus 1 by co-administration of an interleukin-18 expression plasmid as a genetic adjuvant. *J Med Microbiol.*, 2003; 52(Pt3):223-8

## 14. Assay Summary

Total procedure length : 3h45mn

Add 100µl sample and diluted standard and 50µl Biotinylated anti-mIL-2

↓

Incubate 3 hours at room temperature

↓

Wash three times

↓

Add 100µl of Streptavidin-HRP

↓

Incubate 30min at room temperature

↓

Wash three times

↓

Add 100 µl of ready-to-use TMB  
Protect from light. Let the color develop for 10-20 min.

↓

Add 100 µl of H<sub>2</sub>SO<sub>4</sub>

↓

Read Absorbance at 450 nm

### TECHNICAL CONSULTATION

OriGene Technologies, Inc.  
9620 Medical Center Dr., Suite 200  
Rockville, MD 20850

Phone: 1.888.267.4436  
Fax: 301-340-9254  
Email: techsupport@OriGene.com  
Web: www.OriGene.com

**For Research Use Only**  
**Not for use in diagnostic procedures**