



# Human MCP-1 ELISA Kit

Instructions for use

Catalogue numbers:                   1x48 tests: EA101854  
  1x96 tests: EA101855  
  2x96 tests: EA101856

**For research use only**

Fast Track Your Research.....

## Table of Contents

1. Intended use .....	2
2. Introduction .....	2
2.1. Summary .....	2
2.2. Principle of the Method .....	2
3. Reagents provided and reconstitution .....	3
4. Materials required but not provided .....	3
5. Storage Instructions .....	3
6. Specimen collection, processing & storage .....	4
7. Safety & precautions for use .....	5
8. Assay Preparation .....	6
8.1. Assay Design .....	6
8.2. Preparation of Wash Buffer .....	6
8.3. Preparation of Standard Diluent Buffer .....	6
8.4. Preparation of Standard .....	7
8.5. Preparation of Samples .....	7
8.6. Preparation of Biotinylated anti-MCP-1 .....	7
8.7. Preparation of Streptavidin-HRP .....	8
9. Method .....	9
10. Data Analysis .....	10
11. Assay limitations .....	10
12. Performance Characteristics .....	10
12.1. Sensitivity .....	11
12.2. Precision .....	11
12.3. Dilution Parallelism .....	11
12.4. Spike Recovery .....	12
12.5. Specificity .....	12
12.6. Stability .....	12
12.7. Expected values .....	12
13. Bibliography .....	13
14. Assay Summary .....	14

# Human MCP-1 ELISA KIT

## 1. Intended use

The MCP-1 ELISA is an enzyme-linked immunosorbent assay for the quantitative detection of monocyte chemoattractant protein-1 levels in cell culture supernatants, human serum, plasma, amniotic fluid, or other body fluids.

**This kit has been configured for research use only.**

## 2. Introduction

### 2.1. Summary

Chemokines (chemoattractant cytokines) represent a superfamily of small secreted proteins that function as intercellular messengers to control migration and activation of leukocytes involved in inflammatory reactions and immunity (reviewed in 1,2). Furthermore, chemokines are important mediators of many pathologies such as allergic responses, chronic inflammatory and autoimmune diseases, tumor growth and hematopoietic development (3,4).

The CC-chemokine monocyte chemoattractant protein 1 (MCP-1), also known as monocyte chemoattractant and activating factor (MCAF) was characterized as a monocyte-specific chemoattractant that was later shown to attract also T lymphocytes and NK cells (1, 2,5,6,7).

The human mature MCP-1 protein consists of 76 amino acids, derived by cleavage of a hydrophobic signal peptide from the 99 aa precursor protein. MCP-1 is mainly expressed by macrophages in response to a wide range of cytokines such as IL-6, TNF- $\alpha$  and IL-1 $\beta$ , but can, upon stimulation, also be produced by a variety of cells and tissues, such as fibroblasts, endothelial cells or certain tumor cells (5,8).

Because of its target cell specificity, MCP-1 was postulated to play a pathogenic role in a variety of diseases characterized by mononuclear cell infiltration, including atherosclerosis (9), rheumatoid arthritis and allergic responses. (6,7). Elevated levels of MCP-1 have also been found in connection with osseous inflammation (10) and Alzheimer's disease (AD) (11) as well as Myocardial Ischemia (12) and viral infections (13,14).

In acute and chronic-active multiple sclerosis (MS) lesions immunoreactivity for MCP-1 was increased whereas MCP-1 was found to be significantly reduced in cerebrospinal fluids (CSF) and chronic lesions of patients with MS (15,16). In basophils, MCP-1 is highly effective as a stimulus of histamine release but has only weak chemotactic activity. Additionally it has been shown to chemoattract CD4+ and CD8+ T lymphocytes, and expression of the MCP-1 chemokine may affect HIV infection via signaling through the CCR2 receptor (17).

### 2.2. Principle of the Method

A capture Antibody highly specific for MCP-1 has been coated to the wells of the microtiter strip plate provided during manufacture. Binding of MCP-1 in samples and known standards to the capture antibodies is completed and then any excess unbound analyte is removed. During the next incubation period the binding of the biotinylated anti-MCP-1 secondary antibody to the analyte occurs. Any excess unbound secondary antibody is then removed. The HRP conjugate solution is then added to every well including the zero wells, following incubation excess conjugate is removed by careful washing. A chromogen substrate is added to the wells resulting in the progressive development of a blue coloured complex with the conjugate. The colour development is then stopped by the addition of acid turning the resultant final product yellow. The intensity of the produced coloured complex is directly proportional to the concentration of MCP-1 present in the samples and standards. The absorbance of the colour complex is then measured and the generated OD values for each standard are plotted against expected concentration forming a standard curve. This standard curve can then be used to accurately determine the concentration of MCP-1 in any sample tested.

### 3. Reagents provided and reconstitution

Reagents (Store @ 2-8°C)	Quantity 1x48 well kit Cat no. EA101854	Quantity 1x96 well kit Cat no. EA101855	Quantity 2x96 well kit Cat no. EA101856	Reconstitution
96 well microtiter strip plate	1/2	1	2	Ready to use (Pre-coated)
Plastic plate covers	2	2	4	n/a
Standard: 1000pg/ml	1	2	4	Reconstitute as directed on the vial (see Assay preparation, section 8)
Standard Diluent (Buffer)	1 (25ml)	1 (25ml)	1 (25ml)	10x Concentrate, dilute in distilled water (see reagent preparation, section 8)
Biotinylated anti-MCP-1	1 (0.4ml)	1 (0.4ml)	2 (0.4ml)	Dilute in Biotinylated Antibody Diluent (see Assay preparation, section 8)
Biotinylated Antibody Diluent	1 (7ml)	1 (7ml)	1 (13ml)	Ready to use
Streptavidin-HRP	1 (5µl)	2 (5µl)	4 (5µl)	Add 0.5ml of HRP diluent prior to use (see Assay preparation, section 8)
HRP Diluent	1 (23ml)	1 (23ml)	1 (23ml)	Ready to use
Wash Buffer	1 (10ml)	1 (10ml)	2 (10ml)	200x Concentrate dilute in distilled water (see Assay preparation, section 8)
TMB Substrate	1 (11ml)	1 (11ml)	1 (24ml)	Ready to use
H <sub>2</sub> SO <sub>4</sub> stop reagent	1 (11ml)	1 (11ml)	2 (11ml)	Ready to use

### 4. Materials required but not provided

- Microtiter plate reader fitted with appropriate filters (450nm required with optional 630nm reference filter)
- Microplate washer or wash bottle
- 10, 50, 100, 200 and 1,000µl adjustable single channel micropipettes with disposable tips
- 50-300µl multi-channel micropipette with disposable tips
- Multichannel micropipette reagent reservoirs
- Distilled water
- Vortex mixer
- Miscellaneous laboratory plastic and/or glass, if possible sterile

### 5. Storage Instructions

Store kit reagents between 2 and 8°C. Immediately after use remaining reagents should be returned to cold storage (2-8°C). Expiry of the kit and reagents is stated on box front labels. The expiry of the kit components can only be guaranteed if the components are stored properly, and if, in case of repeated use of one component, the reagent is not contaminated by the first handling.

**Wash Buffer:** Once prepared store at 2-8° C for up to 1 week

**Standard Diluent Buffer:** Once prepared store at 2-8° C for up to 1 week

**Standards :** Once prepared use immediately and do not store

**Biotinylated Secondary Antibody:** Once prepared use immediately and do not store

**Streptavidin-HRP:** Once prepared use immediately and do not store

## 6. Specimen collection, processing & storage

Cell culture supernatants, serum, plasma or other biological samples will be suitable for use in the assay. Remove serum from the clot or red cells, respectively, as soon as possible after clotting and separation.

**Cell culture supernatants:** Remove particulates and aggregates by spinning at approximately 1000 x g for 10 min.

**Serum:** Avoid any unintentional stimulation of the cells by the procedure. Use pyrogen/endotoxin free collecting tubes. Serum should be removed rapidly and carefully from the red cells after clotting. For that, after clotting, centrifuge at approximately 1000 x g for 10 min and remove serum.

**Plasma:** EDTA, citrate and heparin plasma can be assayed. Spin samples at 1000 x g for 30 min to remove particulates. Harvest plasma.

**Storage:** If not analyzed shortly after collection, samples should be aliquoted (250-500µl) to avoid repeated freeze-thaw cycles and stored frozen at -70°C. Avoid multiple freeze-thaw cycles of frozen specimens.

**Recommendation:** Do not thaw by heating at 37°C or 56°C. Thaw at room temperature and make sure that sample is completely thawed and homogeneous before use. When possible avoid use of badly haemolysed or lipemic sera. If large amounts of particles are present these should be removed prior to use by centrifugation or filtration.

## 7. Safety & precautions for use

- Handling of reagents, serum or plasma specimens should be in accordance with local safety procedures , e.g.CDC/NIH Health manual : " Biosafety in Microbiological and Biomedical Laboratories" 1984
- Laboratory gloves should be worn at all times
- Avoid any skin contact with H<sub>2</sub>SO<sub>4</sub> and TMB. In case of contact, wash thoroughly with water
- Do not eat, drink, smoke or apply cosmetics where kit reagents are used
- Do not pipette by mouth
- When not in use, kit components should be stored refrigerated or frozen as indicated on vials or bottles labels
- All reagents should be warmed to room temperature before use. Lyophilized standards should be discarded after use
- Once the desired number of strips has been removed, immediately reseal the bag to protect the remaining strips from deterioration
- Cover or cap all reagents when not in use
- Do not mix or interchange reagents between different lots
- Do not use reagents beyond the expiration date of the kit
- Use a clean disposable plastic pipette tip for each reagent, standard, or specimen addition in order to avoid cross contamination, for the dispensing of H<sub>2</sub>SO<sub>4</sub> and substrate solution, avoid pipettes with metal parts
- Use a clean plastic container to prepare the washing solution
- Thoroughly mix the reagents and samples before use by agitation or swirling
- All residual washing liquid must be drained from the wells by efficient aspiration or by decantation followed by tapping the plate forcefully on absorbent paper. Never insert absorbent paper directly into the wells
- The TMB solution is light sensitive. Avoid prolonged exposure to light. Also, avoid contact of the TMB solution with metal to prevent colour development. Warning TMB is toxic avoid direct contact with hands. Dispose off properly
- If a dark blue colour develops within a few minutes after preparation, this indicates that the TMB solution has been contaminated and must be discarded. Read absorbance's within 1 hour after completion of the assay
- When pipetting reagents, maintain a consistent order of addition from well-to-well. This will ensure equal incubation times for all wells
- Follow incubation times described in the assay procedure
- Dispense the TMB solution within 15 min of the washing of the microtitre plate

## 8. Assay Preparation

Bring all reagents to room temperature before use

### 8.1. Assay Design

Determine the number of microwell strips required to test the desired number of samples plus appropriate number of wells needed for running zeros and standards. Each sample, standard and zero should be tested **in duplicate**. Remove sufficient Microwell Strips for testing from the pouch immediately prior to use. Return any wells not required for this assay with desiccant to the pouch. Seal tightly and return to 2-8°C storage.

**Example plate layout**(example shown for a 6 point standard curve)

	Standards (pg/mL)		Sample Wells									
	1	2	3	4	5	6	7	8	9	10	11	12
A	1000	1000										
B	500	500										
C	250	250										
D	125	125										
E	62.5	62.5										
F	31.25	31.25										
G	zero	zero										
H												

*All remaining empty wells can be used to test samples in duplicate*

### 8.2. Preparation of Wash Buffer

Dilute the (200x) wash buffer concentrate 200 fold with distilled water to give a 1x working solution. Pour entire contents (10 ml) of the Washing Buffer Concentrate into a clean 2,000 ml graduated cylinder. Bring final volume to 2,000 ml with glass-distilled or deionized water. Mix gently to avoid foaming. Transfer to a clean wash bottle and store at 2°-8°C for up to 1 week.

### 8.3. Preparation of Standard Diluent Buffer

Add the contents of the vial (10x concentrate) to 225ml of distilled water before use.

This solution can be stored at 2-8°C for up to 1 week.

#### 8.4. Preparation of Standard

Standard vials must be reconstituted with the volume of standard diluent shown on the vial immediately prior to use. This reconstitution gives a stock solution of 1000 pg/ml of MCP-1. **Mix the reconstituted standard gently by inversion only.** Serial dilutions of the standard are made directly in the assay plate to provide the concentration range from 1000 to 31.25 pg/ml. A fresh standard curve should be produced for each new assay.

- Immediately after reconstitution add 200µl of the reconstituted standard to wells A1 and A2, which provides the highest concentration standard at 1000 pg/ml
- Add 100µl of Standard Diluent to the remaining standard wells B1 and B2 to F1 and F2
- Transfer 100µl from wells A1 and A2 to B1 and B2. Mix the well contents by repeated aspirations and ejections taking care not to scratch the inner surface of the wells
- Continue this 1:1 dilution using 100µl from wells B1 and B2 through to wells F1 and F2 providing a serial diluted standard curve ranging from 1000 pg/ml to 31.25 pg/ml
- Discard 100µl from the final wells of the standard curve (F1 and F2)

Alternatively these dilutions can be performed in separate clean tubes and immediately transferred directly into the relevant wells.

#### 8.5. Preparation of Samples

Before testing, human serum or plasma samples have to be diluted 1:2 in Standard Buffer Diluent

#### 8.6. Preparation of Biotinylated anti-MCP-1

It is recommended this reagent is prepared immediately before use. Dilute the biotinylated anti-MCP-1 with the biotinylated antibody diluent in an appropriate clean glass vial using volumes appropriate to the number of required wells. Please see example volumes below:

Number of wells required	Biotinylated Antibody (µl)	Biotinylated Antibody Diluent (µl)
16	40	1060
24	60	1590
32	80	2120
48	120	3180
96	240	6360



## 8.7. Preparation of Streptavidin-HRP

It is recommended to centrifuge vial for a few seconds in a microcentrifuge to collect all the volume at the bottom.

Dilute the 5 $\mu$ l vial with 0.5ml of HRP diluent **immediately before use**. Do-not keep this diluted vial for future experiments. Further dilute the HRP solution to volumes appropriate for the number of required wells in a clean glass vial. Please see example volumes below:

Number of wells required	Streptavidin-HRP ( $\mu$ l)	Streptavidin-HRP Diluent (ml)
16	30	2
24	45	3
32	60	4
48	75	5
96	150	10

## 9. Method

We strongly recommend that every vial is mixed thoroughly without foaming prior to use except the standard vial which must be mixed gently by inversion only.

Prepare all reagents as shown in section 8.

**Note:** Final preparation of Biotinylated anti-MCP-1 (section 8.6) and Streptavidin-HRP (section 8.7) should occur immediately before use.

Assay Step		Details
1.	Addition	<b>Prepare Standard curve</b> as shown in section 8.4
2.	Addition	Add 100µl of each <b>standard, sample and zero</b> in duplicate to appropriate number of wells
3.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for <b>1 hour</b>
4.	Wash	Remove the cover and wash the plate as follows: a) Aspirate the liquid from each well b) Dispense 0.3 ml of <b>1x washing solution</b> into each well c) Aspirate the contents of each well d) Repeat step b and c another two times
5.	Addition	Add 50µl of diluted <b>biotinylated anti-MCP-1</b> to all wells
6.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for <b>30 min</b>
7.	Wash	Repeat wash step 4.
8.	Addition	Add 100µl of <b>Streptavidin-HRP</b> solution into all wells
9.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for <b>30 min</b>
10.	Wash	Repeat wash step 4.
11.	Addition	Add 100µl of ready-to-use <b>TMB Substrate Solution</b> into all wells
12.	Incubation	Incubate in the dark for <b>10-20 minutes*</b> at room temperature. Avoid direct exposure to light by wrapping the plate in aluminium foil.
13.	Addition	Add 100µl of <b>H<sub>2</sub>SO<sub>4</sub>:Stop Reagent</b> into all wells
<b>Read the absorbance</b> value of each well (immediately after step 13.) on a spectrophotometer using 450 nm as the primary wavelength and optionally 630 nm as the reference wave length (610 nm to 650 nm is acceptable).		

*\*Incubation time of the substrate solution is usually determined by the ELISA reader performance. Many ELISA readers only record absorbance up to 2.0 O.D. Therefore the colour development within individual microwells must be observed by the analyst, and the substrate reaction stopped before positive wells are no longer within recordable range*

## 10. Data Analysis


Calculate the average absorbance values for each set of duplicate standards and samples. Ideally duplicates should be within 20% of the mean.

Generate a linear standard curve by plotting the average absorbance of each standard on the vertical axis versus the corresponding Human MCP-1 standard concentration on the horizontal axis.

The amount of MCP-1 in each sample is determined by extrapolating OD values against MCP-1 standard concentrations using the standard curve.

**Example MCP-1 Standard curve**

Standard	MCP-1 Conc	OD (450nm) mean	CV (%)
1	1000	2.669	1.0
2	500	1.358	0.9
3	250	0.709	4.0
4	125	0.334	7.0
5	62.5	0.136	7.3
6	31.25	0.067	10.6
Zero	0	0.039	-



**Note:** curve shown above should not be used to determine results. Every laboratory must produce a standard curve for each set of microwell strips assayed.

For samples human serum or plasmas which have been diluted 1:2 according to the protocol, the calculated concentration should be multiplied by the dilution factor (x2)

## 11. Assay limitations

Do not extrapolate the standard curve beyond the maximum standard curve point. The dose-response is non-linear in this region and good accuracy is difficult to obtain. Concentrated samples above the maximum standard concentration must be diluted with Standard diluent or with your own sample buffer to produce an OD value within the range of the standard curve. Following analysis of such samples always multiply results by the appropriate dilution factor to produce actual final concentration.

The influence of various drugs on end results has not been investigated. Bacterial or fungal contamination and laboratory cross-contamination may also cause irregular results.

Improper or insufficient washing at any stage of the procedure will result in either false positive or false negative results. Completely empty wells before dispensing fresh Washing Buffer, fill with Washing Buffer as indicated for each wash cycle and do not allow wells to sit uncovered or dry for extended periods.

Disposable pipette tips, flasks or glassware are preferred, reusable glassware must be washed and thoroughly rinsed of all detergents before use.

As with most biological assays conditions may vary from assay to assay therefore **a fresh standard curve must be prepared and run for every assay.**

## 12. Performance Characteristics

## 12.1. Sensitivity

The sensitivity, minimum detectable dose of MCP-1 using this OriGeneMCP-1 ELISA kit was found to be <5.8pg/ml. This was determined by adding 2 standard deviations to the mean OD obtained when the zero standard was assayed 40 times.

## 12.2. Precision

### Intra-assay

Reproducibility within the assay was evaluated in three independent experiments. Each assay was carried out with 6 replicates (3 duplicates) in 2 human pooled serum, 2 in culture media and 2 in standard diluent with samples containing different concentrations of MCP-1. 2 standard curves were run on each plate **The overall intra-assay coefficient of variation has been calculated to be 1.8% .**

Session	Sample	Mean MCP-1 pg/ml	SD	CV
Session 1	Sample 1	649.05	22.33	3.44
	Sample 2	186.76	3.68	1.97
	Sample 3	644.75	7.03	1.09
	Sample 4	202.85	2.93	1.44
	Sample 5	504.84	8.71	1.72
	Sample 6	152.98	2.14	1.40
Session 2	Sample 1	645.27	5.43	0.84
	Sample 2	201.14	3.13	1.55
	Sample 3	641.30	4.36	0.68
	Sample 4	209.78	6.34	3.02
	Sample 5	518.12	11.37	2.2
	Sample 6	159.19	5.49	3.45
Session 3	Sample 1	666.18	5.97	0.9
	Sample 2	221.83	3.87	1.74
	Sample 3	649.02	19.01	2.93
	Sample 4	232.77	1.56	0.67
	Sample 5	521.16	9.91	1.9
	Sample 6	167.73	1.12	0.67

### Inter-assay

Assay to assay reproducibility within one laboratory was evaluated in three independent experiments by two technicians. Each assay was carried out with 6 replicates (3 duplicates) in 2 human pooled serum, 2 in culture media and 2 in standard diluent with samples containing different concentrations of MCP-1. 2 standard curves were run on each plate. **The calculated overall coefficient of variation was 3.7%.**

	Sample 1	Sample 2	Sample 3	Sample 4	Sample 5	Sample 6
Mean MCP-1 pg/ml	650	201	644	208	536	163
SD	17	7	16	11	28	6
CV	2.7	3.2	2.5	5.1	5.3	3.6

## 12.3. Dilution Parallelism

Three spiked human serum with different levels of recombinant MCP-1 were analysed at three serial two fold dilutions (1:2-1:16) with two replicates each. Recoveries ranged from 61% to 118% with an overall **mean recovery of 95%**.

#### **12.4. Spike Recovery**

The spike recovery was evaluated by spiking three concentrations of recombinant MCP-1 in human serum and culture media in three experiments. Recoveries ranged from 95% to 133% with an overall **mean recovery of 114%**.

#### **12.5. Specificity**

The assay recognizes natural human MCP-1. To define specificity of this ELISA, several proteins were tested for cross reactivity. There was no cross reactivity observed for any protein tested (GM-CSF, IFN $\alpha$ 2a, IL-2, IL-6, IL-1 $\beta$ , TNF $\alpha$ , IFN $\gamma$ , IL-8, IL-12 and IL-17A).

#### **12.6. Stability**

##### **Storage Stability**

Aliquots of spiked serum sample and culture media were stored at  $-20^{\circ}\text{C}$ ,  $2-8^{\circ}\text{C}$ , room temperature (RT) and at  $37^{\circ}\text{C}$  and the MCP-1 level determined after 24h. There was no significant loss of MCP-1 at  $2-8^{\circ}\text{C}$ , 8% loss at RT and 35% after 24h at  $37^{\circ}\text{C}$ , when spiked in serum. As a result, we would advise that samples are stored at either  $-20^{\circ}\text{C}$  or  $2-8^{\circ}\text{C}$ . When MCP-1 is spiked in culture media, no significant loss is observed.

##### **Freeze-thaw Stability**

Aliquots of spiked serum were stored frozen at  $-20^{\circ}\text{C}$  and thawed up to 5 times and MCP-1 level was determined. There was no significant loss of MCP-1 after 3 cycles and approximately 10% after 5 cycles. When MCP-1 is spiked in culture media, no significant loss is observed after 5 cycles of freezing and thawing.

#### **12.7. Expected serum values**


A panel of 20 serum coming from apparently healthy blood donors was tested for MCP-1. The detected human MCP-1 levels ranged between 117 and 910 with a mean of 512 pg/ml.

A panel of 20 plasma coming from apparently healthy blood donors was tested for MCP-1. The detected human MCP-1 levels ranged between 180 and 790 with a mean of 382 pg/ml.

### 13. Bibliography

- 1) Rollins BJ.1997.Blood 90:909.
- 2) Baggiolini M., Dewald B. and Moser B.1997.Annu.Rev. Immunol. 15:675.
- 3) Proost P., Wuyts A. and van Damme J.1996.Int. j. Clin. Lab. Res.26:211.
- 4) Taub D.D.1996.Cytokine Growth Factor Rev. 7:355.
- 5) Yoshimura T., Yuhki N., Moore SK., Appella E., Lerman MI., Leonard EJ.1989.FEBS Lett 244:487.
- 6) Carr MW., Roth SJ., Luther E., Rose SS., Springer TA.1994.Proc Natl Acad Sci USA 91:3652.
- 7) Taub DD., Proost P., Murphy WJ., Anver M., Longo DL., van Damme J., Oppenheim JJ.1995.J.Clin.Inverst. 95:1370.
- 8) Shyy YJ., Li YS., Kolattukudy PE.1990. Biochem Biophys Res Commun 169:346.
- 9) Namiki M., Kawashima S., Yamashita T., Ozaki M., Hirase T., Ishida T., Inoue N., Hirata K., Matsukawa ., Morishita R., Kaneda Y., Yokoyama M.2002.Arterioscler Thromb Vasc Biol. 22:115.
- 10) Graves DT., Jiang Y., Valente AJ.1999.Front Biosci. 4:D571.
- 11) Grammas P., Ovase R.2001.Neurobiol Aging. 22:837.
- 12) Tarzami ST., Cheng R., Miao W., Kitsis RN., Berman JW.2002. J Mol Cell Cardiol. 34:209.
- 13) Mateo L., La Linn M., McColl SR., Cross S., Gardner J., Suhrbier A.2000.Intervirology. 43:55.
- 14) Edwards JA., Denis F., Talbot PJ.2000.J Neuroimmunol. 108:73.
- 15) McManus C., Berman JW., Berett FM., Staunton H., Farrell M., Brosnan CF.1998.J Neuroimmunol. 86:20.
- 16) Mahad DJ., Howell SJ., Woodroffe MN.2002.J Neurol Neurosurg Psychiatry. 72:498.
- 17) Kuna P., Reddigari SR., Rucinski D., Oppenheim JJ., Kaplan AP.1992. J Exp Med 175:489.

## 14. Assay Summary



### TECHNICAL CONSULTATION

OriGene Technologies, Inc.  
9620 Medical Center Dr., Suite 200  
Rockville, MD 20850

Phone: 1.888.267.4436  
Fax: 301-340-9254  
Email: techsupport@OriGene.com  
Web: www.OriGene.com

For Research Use Only  
Not for use in diagnostic procedures