



# Human IL-10 ELISA Kit

## Instructions for use

Catalogue numbers:

1x48 tests: EA102154
1x96 tests: EA102155
2x96 tests: EA102156

**For research use only**

Fast Track Your Research.....

## Table of Contents

1.	Intended use.....	3
2.	Introduction.....	3
2.1.	Summary .....	3
2.2.	Principle of the method.....	4
3.	Reagents provided and reconstitution .....	4
4.	Materials required but not provided .....	5
5.	Storage Instructions.....	5
6.	Specimen collection, processing & storage .....	5
7.	Safety & precautions for use .....	6
8.	Assay Preparation .....	7
8.1.	Assay Design.....	7
8.2.	Preparation of Wash Buffer .....	7
8.3.	Preparation of Standard Diluent Buffer.....	7
8.4.	Preparation of Standard .....	8
8.5.	Preparation of Controls.....	8
8.6.	Preparation of Biotinylated anti-IL-10 .....	9
8.7.	Preparation of Streptavidin-HRP .....	9
9.	Method.....	10
10.	Data Analysis.....	11
11.	Assay limitations .....	11
12.	Performance Characteristics .....	12
12.1.	Sensitivity.....	12
12.2.	Specificity .....	12
12.3.	Precision.....	12
12.4.	Dilution Parallelism .....	13
12.5.	Spike Recovery.....	13
12.6.	Stability .....	13
12.7.	Expected serum values .....	13
12.8.	Standard Calibration .....	13
13.	Bibliography.....	14
14.	OriGeneHuman IL-10 ELISA references .....	15
15.	Assay Summary.....	16
16.	International Summaries.....	17
16.1.	French .....	17
16.2.	Spanish.....	19

# Human IL-10 ELISA KIT

## 1. Intended use

The OriGene IL-10 ELISA kit is a solid phase sandwich ELISA for the *in-vitro* qualitative and quantitative determination of IL-10 in supernatants, buffered solutions or serum and plasma samples. This assay will recognise both natural and recombinant human IL-10.

**This kit has been configured for research use only. Not suitable for use in therapeutic procedures.**

## 2. Introduction

### 2.1. Summary

Interleukin-10 is a pleiotropic cytokine playing an important role as a regulator of lymphoid and myeloid cell function. Due to the ability of IL-10 to block cytokine synthesis and several accessory cell functions of macrophages this cytokine is a potent suppressor of the effector functions of macrophages, T-cells and NK cells. In addition, IL-10 participates in regulating proliferation and differentiation of B-cells, mast cells and thymocytes (9). The primary structure of human IL-10 has been determined by cloning the cDNA encoding the cytokine (15). The corresponding protein exerts 160 amino acids with a predicted molecular mass of 18.5 kDa (8, 15). Based on its primary structure, IL-10 is a member of the four -helix bundle family of cytokines (11). In solution human IL-10 is a homodimer with an apparent molecular mass of 39 kDa (14). Although it contains an N-linked glycosylation site, it lacks detectable carbohydrates (15). Recombinant protein expressed in *E. coli* thus retains all known biological activities. The human IL-10 gene is located on chromosome 1 and is present as a single copy in the genome (6). The human IL-10 exhibits strong DNA and amino acid sequence homology to the murine IL-10 and an open reading frame in the Epstein- Barr virus genome, BCRF1 (1, 8, 15) which shares many of the cellular cytokine's biological activities and may therefore play a role in the host- virus interaction.

The immunosuppressive properties of IL-10 (4) suggest a possible clinical use of IL-10 in suppressing rejections of grafts after organ transplantations. IL-10 can furthermore exert strong anti-inflammatory activities (4).

### IL-10 in disease

IL-10 expression was shown to be elevated in **parasite infections** like in *Schistosoma mansoni* (7), *Leishmania* (5), *Toxoplasma gondii* (12) and *Trypanosoma* (13) infection. Furthermore, high IL-10 expression was detected in **mycobacterial infections** as shown for *Mycobacterium leprae* (3), ***Mycobacterium tuberculosis*** (2) and ***Mycobacterium avium*** infections. High expression levels of IL-10 are also found in **retroviral infections** inducing immunodeficiency (10).

## 2.2. Principle of the method

A capture Antibody highly specific for IL-10 has been coated to the wells of the microtitre strip plate provided during manufacture. Binding of IL-10 in samples and known standards to the capture antibodies is completed and then any excess unbound analyte is removed. During the next incubation period the binding of the biotinylated anti-IL-10 secondary antibody to the analyte occurs. Any excess unbound secondary antibody is then removed. The HRP conjugate solution is then added to every well including the zero wells, following incubation excess conjugate is removed by careful washing. A chromogen substrate is added to the wells resulting in the progressive development of a blue coloured complex with the conjugate. The colour development is then stopped by the addition of acid turning the resultant final product yellow. The intensity of the produced coloured complex is directly proportional to the concentration of IL-10 present in the samples and standards. The absorbance of the colour complex is then measured and the generated OD values for each standard are plotted against expected concentration forming a standard curve. This standard curve can then be used to accurately determine the concentration of IL-10 in any sample tested.

## 3. Reagents provided and reconstitution

Reagents (Store@2-8°C)	Quantity 1x48 well kit Cat no. EA102154	Quantity 1x96 well kit Cat no. EA102155	Quantity 2x96 well kit Cat no. EA102156	Reconstitution
96 well microtitre strip plate	1/2	1	2	Ready to use (Pre-coated)
Plastic plate covers	2	2	4	n/a
Standard: 400pg/ml	1	2	4	Reconstitute as directed on the vial (see reagent preparation, section 8)
Standard Diluent (Buffer)	1 (25ml)	1 (25ml)	1 (25ml)	10x Concentrate, dilute in distilled water (see reagent preparation, section 8)
Standard Diluent (Human serum)	1 (7ml)	1 (7ml)	2 (7ml)	Ready to use
Control	1	2	4	Reconstitute as directed on the vial (see reagent preparation, section 8)
Biotinylated anti-IL-10	1 (0.4ml)	1 (0.4ml)	2 (0.4ml)	Dilute in biotinylated antibody diluent (see reagent preparation, section 8)
Biotinylated Antibody diluent	1 (7ml)	1 (7ml)	1 (13ml)	Ready to use
Streptavidin-HRP	1(5µl)	2 (5µl)	4 (5µl)	Add 0.5ml of HRP diluent prior to use (see reagent preparation, section 8)
HRP Diluent	1 (23ml)	1 (23ml)	1 (23ml)	Ready to use
Wash Buffer	1 (10ml)	1 (10ml)	2 (10ml)	200x Concentrate dilute in distilled water (see reagent preparation, section 8)
TMB Substrate	1 (11ml)	1 (11ml)	1 (24ml)	Ready to use
H <sub>2</sub> SO <sub>4</sub> stop reagent	1 (11ml)	1 (11ml)	2 (11ml)	Ready to use

## **4. Materials required but not provided**

- Microtitre plate reader fitted with appropriate filters (450nm required with optional 620nm reference filter)
- Microplate washer or wash bottle
- 10, 50, 100, 200 and 1,000 $\mu$ l adjustable single channel micropipettes with disposable tips
- 50-300 $\mu$ l multi-channel micropipette with disposable tips
- Multichannel micropipette reagent reservoirs
- Distilled water
- Vortex mixer
- Miscellaneous laboratory plastic and/or glass, if possible sterile

## **5. Storage Instructions**

Store kit reagents between 2 and 8°C. Immediately after use remaining reagents should be returned to cold storage (2-8°C). Expiry of the kit and reagents is stated on box front labels. The expiry of the kit components can only be guaranteed if the components are stored properly, and if, in case of repeated use of one component, the reagent is not contaminated by the first handling.

## **6. Specimen collection, processing & storage**

Cell culture supernatants, human serum, plasma or other biological samples will be suitable for use in the assay. Remove serum from the clot or red cells, respectively, as soon as possible after clotting and separation.

**Cell culture supernatants:** Remove particulates and aggregates by spinning at approximately 1000 x g for 10 min.

**Serum:** Use pyrogen/endotoxin free collecting tubes. Serum should be removed rapidly and carefully from the red cells after clotting. Following clotting, centrifuge at approximately 1000 x g for 10 min and remove serum.

**Plasma:** EDTA, citrate and heparin plasma can be assayed. Spin samples at 1000 x g for 30 min to remove particulates. Harvest plasma.

**Storage:** If not analyzed shortly after collection, samples should be aliquoted (250-500 $\mu$ l) to avoid repeated freeze-thaw cycles and stored frozen at -70°C. Avoid multiple freeze-thaw cycles of frozen specimens.

**Recommendation:** Do not thaw by heating at 37°C or 56°C. Thaw at room temperature and make sure that sample is completely thawed and homogeneous before use. When possible avoid use of badly haemolysed or lipemic sera. If large amounts of particles are present these should be removed prior to use by centrifugation or filtration.

## **7. Safety & precautions for use**

- Handling of reagents, serum or plasma specimens should be in accordance with local safety procedures, e.g.CDC/NIH Health manual : "Biosafety in Microbiological and Biomedical Laboratories" 1984.
- The human serum included in this kit have been tested and found non reactive for HbsAg, anti HIV1 & 2 and anti VHC. Nevertheless, no known method can offer complete assurance that human blood derivatives will not transmit hepatitis, AIDS or other infections. Therefore handling of reagents, serum or plasma specimens should be in accordance with local safety procedures.
- Laboratory gloves should be worn at all times.
- Avoid any skin contact with  $H_2SO_4$  and TMB. In case of contact, wash thoroughly with water.
- Do not eat, drink, smoke or apply cosmetics where kit reagents are used.
- Do not pipette by mouth.
- When not in use, kit components should be stored refrigerated or frozen as indicated on vials or bottles labels.
- All reagents should be warmed to room temperature before use. Lyophilized standards should be discarded after use.
- Once the desired number of strips has been removed, immediately reseal the bag to protect the remaining strips from deterioration.
- Cover or cap all reagents when not in use.
- Do not mix or interchange reagents between different lots.
- Do not use reagents beyond the expiration date of the kit.
- Use a clean disposable plastic pipette tip for each reagent, standard, or specimen addition in order to avoid cross contamination, for the dispensing of  $H_2SO_4$  and substrate solution, avoid pipettes with metal parts.
- Use a clean plastic container to prepare the washing solution.
- Thoroughly mix the reagents and samples before use by agitation or swirling.
- All residual washing liquid must be drained from the wells by efficient aspiration or by decantation followed by tapping the plate forcefully on absorbent paper. Never insert absorbent paper directly into the wells.
- The TMB solution is light sensitive. Avoid prolonged exposure to light. Also, avoid contact of the TMB solution with metal to prevent colour development. Warning TMB is toxic avoid direct contact with hands. Dispose off properly.
- If a dark blue colour develops within a few minutes after preparation, this indicates that the TMB solution has been contaminated and must be discarded. Read absorbance's within 1 hour after completion of the assay.
- When pipetting reagents, maintain a consistent order of addition from well-to-well. This will ensure equal incubation times for all wells.
- Follow incubation times described in the assay procedure.
- Dispense the TMB solution within 15 min of the washing of the microtitre plate.

## 8. Assay Preparation

**Bring all reagents to room temperature before use**

### 8.1. Assay Design

Determine the number of microwell strips required to test the desired number of samples plus appropriate number of wells needed for running zeros and standards. Each sample, standard, zero and control should be tested **in duplicate**. Remove sufficient Microwell Strips for testing from the plastic pouch immediately prior to use. Return any wells not required for this assay with desiccant to the pouch. Seal tightly and return to 2-8°C storage.

**Example plate layout** (example shown for a 6 point standard curve)

	Standards / Controls		Sample Wells									
	1	2	3	4	5	6	7	8	9	10	11	12
A	400	400										
B	200	200										
C	100	100										
D	50	50										
E	25	25										
F	12.5	12.5										
G	zero	zero										
H	Ctrl	Ctrl										

*All remaining empty wells can be used to test samples in duplicate*

### 8.2. Preparation of Wash Buffer

Dilute the (200x) wash buffer concentrate 200 fold with distilled water to give a 1x working solution. Pour entire contents (10 ml) of the Washing Buffer Concentrate into a clean 2,000 ml graduated cylinder. Bring final volume to 2,000 ml with glass-distilled or deionized water. Mix gently to avoid foaming. Transfer to a clean wash bottle and store at 2°-25°C.

### 8.3. Preparation of Standard Diluent Buffer

Add the contents of the vial (10x concentrate) to 225 ml of distilled water before use.

This Solution can be stored at 2-8°C for up to 1 week.

## **8.4. Preparation of Standard**

Depending on the type of samples you are assaying, the kit may include two standard diluents. Because biological fluids might contain proteases or cytokine-binding proteins that could modify the recognition of the cytokine you want to measure, you should reconstitute standard vials with the most appropriate Standard Diluent.

For **serum and plasma** samples: use Standard Diluent - human serum.

For **cell culture supernatants**: use Standard Diluent Buffer.

Standard vials must be reconstituted with the volume of standard diluent shown on the vial immediately prior to use. This reconstitution gives a stock solution of 400pg/ml of IL-10. Mix the reconstituted standard gently by repeated aspiration/inversion. Serial dilutions of the standard are made directly in the assay plate to provide the concentration range from 400 to 12.5pg/ml. A fresh standard curve should be produced for each new assay.

- Immediately after reconstitution add 200µl of the reconstituted standard to wells A1 and A2, which provides the highest concentration standard at 400pg/ml.
- Add 100µl of appropriate standard diluent to the remaining standard wells B1 and B2 to F1 and F2.
- Transfer 100µl from wells A1 and A2 to B1 and B2. Mix the well contents by repeated aspirations and ejections taking care not to scratch the inner surface of the wells.
- Continue this 1:1 dilution using 100µl from wells B1 and B2 through to wells F1 and F2 providing a serial diluted standard curve ranging from 400pg/ml to 12.5pg/ml.
- Discard 100µl from the final wells of the standard curve (F1 and F2).

Alternatively these dilutions can be performed in separate clean tubes and immediately transferred directly into the relevant wells.

## **8.5. Preparation of Controls**

Freeze-dried control vials should also be reconstituted with the most appropriate Standard Diluent to your samples.

For **serum and plasma samples**: use Standard Diluent - human serum.

For **cells culture supernatants**: use Standard Diluent Buffer.

The supplied Controls must be reconstituted with the volume of Standard Diluent indicated on the vial. Reconstitution of the freeze-dried material with the indicated volume, will give a solution at the concentration stated on the vial. Do not store after use.

## 8.6. Preparation of Biotinylated anti-IL-10

It is recommended this reagent is prepared immediately before use. Dilute the biotinylated anti-IL-10 with the biotinylated antibody diluent in an appropriate clean glass vial using volumes appropriate to the number of required wells. Please see example volumes below:

Number of wells required	Biotinylated Antibody ( $\mu$ l)	Biotinylated Antibody Diluent ( $\mu$ l)
16	40	1060
24	60	1590
32	80	2120
48	120	3180
96	240	6360

## 8.7. Preparation of Streptavidin-HRP

It is recommended to centrifuge vial for a few seconds in a microcentrifuge to collect all the volume at the bottom.

Dilute the 5 $\mu$ l vial with 0.5ml of HRP diluent **immediately before use**. Do-not keep this diluted vial for future experiments. Further dilute the HRP solution to volumes appropriate for the number of required wells in a clean glass vial. Please see example volumes below:

Number of wells required	Streptavidin-HRP ( $\mu$ l)	Streptavidin-HRP Diluent (ml)
16	30	2
24	45	3
32	60	4
48	75	5
96	150	10

## 9. Method

We strongly recommend that every vial is mixed without foaming prior to use.

Prepare all reagents as shown in section 8.

Note: final preparation of **Biotinylated Secondary Antibody** (section 8.6) and **Streptavidin-HRP** (section 8.7) should occur immediately before use.

Assay Step		Details
1.	Addition	<b>Prepare Standard curve</b> as shown in section 8.4
2.	Addition	Add 100µl of each, <b>Sample and zero (appropriate standard diluent)</b> in duplicate to appropriate number of wells
3.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for <b>1 hour</b>
4.	Wash	Remove the cover and wash the plate as follows: a) Aspirate the liquid from each well b) Dispense 0.3 ml of <b>1x washing solution</b> into each well c) Aspirate the contents of each well d) Repeat step b and c another two times
5.	Addition	Add 50µl of diluted <b>biotinylated anti-IL-10</b> to all wells
6.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for <b>1 hour</b>
7.	Wash	Repeat wash step 4.
8.	Addition	Add 100µl of <b>Streptavidin-HRP</b> solution into all wells
9.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for <b>30 min</b>
10.	Wash	Repeat wash step 4.
11.	Addition	Add 100µl of ready-to-use <b>TMB Substrate Solution</b> into all wells
12.	Incubation	Incubate in the dark for <b>10-15 minutes*</b> at room temperature. Avoid direct exposure to light by wrapping the plate in aluminium foil.
13.	Addition	Add 100µl of <b>H<sub>2</sub>SO<sub>4</sub>:Stop Reagent</b> into all wells
<b>Read the absorbance</b> value of each well (immediately after step 11.) on a spectrophotometer using 450 nm as the primary wavelength and optionally 620 nm as the reference wave length (610 nm to 650 nm is acceptable).		

\*Incubation time of the substrate solution is usually determined by the ELISA reader performance. Many ELISA readers only record absorbance up to 2.0 O.D. Therefore the colour development within individual microwells must be observed by the analyst, and the substrate reaction stopped before positive wells are no longer within recordable range

## 10. Data Analysis


Calculate the average absorbance values for each set of duplicate standards, controls and samples. Ideally duplicates should be within 20% of the mean.

Generate a linear standard curve by plotting the average absorbance of each standard on the vertical axis versus the corresponding IL10 standard concentration on the horizontal axis.

The amount of IL-10 in each sample is determined by extrapolating OD values against IL-10 standard concentrations using the standard curve.

### Example IL-10 Standard curve

Standard	IL-10 Conc pg/ml	OD (450nm) mean	CV (%)
1	400	2,163	0,5
2	200	1,116	0,7
3	100	0,618	1,8
4	50	0,341	2,3
5	25	0,191	5,6
6	12,5	0,105	0,7
zero	0	0,021	29,6



**Note:** curve shown above should not be used to determine results. Every laboratory must produce a standard curve for each set of microwell strips assayed.

## 11. Assay limitations

Do not extrapolate the standard curve beyond the maximum standard curve point. The dose-response is non-linear in this region and good accuracy is difficult to obtain. Concentrated samples above the maximum standard concentration must be diluted with Standard diluent or with your own sample buffer to produce an OD value within the range of the standard curve. Following analysis of such samples always multiply results by the appropriate dilution factor to produce actual final concentration.

The influence of various drugs on end results has not been investigated. Bacterial or fungal contamination and laboratory cross-contamination may also cause irregular results.

Improper or insufficient washing at any stage of the procedure will result in either false positive or false negative results. Completely empty wells before dispensing fresh Washing Buffer, fill with Washing Buffer as indicated for each wash cycle and do not allow wells to sit uncovered or dry for extended periods.

Disposable pipette tips, flasks or glassware are preferred, reusable glassware must be washed and thoroughly rinsed of all detergents before use.

As with most biological assays conditions may vary from assay to assay therefore **a fresh standard curve must be prepared and run for every assay**.

## 12. Performance Characteristics

### 12.1. Sensitivity

The sensitivity, minimum detectable dose of IL-10 using this OriGeneIL-10 ELISA kit was found to be less than **4.9pg/ml**. This was determined by adding 3 standard deviations to the mean OD obtained when the zero standard was assayed 39 times.

### 12.2. Specificity

The assay recognizes both natural and recombinant human IL-10. To define the specificity of this ELISA several proteins were tested for cross reactivity. There was no cross reactivity observed for any protein tested (IL-1 $\beta$ , IL-2, IL-4, IL-6, IL-8, IL-12, IL-13, IL-17A, IFN $\gamma$  and TNF $\alpha$ ).

### 12.3. Precision

#### Intra-assay

Reproducibility within the assay will be evaluated in three independent experiments. Each assay was carried out with 6 replicates (3 duplicates) of 2 human pooled serum, 2 in culture media and 2 in standard diluent with samples containing different concentrations of IL-10. 2 standard curves were run on each plate. **The overall intra-assay coefficient of variation has been calculated to be 3.5%.**

Session	Sample	Mean IL-10 pg/ml	SD	CV %
Session 1	Sample 1	226,14	2,41	1,07
	Sample 2	161,75	2,62	1,62
	Sample 3	304,46	11,95	3,93
	Sample 4	200,15	6,17	3,08
	Sample 5	209,90	13,80	6,57
	Sample 6	128,01	8,00	6,25
Session 2	Sample 1	225,88	0,25	0,11
	Sample 2	150,97	2,30	1,52
	Sample 3	304,57	28,76	9,44
	Sample 4	185,07	7,71	4,17
	Sample 5	212,52	10,44	4,91
	Sample 6	123,81	3,55	2,87
Session 3	Sample 1	172,89	7,84	4,54
	Sample 2	124,87	4,21	3,37
	Sample 3	254,52	1,35	0,53
	Sample 4	160,08	3,28	2,05
	Sample 5	174,57	4,42	2,53
	Sample 6	110,17	5,40	4,90

## **Inter-assay**

Assay to assay reproducibility within one laboratory will be evaluated in three independent experiments by two technicians. Each assay will be carry out with 6 replicates (3 duplicates) of 2 human pooled serum, 2 in culture media and 2 in standard diluent with samples containing different concentrations of IL-10. 2 standard curves were run on each plate. **The calculated overall coefficient of variation was 7.5%.**

	<b>Sample 1</b>	<b>Sample 2</b>	<b>Sample 3</b>	<b>Sample 4</b>	<b>Sample 5</b>	<b>Sample 6</b>
<b>Mean IL-10 pg/ml</b>	220	154	308	194	212	129
<b>SD</b>	14	11	24	15	16	11
<b>CV %</b>	6,4	6,8	7,8	7,9	7,5	8,3

## **12.4. Dilution Parallelism**

Two spiked human serum with different levels of IL-10 were analysed at three serial two fold dilutions (1:2-1:16) with two replicates each. Recoveries ranged from 75% to 123% with an overall **mean recovery of 102%**.

## **12.5. Spike Recovery**

The spike recovery was evaluated by spiking two concentrations of recombinant IL-10 in human serum and culture media in three experiments. Recoveries ranged from 78% to 123% with an overall **mean recovery of 103%**.

## **12.6. Stability**

### **Storage Stability**

Aliquots of spiked serum samples were stored at -20°C, 2-8°C, room temperature (RT) and at 37°C and the IL-10 level determined after 24h. There was no significant loss of IL-10 reactivity during storage at RT, and 2-8°C, however we observe 70% loss of reactivity when stored at 37°C if spiked in serum and 50% loss if spiked into culture media.

### **Freeze-thaw Stability**

Aliquots of spiked serum were stored frozen at -20°C and thawed up to 5 times and the IL-10 level was determined. There was no significant loss of IL-10 after 5 cycles of freezing and thawing.

## **12.7. Expected serum values**

A panel of 20 sera and 20 plasmas coming from apparently healthy blood donors was tested for IL-10. All were below the detection level <4.9pg/ml.

## **12.8. Standard Calibration**

This immunoassay is calibrated against the International Reference Standard NIBSC 93/722. NIBSC 93/722 is quantitated in International Units (IU). 1IU corresponding to 540pg OriGenIL-10.


## 13. Bibliography

1. Baer R., A. T. Bankier, M. D. Biggin, P. L. Deininger, P. J. Farrell, T. J. Gibson, G. Hatfull, G. S. Hudson, S. C. Satchwell, P. S. Tuffnell, and B. G. Barrell. (1984). DNA sequence and expression of the B95-8 Epstein-Barr virus genome. *Nature* 310, 207-211.
2. Barnes P. F., D. Chatterjee, J. S. Abrams, S. Lu, E. Wang, M. Yamamura, P. J. Brennan, and R. L. Modlin. (1992). Cytokine production induced by *Mycobacterium tuberculosis* lipoarabinomannan. Relationship to chemical structure. *J. Immunol.* 149, 541-547.
3. Bloom B. R., and V. Mehra. (1984). Immunological unresponsiveness in leprosy. *Immunol. Rev.* 80, 5-28.
4. De Waal Malefyt R., J. Abrams, B. Bennett, C. G. Figdor, and J. E. de Vries. (1991). Interleukin-10 inhibits cytokine synthesis by human monocytes - an autoregulatory role of IL-10 produced by monocytes. *J. Exp. Med.* 174, 1209-1220.
5. Heinzel F. P., M. D. Sadick, S. S. Mutha, and R. M. Locksley. (1991). Production of interferon gamma, interleukin 2, interleukin 4, and interleukin 10 by CD4 positive lymphocytes in-vivo during healing and progressive murine leishmaniasis. *Proc. Natl. Acad. Sci., USA* 88, 7011-7015.
6. Kim J. M., C. I. Brannan, N. G. Copeland, N. A. Jenkins, T. A. Khan, and K. W. Moore. (1992). Structure of the mouse IL-10 gene and chromosomal localization of the mouse and human genes. *J. Immunol.* 148, 3618-3623.
7. Kullberg M. C., E. J. Pearce, S. E. Hieny, A. Sher, and J. A. Berzofsky. (1992). Infection with *Schistosoma mansoni* alters Th1/Th2 cytokine responses to a non-parasite antigen. *J. Immunol.* 148, 3264-3270.
8. Moore K. W., P. Vieira, D. F. Fiorentino, M. L. Trounstein, T. A. Khan, and T. R. Mosmann. (1990). Homology of cytokine synthesis inhibitory factor (IL-10) to the Epstein Barr Virus gene BCRF1. *Science* 248, 1230-1234.
9. Moore K. W., A. O'Garra, R. de Waal Malefyt, P. Vieira, and T. R. Mosmann. (1993). Interleukin-10. *Ann. Rev. Immunol.* 11, 165-190.
10. Mosier D. E., R. A. Yetter, and H. C. Morse III. (1985). Retroviral induction of acute lymphoproliferative disease and profound immunosuppression in adult C57 Bl/6 mice. *J. Exp. Med.* 161, 766-784.
11. Shanafelt A. B., A. Miyajima, T. Kitamura, and R. A. Kastlein. (1991). The amino-terminal helix of GM-CSF and IL-5 governs high-affinity binding to their receptors. *EMBO J.* 10, 4105-4112.
12. Sher A., R. T. Gazzinelli, I. P. Oswald, M. Clerici, M. Kullberg, E. J. Pearce, J. A. Berzofsky, T. R. Mosmann, S. L. James, H. C. Morse III, and G. M. Shearer. (1992). Role of T-cell derived cytokines in the down-regulation of immune responses in parasitic and retroviral infection. *Immunol. Rev.* 127, 183-204.
13. Silva, J. S., P. J. Morrissey, K. H. Grabstein, K. M. Mohler, D. Anderson, and S. G. Reed. (1992). Interleukin 10 and interferon gamma regulation of experimental trypanosoma cruzi infection. *J. Exp. Med.* 175, 169-174.
14. Spits H., and R. de Waal Malefyt. (1992). Functional characterization of human IL-10. *Int. Arch. Allergy Immunol.* 99, 8-15.
15. Vieira P., R. de Waal Malefyt, M. N. Dang, K. E. Johnson, R. Kastlein, D. F. Fiorentino, J. E. de Vries, M. G. Roncarolo, T. R. Mosmann, and K. W. Moore, (1991). Isolation and expression of human cytokine synthesis inhibitory factor (CSIF/IL-10) cDNA clones: homology to Epstein-Barr virus open reading frame BCRF1. *Proc. Natl. Acad. Sci. USA* 88, 1172-1176.

## 14. OriGeneHuman IL-10 ELISA references

- Abbal C. et al., Int. Immunol., 1999; 11(9) : 1451 - 1462  
Bellanger, A.-P. et al., Clin. Vaccine Immunol., 2013 ; 20(8):1133-1142.  
Beyth S. et al., Blood, 2005; 105(5): 2214 - 2219  
Calabro,M.L. et al., Blood, 2009;113(19): 4525-4533.  
Cerkiene, Z. et al., Am J Reprod Immunol., 2008; 59(2): 118-26.  
Chabbert-de Ponnat, I. et al., Int Immunol.,2005; 17(4): 439-47.  
Chang, Y. et al.,FASEB J.,2010; fj.10-162560.  
Charbonnier, A. S. et al., J Leukoc Biol.,2003; 73(1): 91-9.  
Chenivesse, C. et al., J. Immunol.,2012; 189(1): 128-137.  
Cohen, N. et al.,Blood,2006;107(5): 2037-44.  
Cunin, P. et al., J. Immunol.,2011;186(7): 4175-4182.  
Das, A. et al., J. Immunol.,2012;189:3925-3935  
de Nadai, P. et al., J Immunol.,2006;176(10): 6286-93.  
Doniz-Padilla, L. et al., Eur. J. Endocrinol.,2011 ; 165(1): 129-136.  
Forsbach, A. et al., J. Immunol.,2008;180(6): 3729-3738.  
Gary-Gouy H. et al., Blood, 2002; 100(13): 4537 - 4543  
Gosset P. et al., J. Immunol., 2003 ; 170(10):4943 - 4952  
Gu, Z. J. et al., Blood,1996; 88(10): 3972-86.  
Hamdi, H. et al., Blood,2007; 110: 211 - 219  
Hatzfeld-Charbonnier, A. S. et al.,J Leukoc Biol., 2007;81: 1179 - 1187  
Iwamoto, S. et al.,J Leukoc Biol.,2005;78(2): 383-92.  
Jain, S. et al.,J. Med. Microbiol.,2009; 58(2): 180-184.  
Jouan, J. et al., J. Thorac. Cardiovasc. Surg., 2012;144(2):467-4732.  
Kayaba, H. et al., J Immunol.,2001; 167(2): 995-1003.  
Kerr J. et al., J. Gen. Virol., 2001; 82(Pt 12): 3011-3019  
Kwan, W. H. et al., J Leukoc Biol.,2007; 82: 133 - 141  
Llorente, L. et al.,Arthritis Rheum.,2000; 43(8): 1790-800.  
Lu, Z. Y. et al., FEBS Lett.,1995; 377(3): 515-8.  
Montes, C. L. et al.,Cancer Res.,(2008; 68(3): 870-879.  
Moreaux, J. et al., Blood,2011; 117(4): 1280-1290.  
Musleh, G.S. et al., Eur. J. Cardiothorac. Surg., 2009;35(3):511-514  
Naisbitt D.J. et al., J. Pharmacol. Exp. Ther., 2005; 313(3): 1058 - 1065  
Naisbitt, D. J. et al., Mol Pharmacol.,2003; 63(3): 732-41.  
Oksenhendler E. et al., Blood, 2000; 96(6): 2069 - 2073  
Pene, J. et al., J. Immunol.,2008; 180(11): 7423-7430.  
Perez S A. et al., Blood, 2003; 101(9): 3444 - 3450  
Perez S. A. et al., Blood, 2005; 106(1):158 - 166  
Perez, S. A. et al., Int Immunol.,2006; 18(1): 49-58.  
Perez, S. A. et al.,Clin Cancer Res.,2007; 13(9): 2714-21.  
Perez-Bercoff D. et al., J. Virol., 2003; 77(7): 4081 - 4094  
Pichavant, M. et al., J Immunol.,2006; 177(9): 5912-9.  
Plant, L. J. and A. B. Jonsson, Infect Immun.,2006; 74(1): 442-8.  
Rajappa, M. et al., Angiology,2009; 60(4): 419-426.  
Ralph, C.et al.,Clin. Cancer Res.,2010;16(5): 1662-1672.  
Rodriguez-Zapata, M. et al., Infect. Immun.,2010; 78:3272-3279.  
Sadik, C.D. et al., Nucleic Acids Res.,2009; gkp525  
Saverino D. et al., J. immunol., 2002; 168(1): 207 - 215  
Schif-Zuck, S. et al., J Immunol.,2006; 177(11): 8241-7.  
Schmitt, C. et al., J Leukoc Biol.,2000; 68(6): 836-44.  
Schneider E M. et al., Blood, 2002; 100(8): 2891 - 2898  
Tang, Y. et al., Am J Trop Med Hyg.,2008; 79(2): 154-158.  
Vollmer J. et al., Antimicrob. Agents Chemother., 2004; 48(6): 2314 - 2317  
Vollmer J. et al., J. Leukoc. Biol., 2004; 76(3): 585 - 593  
Woerly G. et al., J. Exp. Med., 1999; 190(4): 487 -496.  
Zhang, Y. et al., Lupus, 2011; 20: 1172 - 1181

## 15. Assay Summary



### TECHNICAL CONSULTATION

OriGene Technologies, Inc.  
9620 Medical Center Dr., Suite 200  
Rockville, MD 20850

Phone: 1.888.267.4436  
Fax: 301-340-9254  
Email: [techsupport@origene.com](mailto:techsupport@origene.com)  
Web: [www.origene.com](http://www.origene.com)

For Research Use Only  
Not for use in diagnostic procedures

## 16. International Summaries

### 16.1. French

#### PREPARATION DES REACTIFS : RESUME

<b>1. Tampon de Lavage</b> <i>(Washing Buffer)</i>	Ajouter 10 ml de <b>Tampon de Lavage concentré</b> ( <i>Washing Buffer Concentrate</i> ) 200 fois (200X) à 1990 ml d'eau distillée																		
<b>2 Tampon de Dilution du Standard</b> <i>(Standard Diluent Buffer)</i>	Ajouter 25 ml de <b>Tampon de Dilution du Standard concentré</b> 10 fois ( <i>Standard Diluent Buffer Concentrate 10X</i> ) à 225 ml d'eau distillée																		
<b>3. Standard IL-10</b> <i>(IL-10 Standard)</i>	Reconstituer le <b>Standard IL-10</b> ( <i>IL-10 Standard</i> ) en ajoutant la quantité indiquée sur le flacon avec le <b>Tampon approprié</b> (Tampon de Dilution du Standard ou Tampon de Dilution du Standard : Sérum Humain)																		
<b>4. Contrôles</b> <i>(Controls)</i>	Reconstituer les <b>Contrôles</b> ( <i>Controls</i> ) en ajoutant la quantité indiquée sur le flacon avec le <b>Tampon approprié</b> (Tampon de Dilution du Standard ou Tampon de Dilution du Standard : Sérum Humain)																		
<b>5. Anti-IL-10 Biotinylé</b> <i>(Biotinylated anti IL-10)</i>	<table><thead><tr><th>Nombre de barrettes</th><th>Anti-IL-10 Biotinylé Concentré (µl) <i>(Biotinylated anti IL-10)</i></th><th>Diluent de l'Anticorps Biotinylé (µl) <i>(Antibody Diluent)</i></th></tr></thead><tbody><tr><td>2</td><td>40</td><td>1060</td></tr><tr><td>3</td><td>60</td><td>1590</td></tr><tr><td>4</td><td>80</td><td>2120</td></tr><tr><td>6</td><td>120</td><td>3180</td></tr><tr><td>12</td><td>240</td><td>6360</td></tr></tbody></table>	Nombre de barrettes	Anti-IL-10 Biotinylé Concentré (µl) <i>(Biotinylated anti IL-10)</i>	Diluent de l'Anticorps Biotinylé (µl) <i>(Antibody Diluent)</i>	2	40	1060	3	60	1590	4	80	2120	6	120	3180	12	240	6360
Nombre de barrettes	Anti-IL-10 Biotinylé Concentré (µl) <i>(Biotinylated anti IL-10)</i>	Diluent de l'Anticorps Biotinylé (µl) <i>(Antibody Diluent)</i>																	
2	40	1060																	
3	60	1590																	
4	80	2120																	
6	120	3180																	
12	240	6360																	
<b>6. Streptavidine-HRP</b> <i>(Streptavidin-HRP)</i>	<table><thead><tr><th>Nombre de barrettes</th><th>Streptavidine-HRP pré-diluée (µl) <i>(Pre-diluted Streptavidin-HRP)</i></th><th>Diluent HRP (ml) <i>(HRP-Diluent)</i></th></tr></thead><tbody><tr><td>2</td><td>30</td><td>2</td></tr><tr><td>3</td><td>45</td><td>3</td></tr><tr><td>4</td><td>60</td><td>4</td></tr><tr><td>6</td><td>75</td><td>5</td></tr><tr><td>12</td><td>150</td><td>10</td></tr></tbody></table>	Nombre de barrettes	Streptavidine-HRP pré-diluée (µl) <i>(Pre-diluted Streptavidin-HRP)</i>	Diluent HRP (ml) <i>(HRP-Diluent)</i>	2	30	2	3	45	3	4	60	4	6	75	5	12	150	10
Nombre de barrettes	Streptavidine-HRP pré-diluée (µl) <i>(Pre-diluted Streptavidin-HRP)</i>	Diluent HRP (ml) <i>(HRP-Diluent)</i>																	
2	30	2																	
3	45	3																	
4	60	4																	
6	75	5																	
12	150	10																	

## RESUME DU PROTOCOLE OPERATOIRE: durée totale : 2h45mn

1. Ajouter 100 µl de **Tampon de Dilution du Standard** (*Standard Diluent Buffer*) approprié, en duplicate, dans les puits Standards (B1 à F2).
2. Ajouter à la pipette 200 µl de **Standard IL-10** (*IL-10 Standard*) reconstitué dans les puits A1 et A2 puis réaliser des dilutions du Standard allant de 35 à 1 pg/ml en transférant 100 µl d'un puits à l'autre. Jeter les 100 µl des derniers puits (F1 et F2).
3. Ajouter 100 µl de **Tampon de Dilution du Standard** approprié (*Standard Diluent Buffer ou Standard Diluent : human serum*) en duplicate dans les puits "blancs".
4. Ajouter 100 µl d'**échantillon** (*Sample*), en duplicate, dans les puits désignés et 100 µl de **contrôle** (*control*), en duplicate dans les puits contrôles.
5. Couvrir les barrettes de puits et incuber pendant 1 heure à température ambiante (18-25°C).
6. Vider et laver les puits 3 fois avec le **Tampon de Lavage** (*Washing Buffer*).
7. Préparer l'**anticorps anti IL-10 Biotinylé** (*Biotinylated anti IL-10*).
8. Ajouter 50 µl d'**anticorps anti IL-10 Biotinylé dilué** (*diluted biotinylated anti IL-10*) dans tous les puits.
9. Couvrir les barrettes de puits et incuber pendant 1 heure à température ambiante (18-25°C).
10. Vider et laver les puits 3 fois avec le **Tampon de Lavage** (*Washing Buffer*).
11. Préparer la Streptavidin-HRP.
12. Ajouter 100 µl de **Streptavidin-HRP diluée** (*diluted HRP-Streptavidin*) dans tous les puits.
13. Couvrir les puits et incuber pendant 30 minutes à température ambiante (18°-25°C).
14. Vider et laver les puits 3 fois avec le **Tampon de Lavage** (*Washing Buffer*).
15. Ajouter 100 µl de **solution de TMB** (*TMB solution*) prête à l'emploi dans tous les puits y compris les "blancs".
16. Incuber les barrettes de puits pendant environ 10-15 minutes à température ambiante (18°-25°C) à l'obscurité.
17. Ajouter 100 µl d'**H<sub>2</sub>SO<sub>4</sub>: Solution Stop** (*H<sub>2</sub>SO<sub>4</sub> : Stop Solution*) dans tous les puits y compris les "blancs".
18. Mesurer l'absorbance à la longueur d'onde 450 nm et optionnellement à 620 nm (entre 610 et 650 nm) comme longueur d'onde de référence.

**Remarque:** Les échantillons présentant une valeur de D.O. excédant la gamme de la courbe standard peuvent donner des concentrations incorrectes. C'est pourquoi, il est recommandé de diluer de tels échantillons avec le Tampon de Dilution du Standard (*Standard Diluent Buffer*) afin de quantifier précisément le véritable taux d'IL-10.

## 16.2. Spanish

### PREPARACIÓN DE LOS PRODUCTOS

<b>1. Tampón de Lavado</b>	Añadir <b>Tampón de Lavado Concentrado</b> 200 X (10 ml) a 1990 ml de agua destilada.		
<b>2 Tampón diluyente del estandar</b>	Añadir <b>Tampón Diluyente del Estándar Concentrado</b> 10 X (25 ml) a 225 ml de agua destilada.		
<b>3. Estándar IL-10</b>	Reconstituir el <b>Estándar IL-10</b> añadiendo el Diluyente del Estándar apropiado, como indica la etiqueta del vial.		
<b>4. Controles</b>	Reconstituir los <b>controles</b> añadiendo el Diluyente del Estándar apropiado, como indica la etiqueta del vial.		
<b>5. Anti-IL-10biotinilado</b>	Número de tiras	Anticuerpo biotinilado concentrado ( $\mu$ l)	Diluyente del anticuerpo biotinilado ( $\mu$ l)
	2	40	1,060
	3	60	1,590
	4	80	2,120
	6	120	3,180
	12	240	6,360
<b>6. Estreptavidina-HRP</b>	Número de tiras	Estreptavidina-HRP prediluida ( $\mu$ l)	Diluyente de HRP (ml)
	2	30	2
	3	45	3
	4	60	4
	6	75	5
	12	150	10

## **RESUMEN DEL PROTOCOLO. El procedimiento total tiene una duración de 2h45min.**

1. Añadir 100 µl del **Tampón Diluyente del Estándar** apropiado, por duplicado, a los pocillos designados para el estándar (B1 to F2).
2. Pipetear 200 µl del **Estándar IL-10** reconstituido en los pocillos A1 y A2 y hacer diluciones seriadas del estándar con el rango de concentraciones de 400 al 12.5 pg/ml, transfiriendo 100 µl de un pocillo al siguiente. Descartar 100 µl de los últimos pocillos.
3. Añadir 100 µl del **Tampón Diluyente del Estándar** apropiado, por duplicado, a los pocillos que van a ser el “blanco”.
4. Añadir 100 µl de las muestras, por duplicado, a los pocillos designados para ello, y 100 µl del Control reconstituido, por duplicado, a los pocillos designados como “control”.
5. Cubrir la placa e incubar durante 1 horas a temperatura ambiente (18-25°C).
6. Vaciar y lavar la placa 3 veces con **Tampón de Lavado**.
7. Preparar el anticuerpo **Anti-IL-10 Biotinilado**.
8. Añadir 50 µl del **anti-IL-10 Biotinilado** y diluido, a todos los pocillos.
9. Cubrir la placa e incubar durante 1 horas a temperatura ambiente (18-25°C).
10. Vaciar y lavar la placa 3 veces con **Tampón de Lavado**.
11. Preparar la **Estreptavidina-HRP**.
12. Añadir 100 µl de **Estreptavidina-HRP** diluida a todos los pocillos.
13. Cubrir la placa e incubar 30 minutos a temperatura ambiente (18° to 25°C).
14. Vaciar y lavar la placa 3 veces con **Tampón de Lavado**.
15. Añadir 100 µl de solución **TMB preparado para utilizar**, a todos los pocillos, incluidos los pocillos con “blancos”.
16. Incubar la placa durante 10-15 minutos a temperatura ambiente (18° to 25°C) y en oscuridad.
17. Añadir 100 µl de H<sub>2</sub>SO<sub>4</sub>: **Solución de Parada**, a todos los pocillos, incluidos los pocillos con los “blancos”.
18. Medir la intensidad de color a 450 nm y a 620 nm como longitud de onda de referencia (de 610 nm a 650 nm sería aceptable).

**Nota:** El cálculo de concentraciones de muestras con densidad óptica que supere el rango de la curva estándar, resultaría incorrecto, dando niveles de IL-10 más bajos de lo real. Estas muestras, requerirían ser diluidas con el Tampón de Dilución de Estándar adecuado, para poder precisar la cantidad real de IL-10 .