

## Mouse CD206/CLEC13D ELISA Kit

Catalog Number: EA102727

### Assay Principle

The OriGene Mouse Mrc1 Pre-Coated ELISA (Enzyme-Linked Immunosorbent Assay) kit is a solid phase immunoassay specially designed to measure Mouse Mrc1 with a 96-well strip plate that is pre-coated with antibody specific for Mrc1. The detection antibody is a biotinylated antibody specific for Mrc1. The capture antibody is a monoclonal antibody from rat, the detection antibody is a biotinylated polyclonal antibody from goat. The kit contains recombinant Mouse Mrc1 with immunogen: Expression system for standard: NSO, Immunogen sequence: L20-A1388. The kit is analytically validated with ready to use reagents.

To measure Mouse Mrc1, add standards and samples to the wells, then add the biotinylated detection antibody. Wash the wells with PBS or TBS buffer, and add Avidin-Biotin-Peroxidase Complex (ABC-HRP). Wash away the unbound ABC-HRP with PBS or TBS buffer and add TMB. TMB is substrate to HRP and will be catalyzed to produce a blue color product, which changes into yellow after adding acidic stop solution. The density of the yellow product is linearly proportional to Mouse Mrc1 in the sample. Read the density of the yellow product in each well using a plate reader, and benchmark the sample wells' readings against the standard curve to determine the concentration of Mouse Mrc1 in the sample.

### Overview

<b>Product Name</b>	Mouse CD206/CLEC13D ELISA Kit
<b>Reactive Species</b>	Mouse
<b>Size</b>	96wells/kit, with removable strips.
<b>Description</b>	Sandwich High Sensitivity ELISA kit for Quantitative Detection of mouse CD206/CLEC13D. 107wells/kit, with removable strips.
<b>Sensitivity</b>	<50pg/ml *The sensitivity or the minimum detectable dose (MDD) is the lower limit of target protein that can be detected by the kit. It is determined by adding two standard deviations to the mean O.D. value of twenty (20) blank wells and calculating the corresponding concentration.
<b>Detection Range</b>	0.78ng/ml-50ng/ml
<b>Storage Instructions</b>	Store at 4°C for 6 months, at -20°C for 12 months. Avoid multiple freeze-thaw cycles (Shipped with wet ice.)
<b>Uniprot ID</b>	Q61830

## Technical Details

<b>Capture/Detection Antibodies</b>	<i>The capture antibody is a monoclonal antibody from rat, the detection antibody is a biotinylated polyclonal antibody from goat.</i>
<b>Specificity</b>	<i>Natural and recombinant Mouse Mrc1</i>
<b>Immunogen</b>	<i>Expression system for standard: NSO, Immunogen sequence: L20-A1388</i>
<b>Cross Reactivity</b>	<i>There is no detectable cross-reactivity with other relevant proteins.</i>

## Notice Before Application

Please read the following instructions before starting the experiment.

1. To inspect the validity of experiment operation and the appropriateness of sample dilution proportion, pilot experiment using standards and a small number of samples is recommended.
2. Before using the Kit, spin tubes and bring down all components to the bottom of tubes.
3. Don't let 96-well plate dry, for dry plate will inactivate active components on plate.
4. Don't reuse tips and tubes to avoid cross contamination.
5. Avoid using the reagents from different batches together.

## Kit Components/Materials Provided

<b>Description</b>	<b>Quantity</b>	<b>Volume</b>
<b>Anti-Mouse Mrc1 Pre-coated 96-well strip microplate</b>	<b>1</b>	<b>12 strips of 8 wells</b>
<b>Mouse Mrc1 Standard</b>	<b>2</b>	<b>50ng/tube</b>
<b>Mouse Mrc1 Biotinylated antibody (100x)</b>	<b>1</b>	<b>130 µl</b>
<b>Avidin-Biotin-Peroxidase Complex (100x)</b>	<b>1</b>	<b>130 µl</b>
<b>Sample Diluent</b>	<b>1</b>	<b>30ml</b>
<b>Antibody Diluent</b>	<b>1</b>	<b>12ml</b>
<b>Avidin-Biotin-Peroxidase Diluent</b>	<b>1</b>	<b>12ml</b>
<b>Color Developing Reagent (TMB)</b>	<b>1</b>	<b>10ml</b>
<b>Stop Solution</b>	<b>1</b>	<b>10ml</b>
<b>Plate Sealers</b>	<b>4</b>	<b>Piece</b>

\*Why there is no wash buffer? Our Avidin-Biotin-Peroxidase Diluent contains the detergent (TWEEN) normally present in other companies' ELISA kits. This saves you the step of having to wash with the special wash buffer and achieves similar or better signal to noise ratio. The wash can use regular wash buffers (PBS, TBS etc.) commonly found in labs.

## Required Materials That Are Not Supplied

Microplate Reader capable of reading absorbance at 450nm.

Automated plate washer (optional)

1000ml of 1X wash buffer (TBS or PBS)

Pipettes and pipette tips capable of precisely dispensing 0.5 µl through 1 ml volumes of aqueous solutions.

Multichannel pipettes are recommended for large amount of samples.

Deionized or distilled water.

500ml graduated cylinders.


Test tubes for dilution.

## Mouse CD206/CLEC13D ELISA Kit (EA102727) Standard Curve Example

Highest O.D. value might be higher or lower than in the example. The experiment result is statistically significant if the highest O.D. value is no less than 1.0.

<b>Concentration 0 (ng/ml)</b>	<b>0.78</b>	<b>1.56</b>	<b>3.12</b>	<b>6.25</b>	<b>12.5</b>	<b>25</b>	<b>50</b>
<b>O.D.</b>	<b>0.049</b>	<b>0.188</b>	<b>0.306</b>	<b>0.514</b>	<b>0.979</b>	<b>1.383</b>	<b>2.056</b>

**Mouse CD206/CLEC13D ELISA Kit Standard Curve**



A standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

## Intra/Inter Assay Variability

OriGene spend great efforts in documenting lot to lot variability and make sure our assay kits produce robust data that are reproducible.

**Intra-Assay Precision (Precision within an assay):** Three samples of known concentration were tested on one plate to assess intra-assay precision.

**Inter-Assay Precision (Precision across assays):** Three samples of known concentration were tested in separate assays to assess inter-assay

precision.

Sample	Intra-Assay Precision			Inter-Assay Precision		
	1	2	3	1	2	3
n	16	16	16	24	24	24
Mean(pg/ml)	1150	6191	22279	1234	6199	23488
Standard deviation	59.80	266.21	913.44	81.44	359.54	1550.21
CV(%)	5.2%	4.3%	4.1%	6.6%	5.8%	6.6%

## Reproducibility

To assay reproducibility, three samples with differing target protein concentrations were assayed using four different lots.

Lots	Lot1 (pg/ml)	Lot2 (pg/ml)	Lot3 (pg/ml)	Lot4 (pg/ml)	Mean (pg/ml)	Standard Deviation	CV (%)
Sample 1	1150	1144	1176	1080	1137	35.30	3.1%
Sample 2	6191	5982	6880	6356	6352	332.27	5.2%
Sample 3	22279	22618	21731	19374	21500	1267.87	5.8%

\*number of samples for each test n=16.

## Preparation Before The Experiment

Item	Preparation
All reagents	Bring all reagents to 37°C prior to use. The assay can also be done at room temperature however we recommend doing it at 37°C for best consistency with our QC results. Also the TMB incubation time estimate (15-25min) is based on 37°C.
Wash buffer	Prepare 1000ml of 1X PBS or TBS for wash buffer.
Biotinylated Anti-Mouse Mrc1 antibody	It is recommended to prepare this reagent immediately prior to use by diluting the Mouse Mrc1 Biotinylated antibody (100x) 1: 100 with Antibody Diluent. Prepare 100 µl by adding 1 µl of Biotinylated antibody (100x) to 99 µl of Antibody Diluent for each well. Mix gently and thoroughly and use within 2 hours of generation.
Avidin-Biotin-Peroxidase Complex	It is recommended to prepare this reagent immediately prior to use by diluting the Avidin-Biotin-Peroxidase Complex (100x) 1: 100 with Avidin-Biotin-Peroxidase Diluent. Prepare 100 µl by adding 1 µl of Avidin-Biotin-Peroxidase Complex (100x) to 99 µl of Avidin-Biotin-Peroxidase Diluent for each well. Mix gently and thoroughly and use within 2 hours of generation.
Mouse Mrc1 Standard	It is recommended that the standards be prepared no more than 2 hours prior to performing the

	<p>experiment. Use one 50ng of lyophilized Mouse Mrc1 standard for each experiment. Gently spin the vial prior to use. Reconstitute the standard to a stock concentration of 50ng/ml using 1ml of sample diluent. Allow the standard to sit for a minimum of 10 minutes with gentle agitation prior to making dilutions.</p>
Microplate	<p>The included microplate is coated with capture antibodies and ready-to-use. It does not require additional washing or blocking. The unused well strips should be sealed and stored in the original packaging.</p>

## Dilution of Mouse Mrc1 Standard

1. Number tubes 1-8. Final Concentrations to be Tube #1 – 50000pg/ml, #2 – 25000pg/ml, #3 – 12500pg/ml, #4 – 6250pg/ml, #5 – 3125pg/ml, #6 – 1562.5pg/ml, #7 – 781.25pg/ml, #8 – 0.0 (Blank).
2. For standard #1, add 1000µl of undiluted standard stock solution to tube #1.
3. Add 300 µl of sample diluent to tubes # 2-7.
4. To generate standard #2, add 300 µl of standard #1 from tube #1 to tube #2 for a final volume of 600 µl. Mix thoroughly.
5. To generate standard #3, add 300 µl of standard #2 from tube #2 to tube #3 for a final volume of 600 µl. Mix thoroughly.
6. Continue the serial dilution for tube #4-7.
7. Tube #8 is a blank standard to be used with every experiment.

## Sample Preparation and Storage

These sample collection instructions and storage conditions are intended as a general guideline and the sample stability has not been evaluated.

Sample Type	Procedure
Cell culture supernatants	Clear sample of particulates by centrifugation, assay immediately or store samples at -20°C.
Serum	Use a serum separator tube (SST) and allow serum to clot at room temperature for about four hours. Then, centrifuge for 15 min at approximately 1,000 x g. assay immediately or store samples at -20°C.
Plasma	Collect plasma using heparin, EDTA or citrate as an anticoagulant. Centrifuge for 15 min at approximately 1,000 x g. Assay immediately or store samples at -20°C. *Note: it is important to not use anticoagulants other than the ones described above to treat plasma for other anticoagulants could block the antibody binding site.

## Sample Dilution

The target protein concentration should be estimated and appropriate sample dilutions should be selected such that the final protein concentration lies near the middle of the linear dynamic range of the assay.

It is recommended to prepare 150µl of sample for each replicate to be assayed. The samples should be diluted with sample diluent and mixed

*gently.*

## **Assay protocol**

*It is recommended that all reagents and materials be equilibrated to 37°C/room temperature prior to the experiment (see Preparation Before The Experiment if you have missed this information).*

1. Prepare all reagents and working standards as directed previously.
2. Remove excess microplate strips from the plate frame and seal and store them in the original packaging.
3. Add 100 µl of the standard, samples, or control per well. Add 100 µl of the sample diluent buffer into the control well (Zero well). At least two replicates of each standard, sample, or control is recommended.
4. Cover with the plate sealer provided and incubate for 120 minutes at RT (or 90 min. at 37 °C).
5. Remove the cover and discard the liquid in the wells into an appropriate waste receptacle. Invert the plate on the benchtop onto a paper towel and tap the plate to gently blot any remaining liquid. It is recommended that the wells are not allowed to completely dry at any time.
6. Add 100 µl of the prepared 1x Biotinylated Anti-Mouse Mrc1 antibody to each well.
7. Cover with plate sealer and incubate for 90 minutes at RT (or 60 minutes at 37°C).
8. Wash the plate 3 times with the 1x wash buffer.
  - a. Discard the liquid in the wells into an appropriate waste receptacle. Then, invert the plate on the benchtop onto a paper towel and tap the plate to gently blot any remaining liquid. It is recommended that the wells are not allowed to completely dry at any time.
  - b. Add 300 µl of the 1x wash buffer to each assay well. (For cleaner background incubate for 60 seconds between each wash).
  - c. Repeat steps a-b 2 additional times.
9. Add 100 µl of the prepared 1x Avidin-Biotin-Peroxidase Complex into each well. Cover with the plate sealer provided and incubate for 40 minutes at RT (or 30 minutes at 37°C).
10. Wash the plate 5 times with the 1x wash buffer.
  - a. Discard the liquid in the wells into an appropriate waste receptacle. Then, invert the plate on the benchtop onto a paper towel and tap the plate to gently blot any remaining liquid. It is recommended that the wells are not allowed to completely dry at any time.
  - b. Add 300 µl of the 1x wash buffer to each assay well. (For cleaner background incubate for 60 seconds between each wash).
  - c. Repeat steps a-b 4 additional times.
11. Add 90 µl of Color Developing Reagent to each well. Cover with the plate sealer provided and incubate in the dark for 30 minutes at RT (or 15-25 minutes at 37°C). (The optimal incubation time must be empirically determined. A guideline to look for is blue shading the top four standard wells, while the remaining standards remain clear.)
12. Add 100 µl of Stop Solution to each well. The color should immediately change to yellow.
13. Within 30 minutes of stopping the reaction, the O.D. absorbance should be read with a microplate reader at 450nm.

## **Data Analysis**

*Average the duplicate readings for each standard, sample, and control. Subtract the average zero standard O.D. reading.*

*It is recommended that a standard curve be created using computer software to generate a four parameter logistic (4-PL) curve-fit. A free program capable of generating a four parameter logistic (4-PL) curve-fit can be found online at: [www.myassays.com/four-parameter-logistic-curve.assay](http://www.myassays.com/four-parameter-logistic-curve.assay).*

*Alternatively, plot the mean absorbance for each standard against the concentration. The measured concentration in the sample can be*

---

*interpolated by using linear regression of each average relative OD against the standard curve generated using curve fitting software. This will generate an adequate but less precise fit of the data.*

*For diluted samples, the concentration reading from the standard curve must be multiplied by the dilution factor.*

## **Background on Mrc1**

*The mannose receptor (Cluster of Differentiation 206, CD206) is a C-type lectin primarily present on the surface of macrophages and immature dendritic cells, but is also expressed on the surface of skin cells such as human dermal fibroblasts and keratinocytes. It is the first member of a family of endocytic receptors that includes Endo180 (CD280), M-type PLA2R, and DEC-205 (CD205). This gene is mapped to 2 A2; 2 10.46 cM. The recognition of complex carbohydrate structures on glycoproteins is an important part of several biological processes, including cell-cell recognition, serum glycoprotein turnover, and neutralization of pathogens. The protein encoded by this gene is a type I membrane receptor that mediates the endocytosis of glycoproteins by macrophages. The protein has been shown to bind high-mannose structures on the surface of potentially pathogenic viruses, bacteria, and fungi so that they can be neutralized by phagocytic engulfment.*