

9620 Medical Center Drive, Suite 200, Rockville, MD 20850Phone: 1.888.267.4436Fax: 301-340-9254Email: techsupport@origene.comWeb: www.origene.com

Monkey primate IGF-2 ELISA Kit

Catalog Number: EA102810

Assay Principle

The Boster Monkey IGF2 Pre-Coated ELISA (Enzyme-Linked Immunosorbent Assay) kit is a solid phase immunoassay specially designed to measure Monkey IGF2 with a 96-well strip plate that is pre-coated with antibody specific for IGF2. The detection antibody is a biotinylated antibody specific for IGF2. The capture antibody is monoclonal antibody from mouse, the detection antibody is polyclonal antibody from goat. The kit contains recombinant Monkey IGF2 with immunogen: Expression system for standard: E.coli; Immunogen sequence: A25-E91. The kit is analytically validated with ready to use reagents.

To measure Monkey IGF2, add standards and samples to the wells, then add the biotinylated detection antibody. Wash the wells with PBS or TBS buffer, and add Avidin-Biotin-Peroxidase Complex (ABC-HRP). Wash away the unbounded ABC-HRP with PBS or TBS buffer and add TMB. TMB is substrate to HRP and will be catalyzed to produce a blue color product, which changes into yellow after adding acidic stop solution. The density of the yellow product is linearly proportional to Monkey IGF2 in the sample. Read the density of the yellow product in each well using a plate reader, and benchmark the sample wells' readings against the standard curve to determine the concentration of Monkey IGF2 in the sample. For more information on assay principle, protocols, and troubleshooting tips, see Boster's ELISA Resource Center at https://www.bosterbio.com/elisa-technical-resource-center.

Overview

Product Name	Monkey primate IGF-2 ELISA Kit		
Reactive Species	Monkey		
Size	96wells/kit, with removable strips.		
Description	Sandwich High Sensitivity ELISA kit for Quantitative Detection of monkey primate IGF-2. 96wells/kit, with removable strips.		
Sensitivity	<2pg/ml *The sensitivity or the minimum detectable dose (MDD) is the lower limit of target protein that can be detected by the kit. It is determined by adding two standard deviations to the mean O.D. value of twenty (20) blank wells and calculating the corresponding concentration.		
Detection Range	62.5pg/ml-4000pg/ml		
Storage Instructions	Store at 4°C for 6 months, at -20°C for 12 months. Avoid multiple freeze-thaw cycles(Shipped with wet ice.)		
Uniprot ID	P01344		



9620 Medical Center Drive, Suite 200, Rockville, MD 20850Phone: 1.888.267.4436Fax: 301-340-9254Email: techsupport@origene.comWeb: www.origene.com

Technical Details

Capture/Detection Antibodies	The capture antibody is monoclonal antibody from mouse, the detection antibody is polyclonal antibody from goat.
Specificity	Natural and recombinant Monkey IGF2
Immunogen	Expression system for standard: E.coli; Immunogen sequence: A25-E91
Cross Reactivity	There is no detectable cross-reactivity with IGF-1.

Notice Before Application

Please read the following instructions before starting the experiment.

- 1. To inspect the validity of experiment operation and the appropriateness of sample dilution proportion, pilot experiment using standards and a small number of samples is recommended.
- 2. Before using the Kit, spin tubes and bring down all components to the bottom of tubes.
- 3. Don't let 96-well plate dry, for dry plate will inactivate active components on plate.
- 4. Don't reuse tips and tubes to avoid cross contamination.
- 5. Avoid using the reagents from different batches together.

Kit Components/Materials Provided

Quantity	Volume
1	12 strips of 8 wells
2	10ng/tube
1	130 µl
1	130 µl
1	30ml
1	12ml
1	12ml
1	10ml
1	10ml
4	Piece
	1 2 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1

*Why there is no wash buffer? Our Avidin-Biotin-Peroxidase Diluent contains the detergent (TWEEN) normally present in other companies' ELISA kits. This saves you the step of having to wash with the special wash buffer and achieve similar or better signal to noise ratio. The wash can use regular wash buffers (PBS, TBS etc.) commonly found in labs.



9620 Medical Center Drive, Suite 200, Rockville, MD 20850Phone: 1.888.267.4436Fax: 301-340-9254Email: techsupport@origene.comWeb: www.origene.com

Required Materials That Are Not Supplied

Microplate Reader capable of reading absorbance at 450nm.

Automated plate washer (optional)

1000ml of 1X wash buffer (TBS or PBS)

Pipettes and pipette tips capable of precisely dispensing 0.5 µl through 1 ml volumes of aqueous solutions.

Multichannel pipettes are recommended for large amount of samples.

Deionized or distilled water.

500ml graduated cylinders.


Test tubes for dilution.

Monkeyprimate IGF-2 ELISA Kit (EA102810) Standard Curve Example

Highest O.D. value might be higher or lower than in the example. The experiment result is statistically significant if the highest O.D. value is no less than 1.0.

Concentration 0	4000	2000	4000	2000	4000	2000	4000	
(pg/ml) O.D.	0.136	0.158	0.179	0.200	0.298	0.612	1.166	1.721

Monkey Primate IGF-2 ELISA Kit standard curve



A standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

Intra/Inter Assay Variability

Boster spend great efforts in documenting lot to lot variability and make sure our assay kits produce robust data that are reproducible.

Intra-Assay Precision (Precision within an assay): Three samples of known concentration were tested on one plate to assess intra-assay precision.

Inter-Assay Precision (Precision across assays): Three samples of known concentration were tested in separate assays to assess inter-assay



9620 Medical Center Drive, Suite 200, Rockville, MD 20850Phone: 1.888.267.4436Fax: 301-340-9254Email: techsupport@origene.comWeb: www.origene.com

precision.

Intra-Assay Precision				Inter-Assay	Precision	
Sample	1	2	3	1	2	3
n	16	16	16	24	24	24
Mean(pg/ml)	648	1657	2635	759	1830	3176
Standard deviation	26.6	81.2	152.83	47.82	122.6	236.7
CV(%)	4.1	4.9	5.8	6.7	6.7	7.5

Preparation Before The Experiment

Item	Preparation
All reagents	Bring all reagents to 37°C prior to use. The assay can also be done at room temperature however we recommend doing it at 37°C for best consistency with our QC results. Also the TMB incubation time estimate (15-25min) is based on 37°C.
Wash buffer	Prepare 1000ml of 1X PBS or TBS for wash buffer.
Biotinylated Anti-Monkey IGF2 antibody	It is recommended to prepare this reagent immediately prior to use by diluting the Monkey IGF2 Biotinylated antibody (100x) 1: 100 with Antibody Diluent. Prepare 100 μl by adding 1 μl of Biotinylated antibody (100x) to 99μl of Antibody Diluent for each well. Mix gently and thoroughly and use within 2 hours of generation.
Avidin-Biotin-Peroxidase Complex	It is recommended to prepare this reagent immediately prior to use by diluting the Avidin-Biotin- Peroxidase Complex (100x) 1: 100 with Avidin-Biotin-Peroxidase Diluent. Prepare 100 μl by adding 1 μl of Avidin-Biotin-Peroxidase Complex (100x) to 99 μl of Avidin-Biotin-Peroxidase Diluent for each well. Mix gently and thoroughly and use within 2 hours of generation.
Monkey IGF2 Standard	It is recommended that the standards be prepared no more than 2 hours prior to performing the experiment. Use one 10ng of lyophilized Monkey IGF2 standard for each experiment. Gently spin the vial prior to use. Reconstitute the standard to a stock concentration of 10ng/ml using 1ml of sample diluent. Allow the standard to sit for a minimum of 10 minutes with gentle agitation prior to making dilutions.
Microplate	The included microplate is coated with capture antibodies and ready-to-use. It does not require additional washing or blocking. The unused well strips should be sealed and stored in the original packaging.

Dilution of Monkey IGF2 Standard

1. Number tubes 1-8. Final Concentrations to be Tube #1-4000pg/ml, #2-2000pg/ml, #3-1000pg/ml, #4-500pg/ml, #5-250pg/ml, #6-



9620 Medical Center Drive, Suite 200, Rockville, MD 20850Phone: 1.888.267.4436Fax: 301-340-9254Email: techsupport@origene.comWeb: www.origene.com

125pg/ml, #7 – 62.5pg/ml, #8 – 0.0 (Blank).

2. To generate standard #1, add 400µl of the reconstituted standard stock solution of 10ng/ml and 600µl of sample diluent to tube #1 for a final volume of 1000µl. Mix thoroughly.

- 3. Add 300 μ I of sample diluent to tubes # 2-7.
- 4. To generate standard #2, add 300 µl of standard #1 from tube #1 to tube #2 for a final volume of 600 µl. Mix thoroughly.
- 5. To generate standard #3, add 300 μ l of standard #2 from tube #2 to tube #3 for a final volume of 600 μ l. Mix thoroughly.
- 6. Continue the serial dilution for tube #4-7.
- 7. Tube #8 is a blank standard to be used with every experiment.

Sample Preparation and Storage

These sample collection instructions and storage conditions are intended as a general guideline and the sample stability has not been evaluated.

Sample Type	Procedure
Cell culture supernatants	Clear sample of particulates by centrifugation, assay immediately or store samples at -20°C.
Serum	Use a serum separator tube (SST) and allow serum to clot at room temperature for about four hours. Then, centrifuge for 15 min at approximately 1,000 x g. assay immediately or store samples at -20°C.
Plasma	Collect plasma using heparin or EDTA as an anticoagulant. Centrifuge for 15 min at approximately 1,000 x g. Assay immediately or store samples at -20°C. *Note: it is important to not use anticoagulants other than the ones described above to treat plasma for other anticoagulants could block the antibody binding site.

Sample Dilution

The target protein concentration should be estimated and appropriate sample dilutions should be selected such that the final protein concentration lies near the middle of the linear dynamic range of the assay.

It is recommended to prepare $150\,\mu$ I of sample for each replicate to be assayed. The samples should be diluted with sample diluent and mixed gently.

Assay protocol

It is recommended that all reagents and materials be equilibrated to 37°C/room temperature prior to the experiment (see Preparation Before The Experiment if you have missed this information).

- 1. Prepare all reagents and working standards as directed previously.
- 2. Remove excess microplate strips from the plate frame and seal and store them in the original packaging.

3. Add 100 µl of the standard, samples, or control per well. Add 100 µl of the sample diluent buffer into the control well (Zero well). At least two replicates of each standard, sample, or control is recommended.

OriGene Technologies Inc.

9620 Medical Center Drive, Suite 200, Rockville, MD 20850Phone: 1.888.267.4436Fax: 301-340-9254Email: techsupport@origene.comWeb: www.origene.com

4. Cover with the plate sealer provided and incubate for 120 minutes at RT (or 90 min. at 37 °C).

5. Remove the cover and discard the liquid in the wells into an appropriate waste eceptacle. Invert the plate on the benchtop onto a paper towel and tap the plate to gently blot any remaining liquid. It is recommended that the wells are not allowed to completely dry at any time.

6. Add 100 μ I of the prepared 1x Biotinylated Anti-Monkey IGF2 antibody to each well.

7. Cover with plate sealer and incubate for 90 minutes at RT (or 60 minutes at 37° C).

8. Wash the plate 3 times with the 1x wash buffer.

a. Discard the liquid in the wells into an appropriate waste receptacle. Then, invert the plate on the benchtop onto a paper towel and tap the plate to gently blot any remaining liquid. It is recommended that the wells are not allowed to completely dry at any time.

b. Add 300 μ l of the 1x wash buffer to each assay well. (For cleaner background incubate for 60 seconds between each wash).

c. Repeat steps a-b 2 additional times.

9. Add 100 µl of the prepared 1x Avidin-Biotin-Peroxidase Complex into each well. Cover with the plate sealer provided and incubate for 40 minutes at RT (or 30 minutes at 37°C).

10. Wash the plate 5 times with the 1x wash buffer.

a. Discard the liquid in the wells into an appropriate waste receptacle. Then, invert the plate on the benchtop onto a paper towel and tap the plate to gently blot any remaining liquid. It is recommended that the wells are not allowed to completely dry at any time.

b. Add 300 µl of the 1x wash buffer to each assay well. (For cleaner background incubate for 60 seconds between each wash).

c. Repeat steps a-b 4 additional times.

 $11. \ \textit{Add 90} \mu \textit{l of Color Developing Reagent to each well. Cover with the plate sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided and incubate in the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at RT (or not sealer provided at the dark for 30 minutes at the dark for 30 minutes at RT (or not s$

15-25 minutes at 37°C). (The optimal incubation time must be empirically determined. A guideline to look for is blue shading the top four standard wells, while the remaining standards remain clear.)

12. Add 100 µl of Stop Solution to each well. The color should immediately change to yellow.

13. Within 30 minutes of stopping the reaction, the O.D. absorbance should be read with a microplate reader at 450nm.

Data Analysis

Average the duplicate readings for each standard, sample, and control. Subtract the average zero standard O.D. reading.

It is recommended that a standard curve be created using computer software to generate a four parameter logistic (4-PL) curve-fit. A free program capable of generating a four parameter logistic (4-PL) curve-fit can be found online at: www.myassays.com/four-parameter-logistic- curve.assay.

Alternatively, plot the mean absorbance for each standard against the concentration. The measured concentration in the sample can be interpolated by using linear regression of each average relative OD against the standard curve generated using curve fitting software. This will generate an adequate but less precise fit of the data.

For diluted samples, the concentration reading from the standard curve must be multiplied by the dilution factor.

Background on IGF2

Insulin-like growth factor II is also known as somatomedin A. IGF-2 is a member of the insulin family of polypeptide growth factors that is involved in development and growth. It is paternally expressed in the fetus and placenta. IGF-II is a mitogen for many cell types and an important modulator of muscle growth and differentiation. IGF-II gene is prevalently expressed during prenatal development and its gene activity is regulated by genomic imprinting, in that the allele inherited from the father is active and the allele inherited from the most

OriGene Technologies Inc.

9620 Medical Center Drive, Suite 200, Rockville, MD 20850Phone: 1.888.267.4436Fax: 301-340-9254Email: techsupport@origene.comWeb: www.origene.com

normal tissues. IGF-II appears to be induced by placental lactogen during prenatal development. It is a mediator of prolactin-induced alveologenesis; prolactin, IGF-2, and cyclin D1, all of which are overexpressed in breast cancers, are components of a developmental pathway in the mammary gland. The monkey primate IGF-II gene is located on chromosome 11. The standard product used in this kit is recombinant monkey primate IGF-2, consisting of 67 amino acids in mature form with the molecular mass of 7.5KDa. IGF-1 and IGF-2 share high homology.