

Human CXCL1 Immunoassay

Catalog Number: EA800056

For the quantitative determination of human CXCL1 concentrations in cell culture supernates, serum, and plasma.

For research use only. Not for use in diagnostic procedures.

MANUFACTURED AND DISTRIBUTED BY:

OriGene Technologies,Inc.

9620 Medical Center Drive Suite 200 Rockville, MD 20850,USA

Tel: 1-301-340-3188 Fax: 301-340-9254 Mail: techsupport@origene.com Web: www.origene.com



TABLE OF CONTENTS

SECTION	PAGE	
BACKGROUND		1
PRINCIPLE OF THE ASSAY		1
TECHNICAL HINTS AND LIMITATIONS.		2
PRECAUTIONS		2
KIT COMPONENTS& STORAGE CONDIT	ΓΙΟΝS	3
OTHER SUPPLIES REQUIRED BUT NOT	SUPPLIED	
SPECIMEN COLLECTION & STORAGE		2
REAGENTS PREPARATION		4
ASSAY PROCEDURE		6
CALCULATION OF RESULTS		6
PERFORMANCE CHARACTERISTICS		8
REFERENCES		10




BACKGROUND

GRO α , or CXCL1, was initially discovered as a growth-regulated gene that is overexpressed constitutively in tumorigenic cells and transcribed in normal cells only during growth stimulation. The GRO α cDNA encodes a 107 aa precursor protein from which the N-terminal 34 aa residues are cleaved to generate the 73 aa residue mature GRO α . There are no potential N-linked glycosylation sites in the aa sequence nor does the mature protein appear to undergo any post-translational modifications. GRO α exhibits growth stimulatory activity for the human Hs294T melanoma cell line from which it was initially isolated. As a chemokine α member, GRO α , like IL-8, is a potent chemoattractant for human neutrophils and stimulates neutrophil degranulation and enzyme release from cytochalasin B-treated human neutrophils. It has been demonstrated that GRO α can cause a decrease in the expression of interstitial collagens by rheumatoid synovial fibroblasts. In in vivo experiments, GRO α has been shown to induce an acute inflammatory reaction when injected into the foot-pads of endotoxin resistant mice.

PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for CXCL1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any CXCL1 present is captured by the coated antibody after incubation. Following extensive washing, a biotin-conjugate antibody specific for CXCL1 is added to detect the captured CXCL1 protein in sample. For signal development, horseradish peroxidase (HRP)-conjugated Streptavidin is added, followed by tetramethyl-benzidine (TMB) reagent. Following a wash to remove any unbound combination, and enzyme conjugate is added to the wells. Solution containing sulfuric acid is used to stop color development and the color intensity which is proportional to the quantity of bound protein is measurable at 450nm.

Schematic diagram: Analyte



1. This ELISA should not be used beyond the expiration data on the kit label.



- 2. To avoid cross-contamination, use a fresh reagent reservoir and pipette tips for each step.
- 3. To ensure accurate results, some details, such as technique, plasticware and water sources should be emphasized.
- 4. A thorough and consistent wash technique is essential for proper assay performance.
- 5. A standard curve should be generated for each set of samples assayed.
- 6. It is recommended that all standards and samples be assayed in duplicate.
- 7. Avoid microbial contamination of reagents and buffers. Buffers containing protein should be made under aseptic conditions and be prepared fresh daily.
- 8. In order to ensure the accuracy of the results, the standard curve should be made every time.

PRECAUTIONS

The Stop Solution suggested for use with this kit is an acid solution. Wear protective gloves, clothing, eye, and face protection. Wash hands thoroughly after handling.

KIT COMPONENTS& STORAGE CONDITIONS

PART	SIZE	STORAGE	OF	OPENED/
PARI	SIZE	RECONSTITU	TED MATI	ERIAL



Microwell Plate - antibody coated 96-well Microplate (8 wells ×12 strips)	1 plate	Return unused wells to the foil pouch containing the desiccant pack. Reseal along entire edge of the zip-seal. May be stored for up to 1 month at $2-8^{\circ}C^{**}$
Standard - lyophilized,2000 pg/ml upon reconstitution	2 vials	Aliquot and Store at -20°C** for six months
Concentrated Biotin-Conjugated antibody(100X) - 120 ul/vial	1 vial	Store at 2-8°C **for six months
Concentrated Streptavidin-HRP solution(100X) - 120 ul/vial	1 vial	Store at 2-8°C** for six months
Standard/Sample Diluent - 16 ml/ vial	1 bottle	Store at 2-8°C** for six months
Biotin-Conjugate antibody Diluent - 16 ml/vial	1 bottle	Store at 2-8°C** for six months
Streptavidin-HRP Diluent - 16 ml/vial	1 bottle	Store at 2-8°C** for six months
Wash Buffer Concentrate (20x) - 30 ml/vial	1 bottle	Store at 2-8°C** for six months
Substrate Solution - 12 ml/vial	1 bottle	Store at 2-8°C** for six months
Stop Solution - 12 ml/vial	1 bottle	Store at 2-8°C** for six months
Plate Cover Seals	4 pieces	

^{**}Provided this is within the expiration date of the kit.

OTHER SUPPLIES REQUIRED BUT NOT SUPPLIED

- 1. Microplate reader capable of measuring absorbance at 450 nm.
- 2. Pipettes and pipette tips.



- Deionized or distilled water.
- 4. Squirt bottle, manifold dispenser, or automated microplate washer.
- 5. 500 mL graduated cylinder.

SPECIMEN COLLECTION & STORAGE


Cell Culture Supernates - Centrifuge cell culture media at $1000 \times g$ to remove debris. Assay immediately or aliquot and store samples at \le -20 °C. Avoid repeated freeze-thaw cycles. **Serum** - Use a serum separator tube (SST) and allow samples to clot for 2 hours at room temperature or overnight at 2-8°C. Centrifuge approximately for 15 minutes at $1000 \times g$. Assay immediately or aliquot and store samples at \le -20 °C. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at $1000 \times g$ within 30 minutes of collection. Assay immediately or aliquot and store samples at \leq -20 °C. Avoid repeated freeze-thaw cycles.

Note: The normal human serum or plasma samples are suggested to make a 1:2 dilution.

REAGENTS PREPARATION

- 1. **Temperature returning** Bring all kit components and specimen to room temperature (20-25°C) before use.
- **2. Wash Buffer** Dilute 30mL of Wash Buffer Concentrate with 570mL of deionized or distilled water to prepare 600mL of Wash Buffer. If crystals have formed in the concentrate Wash Buffer, warm to room temperature and mix gently until the crystals have completely dissolved.
- 3. Standard/Specimen Reconstitute the Standard with 1.0mL of Standard/Sample Diluent. This reconstitution produces a stock solution of 2000pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 500µL of Standard/Sample Diluent into 1000pg/ml tube and the remaining tubes. Use the stock solution of 2000pg/mL to produce a 2-fold dilution series (below). Mix each tube thoroughly and change pipette tips between each transfer. The 1000 pg/mL standard serves as the high standard. The Standard/Sample Diluent serves as the zero standard (0 pg/mL).





Preparation of CXCL1 standard dilutions

- *If you do not run out of re-melting standard, store it at -20°C. Diluted standard shall not be reused.
- **4. Working solution of Biotin-Conjugate anti-human CXCL1 antibody:** Make a 1:100 dilution of the concentrated Biotin-Conjugate solution with the Biotin-Conjugate antibody Diluent in a clean plastic tube.
 - *The working solution should be used within one day after dilution.
- **5. Working solution of Streptavidin-HRP**: Make a 1:100 dilution of the concentrated Streptavidin-HRP solution with the Streptavidin-HRP Diluent in a clean plastic tube.
 - *The working solution should be used within one day after dilution.

ASSAY PROCEDURE

Prepare all reagents and standards as directed. Wash the plate 3 times before assay.		
Ţ,		
Add 100µl standard or samples to each well, incubate 90 minutes,37°C.		
Aspirate and wash 4 times		
Add 100µl working solution of Biotin-Conjugate anti-human CXCL1 antibody to each well, incubate 60 minutes,37°C.		



Add 100µl working solution of Streptavidin-HRP to each well, incubate 30 minutes,37°C.

 \square Aspirate and wash 5 times

Add 100µl Substrate solution to each well, incubate 15 minutes,37°C.Protect from light.

 \prod

Add 50µl Stop solution to each well. Read at 450nm within 5 minutes.

CALCULATION OF RESULTS


- 1. The standard curve is used to determine the amount of specimens.
- 2. First, average the duplicate readings for each standard, control, and sample. All O.D. values are subtracted by the mean value of blank control before result interpretation.
- 3. Construct a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph.
- 4. The data may be linearized by plotting the log of the CXCL1 concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
- 5. This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

Typical data using the CXCL1 ELISA

Standard(pg/ ml)	OD.	OD.	Average	Corrected
0	0.045	0.046	0.045	
15.625	0.093	0.094	0.093	0.048
31.25	0.184	0.186	0.185	0.132
62.5	0.329	0.346	0.337	0.285
125	0.512	0.534	0.523	0.4705



250	0.793	0.812	0.802	0.750
500	1.487	1.458	1.4725	1.420
1000	2.536	2.562	2.549	2.496



Representative standard curve for CXCL1 ELISA.

Performance Characteristics

SENSITIVITY: The minimum detectable dose was 8pg/mL.

SPECIFICITY: This assay recognizes both natural and recombinant human CXCL1. The factors listed below were prepared at 100ng/ml in Standard /Sample Diluent and assayed for cross-reactivity and no significant cross-reactivity or interference was observed.

Factors assayed for cross-reactivity



Recombinant human	Recombinant mouse	Recombinant porcine
CXCL2/GROβ	CCL3/MIP-1α	
CXCL5/ENA-78	CCL4/MIP-1β	
CXCL7/NAP-2	CXCL1/KC	
CXCL8/IL-8		
CCL2/MCP-1		
CCL3/MIP-1α		
CCL4/MIP-1β		
CCL5/RANTES		

REPEATABILITY: The coefficient of variation of both intra-assay and inter-assay were less than 10%.

RECOVERY: The recovery of CXCL1 spiked to three different levels in four samples throughout the range of the assay in various matrices was evaluated.

Recovery of CXCL1 in two matrices

Sample Type	Average % of Expected Range (%)	Range (%)
Citrate plasma	92	84-102
Cell culture supernatants	94	85-104

LINEARITY: To assess the linearity of the assay, three samples were spiked with high concentrations of CXCL1 in various matrices and diluted with the appropriate Sample Diluent to produce samples with values within the dynamic range of the assay. (The plasma samples were initially diluted 1:1)

Dilution ratio	Recovery (%)	Citrate plasma	Cell culture supernatants
1:2	Average% of Expected	95	102



	Range (%)	87-104	94-110
1.4	Average% of Expected	97	104
1:4	Range (%)	89-105	97-112
1:8	Average% of Expected	96	105
1.0	Range (%)	90-104	96-114
1.16	Average% of Expected	98	103
1:16	Range (%)	91-107	94-110

REFERENCES

- 1. Oppenheim, J.J. et al. (1991) Annu. Rev. Immunol. 9:617.
- 2. Schall, T.J. (1991) Cytokine 3:165.
- 3. Oppenheim, J.J. (1993) Proc. 3rd International Symposium Chemotactic Cytokines.
- 4. Anisowicz, A. et al. (1987) Proc. Natl. Acad. Sci. USA 84:7188.
- 5. Sager, R. et al. (1992) Cytokine 4:96.



- 6. Richmond, A. et al. (1988) EMBO J. 7:2025.
- 7. Tekamp-Olson, P. et al. (1990) J. Exp. Med. 172:911.
- 8. Haskill, S. et al. (1990) Proc. Natl. Acad. Sci. USA 87:7732.
- 9. Schröder, J-M. (1992) J. Invest. Dermatol. 98:241.
- 10. Unemoir, E.N. et al. (1993) J. Biol. Chem. 268:1338.
- 11. Schumacher, C. et al. (1992) Proc. Natl. Acad. Sci. USA 89:10542.
- 12. Horuk, R. et al. (1993) J. Biol. Chem. 268:541.