



Human CXCL4 Immunoassay

Catalog Number: EA800058

For the quantitative determination of human CXCL2 concentrations in cell culture supernates, serum, and plasma.

For research use only. Not for use in diagnostic procedures.

MANUFACTURED AND DISTRIBUTED BY:

OriGene Technologies, Inc.

9620 Medical Center Drive Suite 200
Rockville, MD 20850, USA

TABLE OF CONTENTS

SECTION	PAGE
BACKGROUND.....	1
PRINCIPLE OF THE ASSAY.....	1
TECHNICAL HINTS AND LIMITATIONS.....	2
PRECAUTIONS.....	2
KIT COMPONENTS& STORAGE CONDITIONS.....	3
OTHER SUPPLIES REQUIRED BUT NOT SUPPLIED.....	4
SPECIMEN COLLECTION & STORAGE.....	4
REAGENTS PREPARATION.....	4
ASSAY PROCEDURE	6
CALCULATION OF RESULTS.....	6
PERFORMANCE CHARACTERISTICS.....	8
REFERENCES.....	10



BACKGROUND

Platelet Factor 4 (PF4), also known as CXCL4, is an 8 kDa member of the CXC chemokine family that shares features with CXCL8/IL-8 and CXCL7/thymus chemokine-1/ NAP-2/ β -thromboglobulin. Mature human PF4 shares 65-76% amino acid sequence identity with mouse, rat, bovine, ovine and porcine PF4. The active protein is a tetramer of PF4 subunits that forms a ring of heparin-binding positive charges from sites at the C-terminal region of each monomer. Megakaryocytes synthesize PF4 and store it in platelet α -granules. Secretion from activated platelets can produce micromolar levels in serum and over 100-fold higher within clots. In contrast to other CXC chemokines, PF4 does not contain an ELR motif and lacks binding to nearly all chemokine receptors. A potential high-affinity G-protein-coupled receptor for PF4, the CXCR3 isoform CXCR3B, is expressed in human but not mouse. In most cases, it is likely that cell surface binding and signaling properties of PF4 are due to binding of glycosaminoglycan chains, particularly chondroitin sulfates.

PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for CXCL4 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any CXCL4 present is captured by the coated antibody after incubation. Following extensive washing, a biotin-conjugate antibody specific for CXCL4 is added to detect the captured CXCL4 protein in sample. For signal development, horseradish peroxidase (HRP)-conjugated Streptavidin is added, followed by tetramethyl-benzidine (TMB) reagent. Following a wash to remove any unbound combination, and enzyme conjugate is added to the wells. Solution containing sulfuric acid is used to stop color development and the color intensity which is proportional to the quantity of bound protein is measurable at 450nm.

Schematic diagram:



TECHNICAL HINTS AND LIMITATIONS



1. This ELISA should not be used beyond the expiration data on the kit label.
2. To avoid cross-contamination, use a fresh reagent reservoir and pipette tips for each step.
3. To ensure accurate results, some details, such as technique, plasticware and water sources should be emphasized.
4. A thorough and consistent wash technique is essential for proper assay performance.
5. A standard curve should be generated for each set of samples assayed.
6. It is recommended that all standards and samples be assayed in duplicate.
7. Avoid microbial contamination of reagents and buffers. Buffers containing protein should be made under aseptic conditions and be prepared fresh daily.
8. In order to ensure the accuracy of the results, the standard curve should be made every time.

PRECAUTIONS

The Stop Solution suggested for use with this kit is an acid solution. Wear protective gloves, clothing, eye, and face protection. Wash hands thoroughly after handling.

KIT COMPONENTS & STORAGE CONDITIONS

PART	SIZE	STORAGE OF OPENED/ RECONSTITUTED MATERIAL
Microwell Plate - antibody coated 96-well Microplate (8 wells ×12 strips)	1 plate	Return unused wells to the foil pouch containing the desiccant pack. Reseal along entire edge of the zip-seal. May be stored for up to 1 month at 2 – 8°C**
Standard - lyophilized,1000 pg/ml upon reconstitution	2 vials	Aliquot and Store at -20°C** for six months
Concentrated Biotin-Conjugated antibody(100X) - 120 ul/vial	1 vial	Store at 2-8°C **for six months
Concentrated Streptavidin-HRP solution(100X) - 120 ul/vial	1 vial	Store at 2-8°C** for six months
Standard/Sample Diluent - 16 ml/vial	1 bottle	Store at 2-8°C** for six months
Biotin-Conjugate antibody Diluent - 16 ml/vial	1 bottle	Store at 2-8°C** for six months
Streptavidin-HRP Diluent - 16 ml/vial	1 bottle	Store at 2-8°C** for six months
Wash Buffer Concentrate (20x) - 30 ml/vial	1 bottle	Store at 2-8°C** for six months
Substrate Solution - 12 ml/vial	1 bottle	Store at 2-8°C** for six months
Stop Solution - 12 ml/vial	1 bottle	Store at 2-8°C** for six months
Plate Cover Seals	4 pieces	

**Provided this is within the expiration date of the kit.

OTHER SUPPLIES REQUIRED BUT NOT SUPPLIED

1. Microplate reader capable of measuring absorbance at 450 nm.
2. Pipettes and pipette tips.
3. Deionized or distilled water.
4. Squirrt bottle, manifold dispenser, or automated microplate washer.
5. 500 mL graduated cylinder.

SPECIMEN COLLECTION & STORAGE

Cell Culture Supernates - Centrifuge cell culture media at 1000×g to remove debris. Assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.


Serum - Use a serum separator tube (SST) and allow samples to clot for 2 hours at room temperature or overnight at 2-8°C. Centrifuge approximately for 15 minutes at 1000×g. Assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000×g within 30 minutes of collection. Assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Note: The normal human serum or plasma samples are suggested to make a 1:2 dilution.

REAGENTS PREPARATION

1. **Temperature returning** - Bring all kit components and specimen to room temperature (20-25°C) before use.
2. **Wash Buffer** - Dilute 30mL of Wash Buffer Concentrate with 570mL of deionized or distilled water to prepare 600mL of Wash Buffer. If crystals have formed in the concentrate Wash Buffer, warm to room temperature and mix gently until the crystals have completely dissolved.
3. **Standard/Specimen** - Reconstitute the Standard with 1.0mL of Standard/Sample Diluent. This reconstitution produces a stock solution of 1000pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 500µL of Standard/Sample Diluent into 500pg/ml tube and the remaining tubes. Use the stock solution of 1000pg/mL to produce a 2-fold dilution series (below). Mix each tube thoroughly and change pipette tips between each transfer. The 1000 pg/mL standard serves as the high standard. The



Preparation of CXCL4 standard dilutions

***If you do not run out of re-melting standard, store it at -20°C. Diluted standard shall not be reused.**

- 4. Working solution of Biotin-Conjugate anti-human CXCL4 antibody:** Make a 1:100 dilution of the concentrated Biotin-Conjugate solution with the Biotin-Conjugate antibody Diluent in a clean plastic tube.

***The working solution should be used within one day after dilution.**

- 5. Working solution of Streptavidin-HRP:** Make a 1:100 dilution of the concentrated Streptavidin-HRP solution with the Streptavidin-HRP Diluent in a clean plastic tube.

***The working solution should be used within one day after dilution.**

ASSAY PROCEDURE

Prepare all reagents and standards as directed. Wash the plate 3 times before assay.



Add 100µl standard or samples to each well, incubate 90 minutes, 37°C.



Aspirate and wash 4 times

Add 100µl working solution of Biotin-Conjugate anti-human CXCL4 antibody to each well, incubate 60 minutes,37°C.

↓ Aspirate and wash 4 times

Add 100µl working solution of Streptavidin-HRP to each well, incubate 30 minutes,37°C.

↓ Aspirate and wash 5 times

Add 100µl Substrate solution to each well, incubate 15 minutes,37°C.Protect from light.

↓

Add 50µl Stop solution to each well. Read at 450nm within 5 minutes.


CALCULATION OF RESULTS

1. The standard curve is used to determine the amount of specimens.
2. First, average the duplicate readings for each standard, control, and sample. All O.D. values are subtracted by the mean value of blank control before result interpretation.
3. Construct a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph.
4. The data may be linearized by plotting the log of the CXCL4 concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
5. This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

Typical data using the CXCL4 ELISA

Standard(pg/ml)	OD.	OD.	Average	Corrected
0	0.065	0.067	0.066	---
15.625	0.106	0.098	0.102	0.036

31.25	0.145	0.152	0.148	0.082
62.5	0.223	0.243	0.233	0.167
125	0.392	0.412	0.402	0.336
250	0.758	0.735	0.746	0.680
500	1.335	1.356	1.345	1.279
1000	2.356	2.381	2.368	2.302



Representative standard curve for CXCL4 ELISA.

Performance Characteristics



SENSITIVITY: The minimum detectable dose was 8pg/mL.

SPECIFICITY: This assay recognizes both natural and recombinant human CXCL4. The factors listed below were prepared at 100ng/ml in Standard /Sample Diluent and assayed for cross-reactivity and no significant cross-reactivity or interference was observed.

Factors assayed for cross-reactivity

Recombinant human	Recombinant mouse	Recombinant porcine
Apolipoprotein H		
CXCL1/GRO α		
CXCL2/GRO β		
CXCL7/NAP-2		
CXCL8/IL-8		
CXCL9/MIG		
CXCL10/IP-10/CRG-2		
CXCL12/SDF-1		
CXCL16		
VEGF		

REPEATABILITY: The coefficient of variation of both intra-assay and inter-assay were less than 10%.

RECOVERY: The recovery of CXCL4 spiked to three different levels in four samples throughout the range of the assay in various matrices was evaluated.

Recovery of CXCL4 in two matrices

Sample Type	Average % of Expected Range (%)	Range (%)
Citrate plasma	93	85-102
Cell culture supernatants	96	89-104

LINEARITY: To assess the linearity of the assay, three samples were spiked with high concentrations of CXCL4 in various matrices and diluted with the appropriate Sample Diluent to



produce samples with values within the dynamic range of the assay. (The plasma samples were initially diluted 1:1)

Dilution ratio	Recovery (%)	Citrate plasma	Cell culture supernatants
1:2	Average% of Expected	91	103
	Range (%)	83-101	95-111
1:4	Average% of Expected	93	105
	Range (%)	84-102	95-116
1:8	Average% of Expected	92	110
	Range (%)	85-101	99-117
1:16	Average% of Expected	90	106
	Range (%)	81-103	95-113

REFERENCES



1. Poncz, M. et al. (1987) *Blood* 69:219.
2. Kowalska, M.A. et al. (2010) *Thromb. Res.* 125:292.
3. Slungaard, A. (2005) *Int. J. Biochem. Cell Biol.* 37:1162.
4. Zhang, X. et al. (1994) *Biochemistry* 33:8361.
5. Briquet-Laugier, V. et al. (2004) *J. Thromb. Haemost.* 2:2231.
6. El Golli, N et al. (2005) *J. Biol. Chem.* 280:30329.
7. Petersen, F. et al. (1998) *J. Immunol.* 161:4347.
8. Petersen, F. et al. (1999) *J. Biol. Chem.* 274:12376.
9. Dudek, A.Z. et al. (2003) *Blood* 101:4687.
10. Slungaard, A. et al. (2003) *Blood* 102:146.
11. Kowalska, M.A. et al. (2007) *Blood* 110:1903.
12. Preston, R.J.S. et al. (2009) *J. Biol. Chem.* 284:5869.
13. Fiore, M.M. and I.J. Mackie (2009) *Biochem. Biophys. Res. Commun.* 379:1072.
14. Amelot, A.A. et al. (2007) *J. Biol. Chem.* 282:710.
15. Eslin, D.E. et al. (2004) *Blood* 104:3173.
16. Warkentin, T.E. et al. (2009) *Blood* 113:4963.
17. Sikara, M.P. et al. (2010) *Blood* 115:713.
18. Lambert, M.P. et al. (2007) *Blood* 110:1153.
19. Lambert, M.P. et al. (2009) *Blood* 114:2290.
20. Nesmelova, I.V. et al. (2005) *J. Biol. Chem.* 280:4948.