

# **Human PDGF-BB Immunoassay**

Catalog Number: EA800095

For the quantitative determination of human Platelet-derived growth factor subunit B (PDGF-BB) concentrations in cell culture supernates, serum, and plasma.

For research use only. Not for use in diagnostic procedures.

#### MANUFACTURED AND DISTRIBUTED BY:

OriGene Technologies,Inc.

9620 Medical Center Drive Suite 200 Rockville, MD 20850,USA

Tel: 1-301-340-3188 Fax: 301-340-9254 Mail: techsupport@origene.com Web: www.origene.com



# TABLE OF CONTENTS

SECTION	PAGE	
BACKGROUND		1
PRINCIPLE OF THE		
ASSAY	1	
TECHNICAL HINTS AND LIMITATIONS		2
PRECAUTIONS	2	)
KIT COMPONENTS& STORAGE CONDITION	NS3	}
OTHER SUPPLIES REQUIRED BUT NOT SUI	PPLIED	4
SPECIMEN COLLECTION & STORAGE		4
REAGENTS PREPARATION	4	ļ
ASSAY PROCEDURE		6
CALCULATION OF RESULTS	6	3
PERFORMANCE CHARACTERISTICS	8	3
REFERENCES	10	7




## BACKGROUND

Interleukin- $1\beta$  (IL-1 beta)Platelet-derived growth factor subunit B is a protein that in humans is encoded by the PDGFB gene. The protein is a member of the platelet-derived growth factor family. The four members of this family are mitogenic factors for cells of mesenchymal origin and are characterized by a motif of eight cysteines. This gene product can exist either as a homodimer (PDGF-BB) or as a heterodimer with the platelet-derived growth factor alpha polypeptide (PDGF-AB), where the dimers are connected by disulfide bonds. PDGFs are mitogenic during early developmental stages, driving the proliferation of undifferentiated mesenchyme and some progenitor populations. During later maturation stages, PDGF signalling has been implicated in tissue remodelling and cellular differentiation, and in inductive events involved in patterning and morphogenesis. In addition to driving mesenchymal proliferation, PDGFs have been shown to direct the migration, differentiation and function of a variety of specialised mesenchymal and migratory cell types, both during development and in the adult animal.

## PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for PDGF-BB has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any PDGF-BB present is captured by the coated antibody after incubation. Following extensive washing, a biotin-conjugate antibody specific for PDGF-BB is added to detect the captured PDGF-BB protein in sample. For signal development, horseradish peroxidase (HRP)-conjugated Streptavidin is added, followed by tetramethyl-benzidine (TMB) reagent. Following a wash to remove any unbound combination, and enzyme conjugate is added to the wells. Solution containing sulfuric acid is used to stop color development and the color intensity which is proportional to the quantity of bound protein is measurable at 450nm.

# **Schematic diagram:**





## TECHNICAL HINTS AND LIMITATIONS

- 1. This ELISA should not be used beyond the expiration data on the kit label.
- 2. To avoid cross-contamination, use a fresh reagent reservoir and pipette tips for each step.
- 3. To ensure accurate results, some details, such as technique, plasticware and water sources should be emphasized.
- 4. A thorough and consistent wash technique is essential for proper assay performance.
- 5. A standard curve should be generated for each set of samples assayed.
- 6. It is recommended that all standards and samples be assayed in duplicate.
- 7. Avoid microbial contamination of reagents and buffers. Buffers containing protein should be made under aseptic conditions and be prepared fresh daily.
- 8. In order to ensure the accuracy of the results, the standard curve should be made every time.

## **PRECAUTIONS**

The Stop Solution suggested for use with this kit is an acid solution. Wear protective gloves, clothing, eye, and face protection. Wash hands thoroughly after handling.



# KIT COMPONENTS& STORAGE CONDITIONS

PART	SIZE	STORAGE OF OPENED/ RECONSTITUTED MATERIAL	
<b>Microwell Plate</b> - antibody coated 96-well Microplate (8 wells ×12 strips)	1 plate	Return unused wells to the foil pour containing the desiccant pack. Reseal along entire edge of the zip-seal. May be store for up to 1 month at 2 – 8°C**	
<b>Standard</b> - lyophilized,4000 pg/ml upon reconstitution	2 vials	Aliquot and Store at -20°C** for six months	
Concentrated Biotin-Conjugated antibody(100X) - 120 ul/vial	1 vial	Store at 2-8°C **for six months	
Concentrated Streptavidin-HRP solution(100X) - 120 ul/vial	1 vial	Store at 2-8°C** for six months	
<b>Standard /sample Diluent</b> - 16 ml/vial	1 bottle	Store at 2-8°C** for six months	
Biotin-Conjugate antibody Diluent - 16 ml/vial	1 bottle	Store at 2-8°C** for six months	
<b>Streptavidin-HRP Diluent</b> - 16 ml/vial	1 bottle	Store at 2-8°C** for six months	
<b>Wash Buffer Concentrate</b> (20x) - 30 ml/vial	1 bottle	Store at 2-8°C** for six months	
<b>Substrate Solution</b> - 12 ml/vial	1 bottle	Store at 2-8°C** for six months	
Stop Solution - 12 ml/vial	1 bottle	Store at 2-8°C** for six months	
Plate Cover Seals	4 pieces		

<sup>\*\*</sup>Provided this is within the expiration date of the kit.



# OTHER SUPPLIES REQUIRED BUT NOT SUPPLIED

- 1. Microplate reader capable of measuring absorbance at 450 nm.
- 2. Pipettes and pipette tips.
- 3. Deionized or distilled water.
- 4. Squirt bottle, manifold dispenser, or automated microplate washer.
- 5. 500 mL graduated cylinder.
- 6. Human IL-1 beta controls (optional; available from Solarbio).

# SPECIMEN COLLECTION & STORAGE

**Cell Culture Supernates** - Centrifuge cell culture media at  $1000 \times g$  to remove debris. Assay immediately or aliquot and store samples at  $\leq$  -20 °C. Avoid repeated freeze-thaw cycles. **Serum** - Use a serum separator tube (SST) and allow samples to clot for 2 hours at room

temperature or overnight at 2-8°C. Centrifuge at approximately for 15 minutes at 1000×g. Assay

immediately or aliquot and store samples at  $\leq$  -20 °C. Avoid repeated freeze-thaw cycles.

**Plasma** - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at  $1000 \times g$  within 30 minutes of collection. Assay immediately or aliquot and store samples at  $\leq$  -20 °C. Avoid repeated freeze-thaw cycles.

Note: The normal human serum or plasma samples are suggested to make a 1:2 dilution.


## REAGENTS PREPARATION

- 1. **Temperature returning** Bring all kit components and specimen to room temperature (20-25°C) before use.
- **2. Wash Buffer** Dilute 30mL of Wash Buffer Concentrate with 570mL of deionized or distilled water to prepare 200mL of Wash Buffer. If crystals have formed in the concentrate Wash Buffer, warm to room temperature and mix gently until the crystals have completely dissolved.
- **3. Standard\Specimen** Reconstitute the Standard with 1.0mL of deionized or distilled water. This reconstitution produces a stock solution of 4000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 500μL of Standard/Specimen Diluent into the 2000 pg/mL tube, and the remaining tubes. Use the high standard to produce a 2-fold dilution series (below). Mix each tube thoroughly and change pipette tips between each transfer. The 2000 pg/mL standard serves as the high standard. The Standard/specimen Diluent serves as the zero standard (0 pg/mL).
  - \*If you do not run out of re-melting standard, store it at -20°C. Diluted standard shall not be reused.
- **4. Working solution of Biotin-Conjugate anti-human PDGF-BB antibody:** Make a 1:100 dilution of the concentrated Biotin-Conjugate solution with the Biotin-Conjugate antibody



Diluent in a clean plastic tube.

- \*The working solution should be used within one day after dilution.
- **5. Working solution of Streptavidin-HRP**: Make a 1:100 dilution of the concentrated Streptavidin-HRP solution with the Streptavidin-HRP Diluent in a clean plastic tube.
  - \*The working solution should be used within one day after dilution.



**Preparation of PDGF-BB standard dilutions** 

# **ASSAY PROCEDURE**

Prepare all reagents and standards as directed.



Add 100µl standard or samples to each well, incubate 90 minutes,37°C.



Aspirate and wash 4 times

Add 100µl working solution of Biotin-Conjugate anti-human PDGF-BB antibody to each well, incubate 60 minutes,37°C.

 $\square$  Aspirate and wash 4 times

Add 100µl working solution of Streptavidin-HRP to each well, incubate 30 minutes,37°C.

 $\bigcirc$  Aspirate and wash 5 times

Add 100µl Substrate solution to each well, incubate 15 minutes,37°C.Protect from light.

IJ

 $Add\ 50\mu l\ Stop$  solution to each well. Read at 450nm within 30 minutes.

## CALCULATION OF RESULTS


- 1. The standard curve is used to determine the amount of specimens.
- 2. First, average the duplicate readings for each standard, control, and sample. All O.D. values are subtracted by the mean value of blank control before result interpretation.
- 3. Construct a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph.
- 4. The data may be linearized by plotting the log of the PDGF-BB concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
- 5. This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

## Typical data using the PDGF-BB ELISA

Standard(pg/ ml)	OD.	OD.	Average	Corrected
0	0.015	0.016	0.015	
31.25	0.071	0.068	0.069	0.054



62.5	0.098	0.096	0.097	0.082
125	0.168	0.172	0.170	0.155
250	0.321	0.317	0.319	0.303
500	0.642	0.665	0.653	0.638
1000	1.151	1.157	1.154	1.138
2000	2.205	2.321	2.263	2.247



Representative standard curve for PDGF-BB ELISA.

# **Performance Characteristics**

**SENSITIVITY:** The minimum detectable dose was 15pg/mL.

**SPECIFICITY:** This assay recognizes both natural and recombinant human PDGF-BB. The factors listed below were prepared at 100ng/ml in Standard /sample Diluent and assayed for cross-reactivity and no significant cross-reactivity or interference was observed.

#### Factors assayed for cross-reactivity



Recombinant human	Recombinant mouse	Recombinant rat
PDGF-AB	PDGF-BB	PDGF-BB
PDGF-AA		

**REPEATABILITY:** The coefficient of variation of both intra-assay and inter-assay were less than 10%.

**RECOVERY:** The recovery of PDGF-BB spiked to three different levels in four samples throughout the range of the assay in various matrices was evaluated.

**Recovery of PDGF-BB in two matrices** 

Sample Type	Average % of Expected Range (%)	Range (%)
Citrate plasma	93	88-98
Cell culture supernatants	105	99-111

**LINEARITY:** To assess the linearity of the assay, three samples were spiked with high concentrations of PDGF-BB in various matrices and diluted with the appropriate Sample Diluent to produce samples with values within the dynamic range of the assay. (The plasma samples were initially diluted 1:1)

Dilution ratio	Recovery (%)	Citrate plasma	Cell culture supernatants
1:2	Average% of Expected	90	101
1.2	Range (%)	85-95	95-107
1.4	Average% of Expected	97	108
1:4	Range (%)	93-101	103-113

# **REFERENCES**

1. Ratner L, et al. (Sep 1985). Nucleic Acids Res 13 (14): 5007–18.



- 2. Clements JM, et al. (Jan 1992). EMBO J 10 (13): 4113–20.
- 3. Hoch RV, Soriano P (2003). Development 130 (20): 4769–4784.