OriGene Technologies, Inc.
9620 Medical Center Drive, Ste 200
Rockville, MD 20850, US
Phone: +1-888-267-4436
https://www.origene.com
techsupport@origene.com
EU: info-de@origene.com
CN: techsupport@origene.cn

Product datasheet for HP205638

COPT2 (SLC31A2) Human qPCR Primer Pair (NM_001860)

Product data:

Product Type:	qPCR Primer Pairs
Gene ID:	1318
Forward Sequence:	ATCAGCCAGCAGACCATCGCAG
Reverse Sequence:	TGAAGTAGCCGATGACCACCTG
ACCN:	NM_001860, NM_001860.1, NM_001860.2, BC026252, BC026252.1, BP341248, NM_001860.3
UniProt ID:	O15432
Synonyms:	COPT2; CTR2; hCTR2
Components:	1 vial of lyophilized qSTAR qPCR primer mix (1 nmol each primer, sufficient for 200 reactions). Before use, reconstitute the primer mix in 200 µl dH2O to make a final concentration of 10 µM.
Quality Control:	The primer mix has been tested to generate satisfactory qPCR data on ABI 7900HT by using the following PCR program: Stage 1: Activation: 50 °C for 2 min; Stage 2: pre-soak:95 °C for 10 min; Stage 3: Denaturation: 95 °C for 15 sec, Annealing: 60°C for 1 min; Stage 4: Melting curve: 95°C for 15 sec, 60°C for 15 sec, 95°C for 15 sec.
Storage:	Store at -20°C.
Stability:	The primer mix is stable for one year from date of shipping.

View online »

This product is to be used for laboratory only. Not for diagnostic or therapeutic use.
©2024 OriGene Technologies, Inc., 9620 Medical Center Drive, Ste 200, Rockville, MD 20850, US